

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** ICES02

**Lab Code:** ACE

**SDG NO.:** Q3399

**Continuing Calib Date:** 10/24/2025

**Initial Calibration Date(s):** 10/13/2025 10/13/2025

**Continuing Calib Time:** 14:28

**Initial Calibration Time(s):** 10:53 12:29

**GC Column:** RTX-CLP **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.28       | 7.28      | 7.18      | 7.38 | 0.00       |
| 2,4-D            | 8.41       | 8.41      | 8.31      | 8.51 | 0.00       |
| 2,4,5-TP(Silvex) | 9.28       | 9.28      | 9.18      | 9.38 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** ICES02

**Lab Code:** ACE

**SDG NO.:** Q3399

**Continuing Calib Date:** 10/24/2025

**Initial Calibration Date(s):** 10/13/2025 10/13/2025

**Continuing Calib Time:** 14:28

**Initial Calibration Time(s):** 10:53 12:29

**GC Column:** RTX-CLP2 **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.74       | 7.74      | 7.64      | 7.84 | 0.00       |
| 2,4-D            | 9.00       | 9.00      | 8.90      | 9.10 | 0.00       |
| 2,4,5-TP(Silvex) | 9.89       | 9.89      | 9.79      | 9.99 | 0.00       |