

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/13/2025 10/13/2025

Continuing Calib Time: 17:53

Initial Calibration Time(s): 10:53 12:29

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.28	7.28	7.18	7.38	0.00
2,4-D	8.41	8.41	8.31	8.51	0.01
2,4,5-TP(Silvex)	9.27	9.28	9.18	9.38	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/13/2025 10/13/2025

Continuing Calib Time: 17:53

Initial Calibration Time(s): 10:53 12:29

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
2,4-DCAA	7.74	7.74	7.64	7.84	0.00
2,4-D	9.00	9.00	8.90	9.10	0.00
2,4,5-TP(Silvex)	9.89	9.89	9.79	9.99	0.00