

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032132.D **Time Analyzed:** 10:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.276	9.180	9.380	768.020	712.500	7.8
2,4-D	8.409	8.312	8.512	751.270	705.000	6.6
2,4-DCAA	7.280	7.182	7.382	810.260	750.000	8.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032132.D **Time Analyzed:** 10:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.892	9.792	9.992	773.160	712.500	8.5
2,4-D	9.001	8.902	9.102	774.470	705.000	9.9
2,4-DCAA	7.744	7.644	7.844	877.970	750.000	17.1