

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032144.D **Time Analyzed:** 15:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.277	9.180	9.380	756.490	712.500	6.2
2,4-D	8.407	8.312	8.512	764.020	705.000	8.4
2,4-DCAA	7.279	7.182	7.382	794.670	750.000	6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032144.D **Time Analyzed:** 15:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.892	9.792	9.992	756.750	712.500	6.2
2,4-D	9.000	8.902	9.102	778.940	705.000	10.5
2,4-DCAA	7.743	7.644	7.844	878.420	750.000	17.1