

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032148.D **Time Analyzed:** 17:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.276	9.180	9.380	749.300	712.500	5.2
2,4-D	8.408	8.312	8.512	750.340	705.000	6.4
2,4-DCAA	7.279	7.182	7.382	797.050	750.000	6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032148.D **Time Analyzed:** 17:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.891	9.792	9.992	752.170	712.500	5.6
2,4-D	9.001	8.902	9.102	766.900	705.000	8.8
2,4-DCAA	7.744	7.644	7.844	865.680	750.000	15.4