

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032151.D **Time Analyzed:** 10:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.276	9.180	9.380	728.970	712.500	2.3
2,4-D	8.409	8.312	8.512	712.540	705.000	1.1
2,4-DCAA	7.279	7.182	7.382	772.240	750.000	3.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032151.D **Time Analyzed:** 10:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.890	9.792	9.992	745.250	712.500	4.6
2,4-D	9.000	8.902	9.102	739.330	705.000	4.9
2,4-DCAA	7.743	7.644	7.844	834.560	750.000	11.3