

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032154.D **Time Analyzed:** 14:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.275	9.180	9.380	738.430	712.500	3.6
2,4-D	8.407	8.312	8.512	733.220	705.000	4.0
2,4-DCAA	7.278	7.182	7.382	790.230	750.000	5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/24/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032154.D **Time Analyzed:** 14:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.891	9.792	9.992	750.010	712.500	5.3
2,4-D	9.000	8.902	9.102	761.860	705.000	8.1
2,4-DCAA	7.743	7.644	7.844	856.540	750.000	14.2