



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8151A</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>PS032134.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170220BS (Column 1)	2,4-D	5	4.80	ug/L	96				45	152	
	2,4,5-TP(Silvex)	5	4.80	ug/L	96				51	134	
PB170220BS (Column 2)	2,4-D	5	4.50	ug/L	90				45	152	
	2,4,5-TP(Silvex)	5	4.80	ug/L	96				51	134	