

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: ICES02

Lab Code: ACE
SDG NO.: Q3399

Instrument ID: ECD_D
Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR1
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.71	4.61	4.81	185217000
		2	5.24	5.14	5.34	185731000
		3	5.94	5.84	6.04	758191000
		4	6.03	5.93	6.13	917583000
		5	6.87	6.77	6.97	148478000
Toxaphene	500	1	6.24	6.14	6.34	32564400
		2	6.44	6.34	6.54	48621800
		3	7.14	7.04	7.24	85710600
		4	7.56	7.46	7.66	95737400
		5	7.92	7.82	8.02	60480500

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: ICES02

Lab Code: ACE
SDG NO.: Q3399

Instrument ID: ECD_D
Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR2
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.90	3.80	4.00	1210370000
		2	4.48	4.38	4.58	1263780000
		3	5.12	5.02	5.22	3771830000
		4	5.18	5.08	5.28	3251870000
		5	6.08	5.98	6.18	1460510000
Toxaphene	500	1	5.47	5.37	5.57	263702000
		2	5.64	5.54	5.74	176560000
		3	6.75	6.65	6.85	658883000
		4	7.19	7.09	7.29	495104000
		5	7.32	7.22	7.42	246033000