

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ICES02  
**Lab Code:** ACE **SDG NO.:** Q3399  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 10/22/2025  
**Lab Sample No.:** PSTDCCC050 **Data File :** PD090755.D **Time Analyzed:** 14:59

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Decachlorobiphenyl   | 9.068 | 8.964                | 9.164 | 46.320             | 50.000            | -7.4 |
| Endrin               | 6.570 | 6.468                | 6.668 | 47.990             | 50.000            | -4.0 |
| gamma-BHC (Lindane)  | 4.326 | 4.225                | 4.425 | 47.990             | 50.000            | -4.0 |
| Heptachlor           | 4.924 | 4.823                | 5.023 | 57.070             | 50.000            | 14.1 |
| Heptachlor epoxide   | 5.685 | 5.584                | 5.784 | 49.030             | 50.000            | -1.9 |
| Methoxychlor         | 7.490 | 7.388                | 7.588 | 49.890             | 50.000            | -0.2 |
| Tetrachloro-m-xylene | 3.545 | 3.445                | 3.645 | 46.510             | 50.000            | -7.0 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ICES02  
**Lab Code:** ACE **SDG NO.:** Q3399  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 10/22/2025  
**Lab Sample No.:** PSTDCCC050 **Data File :** PD090755.D **Time Analyzed:** 14:59

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Decachlorobiphenyl   | 8.062 | 7.962                | 8.162 | 54.910             | 50.000            | 9.8  |
| Endrin               | 5.781 | 5.681                | 5.881 | 48.490             | 50.000            | -3.0 |
| gamma-BHC (Lindane)  | 3.722 | 3.623                | 3.823 | 49.500             | 50.000            | -1.0 |
| Heptachlor           | 4.074 | 3.975                | 4.175 | 49.470             | 50.000            | -1.1 |
| Heptachlor epoxide   | 4.864 | 4.764                | 4.964 | 47.790             | 50.000            | -4.4 |
| Methoxychlor         | 6.745 | 6.645                | 6.845 | 52.980             | 50.000            | 6.0  |
| Tetrachloro-m-xylene | 2.874 | 2.775                | 2.975 | 47.350             | 50.000            | -5.3 |