

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.067	8.964	9.164	44.170	50.000	-11.7
Endrin	6.569	6.468	6.668	46.250	50.000	-7.5
gamma-BHC (Lindane)	4.324	4.225	4.425	47.530	50.000	-4.9
Heptachlor	4.923	4.823	5.023	55.620	50.000	11.2
Heptachlor epoxide	5.685	5.584	5.784	46.870	50.000	-6.3
Methoxychlor	7.489	7.388	7.588	46.180	50.000	-7.6
Tetrachloro-m-xylene	3.544	3.445	3.645	46.320	50.000	-7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.061	7.962	8.162	48.710	50.000	-2.6
Endrin	5.780	5.681	5.881	46.650	50.000	-6.7
gamma-BHC (Lindane)	3.722	3.623	3.823	48.880	50.000	-2.2
Heptachlor	4.074	3.975	4.175	48.710	50.000	-2.6
Heptachlor epoxide	4.864	4.764	4.964	46.510	50.000	-7.0
Methoxychlor	6.745	6.645	6.845	48.070	50.000	-3.9
Tetrachloro-m-xylene	2.874	2.775	2.975	47.070	50.000	-5.9