

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090773.D **Time Analyzed:** 19:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.066	8.964	9.164	48.090	50.000	-3.8
Endrin	6.568	6.468	6.668	49.010	50.000	-2.0
gamma-BHC (Lindane)	4.324	4.225	4.425	49.170	50.000	-1.7
Heptachlor	4.923	4.823	5.023	57.890	50.000	15.8
Heptachlor epoxide	5.685	5.584	5.784	49.640	50.000	-0.7
Methoxychlor	7.489	7.388	7.588	49.960	50.000	-0.1
Tetrachloro-m-xylene	3.544	3.445	3.645	47.760	50.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090773.D **Time Analyzed:** 19:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.061	7.962	8.162	55.760	50.000	11.5
Endrin	5.780	5.681	5.881	49.660	50.000	-0.7
gamma-BHC (Lindane)	3.722	3.623	3.823	50.230	50.000	0.5
Heptachlor	4.073	3.975	4.175	49.520	50.000	-1.0
Heptachlor epoxide	4.863	4.764	4.964	48.340	50.000	-3.3
Methoxychlor	6.744	6.645	6.845	53.040	50.000	6.1
Tetrachloro-m-xylene	2.874	2.775	2.975	48.360	50.000	-3.3