

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: ICE Service Group, Inc.

Project: NWIRP Northrup Grumman Site – Bethpage, NY

Client Sample ID: COMP-ROLLOFFMS

Lab Sample ID: Q3399-02MS Analytical Method: 8081B

Sample Wt/Vol: 100 mL Final Vol: 10000 uL

Prep Method: SW3510 Prep Date 10/22/25

Date Collected:	10/20/25
Date Received:	10/20/25
SDG No.:	Q3399
Matrix:	TCLP
% Solid:	0

Test: TCLP Pesticide

CAS Number	r Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	5.00		1	0.037	0.10	0.50	ug/L	10/22/25 18:38	PB170218
76-44-8	Heptachlor	5.90		1	0.027	0.10	0.50	ug/L	10/22/25 18:38	PB170218
1024-57-3	Heptachlor epoxide	5.00		1	0.096	0.25	0.50	ug/L	10/22/25 18:38	PB170218
72-20-8	Endrin	5.10		1	0.032	0.10	0.50	ug/L	10/22/25 18:38	PB170218
72-43-5	Methoxychlor	5.40		1	0.11	0.25	0.50	ug/L	10/22/25 18:38	PB170218
8001-35-2	Toxaphene	5.00	U	1	1.70	5.00	10.0	ug/L	10/22/25 18:38	PB170218
57-74-9	Chlordane	2.50	U	1	0.88	2.50	5.00	ug/L	10/22/25 18:38	PB170218
SURROGATES										
2051-24-3	Decachlorobiphenyl	21.2			30 - 135		106%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.4			44 - 124		107%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit