

LAB CHRONICLE

OrderID:	Q3399	OrderDate:	10/20/2025 1:23:08 PM
Client:	ICE Service Group, Inc.	Project:	NWIRP Northrup Grumman Site – Bethpage, NY
Contact:	Dennis D. Morgan, II	Location:	D41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3399-01	COMP-ROLLOFF	SOIL			10/20/25			10/20/25
			Diesel Range Organics	8015D		10/22/25	10/22/25	
			Gasoline Range Organics	8015D			10/21/25	
			Herbicide	8151A		10/22/25	10/23/25	
			PCB	8082A		10/22/25	10/22/25	
			Pesticide-TCL	8081B		10/22/25	10/22/25	
Q3399-02	COMP-ROLLOFF	TCLP			10/20/25			10/20/25
			TCLP Herbicide	8151A		10/22/25	10/23/25	
			TCLP Pesticide	8081B		10/22/25	10/22/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3399

Order ID: Q3399

Client: ICE Service Group, Inc.

Project ID: NWIRP Northrup Grumman Site – Be

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-----	-------

Client ID :

Total Concentration: 0.000



SAMPLE DATA

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/22/25
Client Sample ID:	PB170172TB		SDG No.:	Q3399
Lab Sample ID:	PB170172TB		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	100 mL	Final Vol:	10000 uL	Test:
Prep Method:	SW3510	Prep Date	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.10	U	1	0.037	0.10	0.50	ug/L	10/22/25 17:58	PB170218
76-44-8	Heptachlor	0.10	U	1	0.027	0.10	0.50	ug/L	10/22/25 17:58	PB170218
1024-57-3	Heptachlor epoxide	0.25	U	1	0.096	0.25	0.50	ug/L	10/22/25 17:58	PB170218
72-20-8	Endrin	0.10	U	1	0.032	0.10	0.50	ug/L	10/22/25 17:58	PB170218
72-43-5	Methoxychlor	0.25	U	1	0.11	0.25	0.50	ug/L	10/22/25 17:58	PB170218
8001-35-2	Toxaphene	5.00	U	1	1.70	5.00	10.0	ug/L	10/22/25 17:58	PB170218
57-74-9	Chlordane	2.50	U	1	0.88	2.50	5.00	ug/L	10/22/25 17:58	PB170218
SURROGATES										
2051-24-3	Decachlorobiphenyl	24.2			30 - 135		121%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.9			44 - 124		104%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/20/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/20/25
Client Sample ID:	COMP-ROLLOFF		SDG No.:	Q3399
Lab Sample ID:	Q3399-02		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	100 mL	Final Vol:	10000 uL	Test:
Prep Method:	SW3510	Prep Date	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.10	U	1	0.037	0.10	0.50	ug/L	10/22/25 15:54	PB170218
76-44-8	Heptachlor	0.10	U	1	0.027	0.10	0.50	ug/L	10/22/25 15:54	PB170218
1024-57-3	Heptachlor epoxide	0.25	U	1	0.096	0.25	0.50	ug/L	10/22/25 15:54	PB170218
72-20-8	Endrin	0.10	U	1	0.032	0.10	0.50	ug/L	10/22/25 15:54	PB170218
72-43-5	Methoxychlor	0.25	U	1	0.11	0.25	0.50	ug/L	10/22/25 15:54	PB170218
8001-35-2	Toxaphene	5.00	U	1	1.70	5.00	10.0	ug/L	10/22/25 15:54	PB170218
57-74-9	Chlordane	2.50	U	1	0.88	2.50	5.00	ug/L	10/22/25 15:54	PB170218
SURROGATES										
2051-24-3	Decachlorobiphenyl	20.1			30 - 135		101%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	19.6			44 - 124		98%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090647.D	PIBLK-PD090647.D	Decachlorobiphen	1	20	21.3	106		30	135
		Tetrachloro-m-xyl	1	20	19.4	97		44	124
		Decachlorobiphen	2	20	20.3	102		30	135
		Tetrachloro-m-xyl	2	20	20.0	100		44	124
I.BLK-PD090753.D	PIBLK-PD090753.D	Decachlorobiphen	1	20	22.1	111		30	135
		Tetrachloro-m-xyl	1	20	19.9	99		44	124
		Decachlorobiphen	2	20	25.1	126		30	135
		Tetrachloro-m-xyl	2	20	20.7	104		44	124
Q3399-02	COMP-ROLLOFF	Decachlorobiphen	1	20	18.5	93		30	135
		Tetrachloro-m-xyl	1	20	19.6	98		44	124
		Decachlorobiphen	2	20	20.1	101		30	135
		Tetrachloro-m-xyl	2	20	19.2	96		44	124
I.BLK-PD090763.D	PIBLK-PD090763.D	Decachlorobiphen	1	20	21.0	105		30	135
		Tetrachloro-m-xyl	1	20	19.9	100		44	124
		Decachlorobiphen	2	20	21.0	105		30	135
		Tetrachloro-m-xyl	2	20	20.6	103		44	124
PB170218BL	PB170218BL	Decachlorobiphen	1	20	20.4	102		30	135
		Tetrachloro-m-xyl	1	20	19.5	98		44	124
		Decachlorobiphen	2	20	22.3	112		30	135
		Tetrachloro-m-xyl	2	20	20.1	101		44	124
PB170218BS	PB170218BS	Decachlorobiphen	1	20	21.1	106		30	135
		Tetrachloro-m-xyl	1	20	16.9	85		44	124
		Decachlorobiphen	2	20	23.4	117		30	135
		Tetrachloro-m-xyl	2	20	17.3	86		44	124
PB170172TB	PB170172TB	Decachlorobiphen	1	20	21.7	108		30	135
		Tetrachloro-m-xyl	1	20	20.1	100		44	124
		Decachlorobiphen	2	20	24.2	121		30	135
		Tetrachloro-m-xyl	2	20	20.9	104		44	124
Q3399-02MS	COMP-ROLLOFFMS	Decachlorobiphen	1	20	19.1	96		30	135
		Tetrachloro-m-xyl	1	20	21.4	107		44	124
		Decachlorobiphen	2	20	21.2	106		30	135
		Tetrachloro-m-xyl	2	20	20.6	103		44	124
Q3399-02MSD	COMP-ROLLOFFMSD	Decachlorobiphen	1	20	19.2	96		30	135
		Tetrachloro-m-xyl	1	20	21.1	106		44	124
		Decachlorobiphen	2	20	20.7	104		30	135
		Tetrachloro-m-xyl	2	20	20.1	100		44	124
I.BLK-PD090772.D	PIBLK-PD090772.D	Decachlorobiphen	1	20	21.4	107		30	135
		Tetrachloro-m-xyl	1	20	19.9	100		44	124
		Decachlorobiphen	2	20	24.0	120		30	135
		Tetrachloro-m-xyl	2	20	20.5	103		44	124

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090770.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3399-02MS	Client Sample ID:		COMP-ROLLOFFMS								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	5.00	ug/L	100				59	134	
	Heptachlor	5	0	5.90	ug/L	118				54	130	
	Heptachlor epoxide	5	0	5.00	ug/L	100				61	133	
	Endrin	5	0	5.10	ug/L	102				60	138	
	Methoxychlor	5	0	5.20	ug/L	104				54	145	
Lab Sample ID:	Q3399-02MS	Client Sample ID:		COMP-ROLLOFFMS								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	4.80	ug/L	96				59	134	
	Heptachlor	5	0	4.90	ug/L	98				54	130	
	Heptachlor epoxide	5	0	4.90	ug/L	98				61	133	
	Endrin	5	0	5.00	ug/L	100				60	138	
	Methoxychlor	5	0	5.40	ug/L	108				54	145	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>8081B</u>
Client:	<u>ICE Service Group, Inc.</u>	DataFile :	<u>PD090771.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q3399-02MSD	Client Sample ID:		COMP-ROLLOFFMSD								
	(Column 1)											
	gamma-BHC (Lindane)	5	0	4.90	ug/L	98		2		59	134	20
	Heptachlor	5	0	5.80	ug/L	116		2		54	130	20
	Heptachlor epoxide	5	0	5.40	ug/L	108		8		61	133	20
	Endrin	5	0	5.00	ug/L	100		2		60	138	20
	Methoxychlor	5	0	5.20	ug/L	104		0		54	145	20
Lab Sample ID:	Q3399-02MSD	Client Sample ID:		COMP-ROLLOFFMSD								
	(Column 2)											
	gamma-BHC (Lindane)	5	0	4.70	ug/L	94		2		59	134	20
	Heptachlor	5	0	4.80	ug/L	96		2		54	130	20
	Heptachlor epoxide	5	0	5.30	ug/L	106		8		61	133	20
	Endrin	5	0	5.00	ug/L	100		0		60	138	20
	Methoxychlor	5	0	5.30	ug/L	106		2		54	145	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3399

Analytical Method: 8081B

Client: ICE Service Group, Inc.

Datafile : PD090766.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170218BS (Column 1)	gamma-BHC (Lindane)	0.5	0.44	ug/L	88				59	134	
	Heptachlor	0.5	0.52	ug/L	105				54	130	
	Heptachlor epoxide	0.5	0.45	ug/L	90				61	133	
	Endrin	0.5	0.45	ug/L	89				60	138	
	Methoxychlor	0.5	0.46	ug/L	91				54	145	
PB170218BS (Column 2)	gamma-BHC (Lindane)	0.5	0.45	ug/L	90				59	134	
	Heptachlor	0.5	0.46	ug/L	92				54	130	
	Heptachlor epoxide	0.5	0.45	ug/L	90				61	133	
	Endrin	0.5	0.46	ug/L	92				60	138	
	Methoxychlor	0.5	0.48	ug/L	95				54	145	

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170218BL

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: PB170218BL

Lab File ID: PD090765.D

Matrix: (soil/water) water

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 10/22/2025

Date Analyzed (1): 10/22/2025

Date Analyzed (2): 10/22/2025

Time Analyzed (1): 17:30

Time Analyzed (2): 17:30

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
COMP-ROLLOFF	Q3399-02	PD090759.D	10/22/2025	10/22/2025
PB170218BS	PB170218BS	PD090766.D	10/22/2025	10/22/2025
PB170172TB	PB170172TB	PD090767.D	10/22/2025	10/22/2025
COMP-ROLLOFFMS	Q3399-02MS	PD090770.D	10/22/2025	10/22/2025
COMP-ROLLOFFMSD	Q3399-02MSD	PD090771.D	10/22/2025	10/22/2025

COMMENTS: _____



QC SAMPLE DATA

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	
Client Sample ID:	PB170218BL		SDG No.:	Q3399
Lab Sample ID:	PB170218BL		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	TCLP Pesticide
Prep Method:	SW3510	Prep Date 10/22/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25 17:30	PB170218
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/22/25 17:30	PB170218
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/22/25 17:30	PB170218
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/22/25 17:30	PB170218
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25 17:30	PB170218
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25 17:30	PB170218
57-74-9	Chlordane	0.25	U	1	0.088	0.25	0.50	ug/L	10/22/25 17:30	PB170218
SURROGATES										
2051-24-3	Decachlorobiphenyl	22.3			30 - 135		112%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.1			44 - 124		101%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/17/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/17/25
Client Sample ID:	PIBLK-PD090647.D		SDG No.:	Q3399
Lab Sample ID:	I.BLK-PD090647.D		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:		Prep Date:		TCLP Pesticide

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/17/25	pd101725
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/17/25	pd101725
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/17/25	pd101725
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/17/25	pd101725
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/17/25	pd101725
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/17/25	pd101725
57-74-9	Chlordane	0.25	U	1	0.088	0.25	0.50	ug/L	10/17/25	pd101725
SURROGATES										
2051-24-3	Decachlorobiphenyl	21.3			30 - 135		106%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.0			44 - 124		100%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/22/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/22/25
Client Sample ID:	PIBLK-PD090753.D		SDG No.:	Q3399
Lab Sample ID:	I.BLK-PD090753.D		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:		Prep Date:		TCLP Pesticide

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/22/25	PD102225
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/22/25	PD102225
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/22/25	PD102225
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25	PD102225
57-74-9	Chlordane	0.25	U	1	0.088	0.25	0.50	ug/L	10/22/25	PD102225
SURROGATES										
2051-24-3	Decachlorobiphenyl	25.1			30 - 135		126%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.7			44 - 124		104%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/22/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/22/25
Client Sample ID:	PIBLK-PD090763.D		SDG No.:	Q3399
Lab Sample ID:	I.BLK-PD090763.D		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:		Prep Date		TCLP Pesticide

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/22/25	PD102225
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/22/25	PD102225
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/22/25	PD102225
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25	PD102225
57-74-9	Chlordane	0.25	U	1	0.088	0.25	0.50	ug/L	10/22/25	PD102225
SURROGATES										
2051-24-3	Decachlorobiphenyl	21.0			30 - 135		105%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.6			44 - 124		103%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/22/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/22/25
Client Sample ID:	PIBLK-PD090772.D		SDG No.:	Q3399
Lab Sample ID:	I.BLK-PD090772.D		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol:	10000 uL	Test:
Prep Method:		Prep Date:		TCLP Pesticide

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.010	U	1	0.0037	0.010	0.050	ug/L	10/22/25	PD102225
76-44-8	Heptachlor	0.010	U	1	0.0027	0.010	0.050	ug/L	10/22/25	PD102225
1024-57-3	Heptachlor epoxide	0.025	U	1	0.0096	0.025	0.050	ug/L	10/22/25	PD102225
72-20-8	Endrin	0.010	U	1	0.0032	0.010	0.050	ug/L	10/22/25	PD102225
72-43-5	Methoxychlor	0.025	U	1	0.011	0.025	0.050	ug/L	10/22/25	PD102225
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25	PD102225
57-74-9	Chlordane	0.25	U	1	0.088	0.25	0.50	ug/L	10/22/25	PD102225
SURROGATES										
2051-24-3	Decachlorobiphenyl	24.0			30 - 135		120%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	20.5			44 - 124		103%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	
Client Sample ID:	PB170218BS		SDG No.:	Q3399
Lab Sample ID:	PB170218BS		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	TCLP Pesticide
Prep Method:	SW3510	Prep Date 10/22/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	0.45		1	0.0037	0.010	0.050	ug/L	10/22/25 17:44	PB170218
76-44-8	Heptachlor	0.52		1	0.0027	0.010	0.050	ug/L	10/22/25 17:44	PB170218
1024-57-3	Heptachlor epoxide	0.45		1	0.0096	0.025	0.050	ug/L	10/22/25 17:44	PB170218
72-20-8	Endrin	0.46		1	0.0032	0.010	0.050	ug/L	10/22/25 17:44	PB170218
72-43-5	Methoxychlor	0.48		1	0.011	0.025	0.050	ug/L	10/22/25 17:44	PB170218
8001-35-2	Toxaphene	0.50	U	1	0.17	0.50	1.00	ug/L	10/22/25 17:44	PB170218
57-74-9	Chlordane	0.25	U	1	0.088	0.25	0.50	ug/L	10/22/25 17:44	PB170218
SURROGATES										
2051-24-3	Decachlorobiphenyl	23.4			30 - 135		117%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	17.3			44 - 124		86%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/20/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/20/25
Client Sample ID:	COMP-ROLLOFFMS		SDG No.:	Q3399
Lab Sample ID:	Q3399-02MS		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	100 mL	Final Vol:	10000 uL	Test:
Prep Method:	SW3510	Prep Date	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	5.00		1	0.037	0.10	0.50	ug/L	10/22/25 18:38	PB170218
76-44-8	Heptachlor	5.90		1	0.027	0.10	0.50	ug/L	10/22/25 18:38	PB170218
1024-57-3	Heptachlor epoxide	5.00		1	0.096	0.25	0.50	ug/L	10/22/25 18:38	PB170218
72-20-8	Endrin	5.10		1	0.032	0.10	0.50	ug/L	10/22/25 18:38	PB170218
72-43-5	Methoxychlor	5.40		1	0.11	0.25	0.50	ug/L	10/22/25 18:38	PB170218
8001-35-2	Toxaphene	5.00	U	1	1.70	5.00	10.0	ug/L	10/22/25 18:38	PB170218
57-74-9	Chlordane	2.50	U	1	0.88	2.50	5.00	ug/L	10/22/25 18:38	PB170218
SURROGATES										
2051-24-3	Decachlorobiphenyl	21.2			30 - 135		106%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.4			44 - 124		107%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/20/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/20/25
Client Sample ID:	COMP-ROLLOFFMSD		SDG No.:	Q3399
Lab Sample ID:	Q3399-02MSD		Matrix:	TCLP
Analytical Method:	8081B		% Solid:	0
Sample Wt/Vol:	100 mL	Final Vol:	10000 uL	Test:
Prep Method:	SW3510	Prep Date	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
58-89-9	gamma-BHC (Lindane)	4.90		1	0.037	0.10	0.50	ug/L	10/22/25 18:52	PB170218
76-44-8	Heptachlor	5.80		1	0.027	0.10	0.50	ug/L	10/22/25 18:52	PB170218
1024-57-3	Heptachlor epoxide	5.40		1	0.096	0.25	0.50	ug/L	10/22/25 18:52	PB170218
72-20-8	Endrin	5.00		1	0.032	0.10	0.50	ug/L	10/22/25 18:52	PB170218
72-43-5	Methoxychlor	5.30		1	0.11	0.25	0.50	ug/L	10/22/25 18:52	PB170218
8001-35-2	Toxaphene	5.00	U	1	1.70	5.00	10.0	ug/L	10/22/25 18:52	PB170218
57-74-9	Chlordane	2.50	U	1	0.88	2.50	5.00	ug/L	10/22/25 18:52	PB170218
SURROGATES										
2051-24-3	Decachlorobiphenyl	20.7			30 - 135		104%	SPK: 20		
877-09-8	Tetrachloro-m-xylene	21.1			44 - 124		106%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	9.07	9.07	9.06	9.07	9.07	9.07	8.97	9.17
Endrin	6.57	6.57	6.57	6.57	6.57	6.57	6.47	6.67
gamma-BHC (Lindane)	4.32	4.33	4.33	4.32	4.32	4.32	4.22	4.42
Heptachlor	4.92	4.92	4.92	4.92	4.92	4.92	4.82	5.02
Heptachlor epoxide	5.69	5.69	5.68	5.68	5.68	5.68	5.58	5.78
Methoxychlor	7.49	7.49	7.49	7.49	7.49	7.49	7.39	7.59
Tetrachloro-m-xylene	3.54	3.55	3.55	3.54	3.54	3.54	3.44	3.64

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>
Instrument ID:	<u>ECD_D</u>	Calibration Date(s):	<u>10/17/2025</u> <u>10/17/2025</u>
		Calibration Times:	<u>11:34</u> <u>12:28</u>

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 100 = <u>PD090650.D</u>	RT 075 = <u>PD090651.D</u>
RT 050 = <u>PD090652.D</u>	RT 025 = <u>PD090653.D</u>	RT 005 = <u>PD090654.D</u>

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	8.06	8.06	8.06	8.06	8.06	8.06	7.96	8.16
Endrin	5.78	5.78	5.78	5.78	5.78	5.78	5.68	5.88
gamma-BHC (Lindane)	3.72	3.72	3.72	3.72	3.72	3.72	3.62	3.82
Heptachlor	4.08	4.08	4.08	4.07	4.08	4.07	3.97	4.17
Heptachlor epoxide	4.86	4.87	4.86	4.86	4.86	4.86	4.76	4.96
Methoxychlor	6.75	6.75	6.75	6.74	6.75	6.74	6.64	6.84
Tetrachloro-m-xylene	2.87	2.88	2.88	2.87	2.87	2.87	2.77	2.97

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Instrument ID: ECD_D

Calibration Date(s): 10/17/2025 10/17/2025

Calibration Times: 11:34 12:28

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	4361180000	4315270000	4506230000	4707060000	5500070000	4677960000	10
Endrin	5035940000	4962910000	4938250000	4711180000	5120260000	4953710000	3
gamma-BHC (Lindane)	7185880000	6905860000	6799650000	6417980000	6684750000	6798820000	4
Heptachlor	5757910000	5597510000	5545540000	5288860000	5757970000	5589560000	3
Heptachlor epoxide	5795570000	5708170000	5718680000	5584980000	6137920000	5789060000	4
Methoxychlor	2050060000	2015340000	2078640000	2094570000	2389310000	2125590000	7
Tetrachloro-m-xylene	3010860000	2951180000	3030100000	3124300000	3454390000	3114160000	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_D

Contract: ICES02
SDG NO.: Q3399

Calibration Date(s): 10/17/2025 10/17/2025
Calibration Times: 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD090650.D</u>	CF 075 =	<u>PD090651.D</u>		
CF 050 =		<u>PD090652.D</u>	CF 025 =	<u>PD090653.D</u>	CF 005 =	<u>PD090654.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	28856800000	28247200000	29308000000	30057200000	34987600000	30291400000	9
Endrin	36902000000	37827400000	39400900000	40693000000	47646500000	40494000000	11
gamma-BHC (Lindane)	44964800000	44604900000	45972800000	46808300000	53893500000	47248900000	8
Heptachlor	41272300000	41810600000	43239300000	44502900000	52189000000	44602800000	10
Heptachlor epoxide	38203400000	38891500000	40672900000	42374400000	53052500000	42639000000	14
Methoxychlor	16116900000	16316400000	17198900000	17735100000	19771700000	17427800000	8
Tetrachloro-m-xylene	27903400000	27689100000	28617400000	29726500000	34949400000	29777100000	10

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: ICES02

Lab Code: ACE
SDG NO.: Q3399

Instrument ID: ECD_D
Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR1
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.71	4.61	4.81	185217000
		2	5.24	5.14	5.34	185731000
		3	5.94	5.84	6.04	758191000
		4	6.03	5.93	6.13	917583000
		5	6.87	6.77	6.97	148478000
Toxaphene	500	1	6.24	6.14	6.34	32564400
		2	6.44	6.34	6.54	48621800
		3	7.14	7.04	7.24	85710600
		4	7.56	7.46	7.66	95737400
		5	7.92	7.82	8.02	60480500

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance
Contract: ICES02

Lab Code: ACE
SDG NO.: Q3399

Instrument ID: ECD_D
Date(s) Analyzed: 10/17/2025 10/17/2025

GC Column: ZB-MR2
ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.90	3.80	4.00	1210370000
		2	4.48	4.38	4.58	1263780000
		3	5.12	5.02	5.22	3771830000
		4	5.18	5.08	5.28	3251870000
		5	6.08	5.98	6.18	1460510000
Toxaphene	500	1	5.47	5.37	5.57	263702000
		2	5.64	5.54	5.74	176560000
		3	6.75	6.65	6.85	658883000
		4	7.19	7.09	7.29	495104000
		5	7.32	7.22	7.42	246033000

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 14:59

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 14:59

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.08	3.98	4.18	0.01
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Methoxychlor	6.75	6.75	6.65	6.85	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ICES02
Lab Code: ACE
SDG NO.: Q3399
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No.: CCAL01
Date Analyzed: 10/22/2025
Lab Sample No.: PSTDCCC050
Data File : PD090755.D
Time Analyzed: 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.068	8.964	9.164	46.320	50.000	-7.4
Endrin	6.570	6.468	6.668	47.990	50.000	-4.0
gamma-BHC (Lindane)	4.326	4.225	4.425	47.990	50.000	-4.0
Heptachlor	4.924	4.823	5.023	57.070	50.000	14.1
Heptachlor epoxide	5.685	5.584	5.784	49.030	50.000	-1.9
Methoxychlor	7.490	7.388	7.588	49.890	50.000	-0.2
Tetrachloro-m-xylene	3.545	3.445	3.645	46.510	50.000	-7.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090755.D **Time Analyzed:** 14:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.062	7.962	8.162	54.910	50.000	9.8
Endrin	5.781	5.681	5.881	48.490	50.000	-3.0
gamma-BHC (Lindane)	3.722	3.623	3.823	49.500	50.000	-1.0
Heptachlor	4.074	3.975	4.175	49.470	50.000	-1.1
Heptachlor epoxide	4.864	4.764	4.964	47.790	50.000	-4.4
Methoxychlor	6.745	6.645	6.845	52.980	50.000	6.0
Tetrachloro-m-xylene	2.874	2.775	2.975	47.350	50.000	-5.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 17:17

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
gamma-BHC (Lindane)	4.32	4.33	4.23	4.43	0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 17:17

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.08	3.98	4.18	0.01
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Methoxychlor	6.75	6.75	6.65	6.85	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.067	8.964	9.164	44.170	50.000	-11.7
Endrin	6.569	6.468	6.668	46.250	50.000	-7.5
gamma-BHC (Lindane)	4.324	4.225	4.425	47.530	50.000	-4.9
Heptachlor	4.923	4.823	5.023	55.620	50.000	11.2
Heptachlor epoxide	5.685	5.584	5.784	46.870	50.000	-6.3
Methoxychlor	7.489	7.388	7.588	46.180	50.000	-7.6
Tetrachloro-m-xylene	3.544	3.445	3.645	46.320	50.000	-7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090764.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.061	7.962	8.162	48.710	50.000	-2.6
Endrin	5.780	5.681	5.881	46.650	50.000	-6.7
gamma-BHC (Lindane)	3.722	3.623	3.823	48.880	50.000	-2.2
Heptachlor	4.074	3.975	4.175	48.710	50.000	-2.6
Heptachlor epoxide	4.864	4.764	4.964	46.510	50.000	-7.0
Methoxychlor	6.745	6.645	6.845	48.070	50.000	-3.9
Tetrachloro-m-xylene	2.874	2.775	2.975	47.070	50.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025

10/17/2025

Continuing Calib Time: 19:19

Initial Calibration Time(s): 11:34

12:28

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.06	8.96	9.16	-0.01
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
gamma-BHC (Lindane)	4.32	4.33	4.23	4.43	0.01
Heptachlor	4.92	4.92	4.82	5.02	0.00
Heptachlor epoxide	5.69	5.68	5.58	5.78	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Continuing Calib Date: 10/22/2025

Initial Calibration Date(s): 10/17/2025 10/17/2025

Continuing Calib Time: 19:19

Initial Calibration Time(s): 11:34 12:28

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene	2.87	2.88	2.78	2.98	0.01
gamma-BHC (Lindane)	3.72	3.72	3.62	3.82	0.00
Heptachlor	4.07	4.08	3.98	4.18	0.01
Heptachlor epoxide	4.86	4.86	4.76	4.96	0.00
Endrin	5.78	5.78	5.68	5.88	0.00
Methoxychlor	6.74	6.75	6.65	6.85	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090773.D **Time Analyzed:** 19:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.066	8.964	9.164	48.090	50.000	-3.8
Endrin	6.568	6.468	6.668	49.010	50.000	-2.0
gamma-BHC (Lindane)	4.324	4.225	4.425	49.170	50.000	-1.7
Heptachlor	4.923	4.823	5.023	57.890	50.000	15.8
Heptachlor epoxide	5.685	5.584	5.784	49.640	50.000	-0.7
Methoxychlor	7.489	7.388	7.588	49.960	50.000	-0.1
Tetrachloro-m-xylene	3.544	3.445	3.645	47.760	50.000	-4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3399
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/17/2025 10/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/22/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090773.D **Time Analyzed:** 19:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.061	7.962	8.162	55.760	50.000	11.5
Endrin	5.780	5.681	5.881	49.660	50.000	-0.7
gamma-BHC (Lindane)	3.722	3.623	3.823	50.230	50.000	0.5
Heptachlor	4.073	3.975	4.175	49.520	50.000	-1.0
Heptachlor epoxide	4.863	4.764	4.964	48.340	50.000	-3.3
Methoxychlor	6.744	6.645	6.845	53.040	50.000	6.1
Tetrachloro-m-xylene	2.874	2.775	2.975	48.360	50.000	-3.3

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.066	8.970	9.170	22.860	20.000	14.3
Tetrachloro-m-xylene	3.544	3.490	3.590	21.710	20.000	8.6
alpha-BHC	3.994	3.940	4.040	9.050	10.000	-9.5
beta-BHC	4.512	4.460	4.560	10.070	10.000	0.7
gamma-BHC (Lindane)	4.324	4.270	4.370	9.330	10.000	-6.7
Endrin	6.568	6.500	6.640	49.200	50.000	-1.6
4,4'-DDT	7.016	6.950	7.090	100.270	100.000	0.3
Methoxychlor	7.489	7.420	7.560	236.810	250.000	-5.3

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090648.D Date Analyzed: 10/17/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:07

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	21.400	20.000	7.0
Tetrachloro-m-xylene	2.875	2.820	2.930	21.220	20.000	6.1
alpha-BHC	3.386	3.340	3.440	10.300	10.000	3.0
beta-BHC	4.019	3.970	4.070	10.460	10.000	4.6
gamma-BHC (Lindane)	3.723	3.670	3.770	10.340	10.000	3.4
Endrin	5.781	5.710	5.850	47.940	50.000	-4.1
4,4'-DDT	6.174	6.100	6.240	95.670	100.000	-4.3
Methoxychlor	6.745	6.670	6.820	201.830	250.000	-19.3

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090754.D Date Analyzed: 10/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 14:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.070	8.970	9.170	22.040	20.000	10.2
Tetrachloro-m-xylene	3.544	3.490	3.590	20.920	20.000	4.6
alpha-BHC	3.994	3.940	4.040	8.790	10.000	-12.1
beta-BHC	4.513	4.460	4.560	10.210	10.000	2.1
gamma-BHC (Lindane)	4.325	4.270	4.380	9.200	10.000	-8.0
Endrin	6.570	6.500	6.640	48.180	50.000	-3.6
4,4'-DDT	7.018	6.950	7.090	100.910	100.000	0.9
Methoxychlor	7.491	7.420	7.560	235.430	250.000	-5.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/17/2025 10/17/2025

Client Sample No. (PEM): PEM - PD090754.D Date Analyzed: 10/22/2025

Lab Sample No.(PEM): PEM Time Analyzed: 14:45

PEM COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.062	7.960	8.160	25.380	20.000	26.9
Tetrachloro-m-xylene	2.873	2.820	2.920	21.110	20.000	5.6
alpha-BHC	3.385	3.330	3.440	10.320	10.000	3.2
beta-BHC	4.017	3.970	4.070	10.570	10.000	5.7
gamma-BHC (Lindane)	3.721	3.670	3.770	10.640	10.000	6.4
Endrin	5.779	5.710	5.850	49.190	50.000	-1.6
4,4'-DDT	6.173	6.100	6.240	103.470	100.000	3.5
Methoxychlor	6.745	6.670	6.820	217.530	250.000	-13.0

Analytical Sequence

Client: ICE Service Group, Inc.	SDG No.: Q3399
Project: NWIRP Northrup Grumman Site – Bethpage	Instrument ID: ECD_D
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 10/17/2025 10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/17/2025	10:53	PD090647.D	9.07	3.55
PEM	PEM	10/17/2025	11:07	PD090648.D	9.07	3.54
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	9.07	3.54
PSTDICC100	PSTDICC100	10/17/2025	11:34	PD090650.D	9.07	3.54
PSTDICC075	PSTDICC075	10/17/2025	11:48	PD090651.D	9.07	3.55
PSTDICC050	PSTDICC050	10/17/2025	12:01	PD090652.D	9.06	3.55
PSTDICC025	PSTDICC025	10/17/2025	12:15	PD090653.D	9.07	3.54
PSTDICC005	PSTDICC005	10/17/2025	12:28	PD090654.D	9.07	3.54
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	9.07	3.55
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	9.07	3.55
IBLK	IBLK	10/22/2025	14:32	PD090753.D	9.08	3.55
PEM	PEM	10/22/2025	14:45	PD090754.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/22/2025	14:59	PD090755.D	9.07	3.55
COMP-ROLLOFF	Q3399-02	10/22/2025	15:54	PD090759.D	9.07	3.55
IBLK	IBLK	10/22/2025	17:03	PD090763.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/22/2025	17:17	PD090764.D	9.07	3.54
PB170218BL	PB170218BL	10/22/2025	17:30	PD090765.D	9.07	3.54
PB170218BS	PB170218BS	10/22/2025	17:44	PD090766.D	9.07	3.54
PB170172TB	PB170172TB	10/22/2025	17:58	PD090767.D	9.07	3.54
COMP-ROLLOFFMS	Q3399-02MS	10/22/2025	18:38	PD090770.D	9.07	3.55
COMP-ROLLOFFMSD	Q3399-02MSD	10/22/2025	18:52	PD090771.D	9.07	3.54
IBLK	IBLK	10/22/2025	19:06	PD090772.D	9.07	3.54
PSTDCCC050	PSTDCCC050	10/22/2025	19:19	PD090773.D	9.07	3.54

Analytical Sequence

Client: ICE Service Group, Inc.	SDG No.: Q3399
Project: NWIRP Northrup Grumman Site – Bethpage	Instrument ID: ECD_D
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 10/17/2025 10/17/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/17/2025	10:53	PD090647.D	8.06	2.88
PEM	PEM	10/17/2025	11:07	PD090648.D	8.06	2.88
RESCHK	RESCHK	10/17/2025	11:20	PD090649.D	8.06	2.87
PSTDICC100	PSTDICC100	10/17/2025	11:34	PD090650.D	8.06	2.87
PSTDICC075	PSTDICC075	10/17/2025	11:48	PD090651.D	8.06	2.88
PSTDICC050	PSTDICC050	10/17/2025	12:01	PD090652.D	8.06	2.88
PSTDICC025	PSTDICC025	10/17/2025	12:15	PD090653.D	8.06	2.87
PSTDICC005	PSTDICC005	10/17/2025	12:28	PD090654.D	8.06	2.87
PCHLORICC500	PCHLORICC500	10/17/2025	13:09	PD090657.D	8.06	2.87
PTOXICC500	PTOXICC500	10/17/2025	14:17	PD090662.D	8.06	2.88
IBLK	IBLK	10/22/2025	14:32	PD090753.D	8.06	2.87
PEM	PEM	10/22/2025	14:45	PD090754.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/22/2025	14:59	PD090755.D	8.06	2.87
COMP-ROLLOFF	Q3399-02	10/22/2025	15:54	PD090759.D	8.06	2.87
IBLK	IBLK	10/22/2025	17:03	PD090763.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/22/2025	17:17	PD090764.D	8.06	2.87
PB170218BL	PB170218BL	10/22/2025	17:30	PD090765.D	8.06	2.87
PB170218BS	PB170218BS	10/22/2025	17:44	PD090766.D	8.06	2.87
PB170172TB	PB170172TB	10/22/2025	17:58	PD090767.D	8.06	2.87
COMP-ROLLOFFMS	Q3399-02MS	10/22/2025	18:38	PD090770.D	8.06	2.87
COMP-ROLLOFFMSD	Q3399-02MSD	10/22/2025	18:52	PD090771.D	8.06	2.88
IBLK	IBLK	10/22/2025	19:06	PD090772.D	8.06	2.87
PSTDCCC050	PSTDCCC050	10/22/2025	19:19	PD090773.D	8.06	2.87

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-ROLLOFFMS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3399-02MS

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	5.20	3.8
	2	6.75	6.70	6.80	5.40	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	5.00	4.1
	2	3.72	3.67	3.77	4.80	
Heptachlor	1	4.92	4.87	4.97	5.90	18.5
	2	4.08	4.03	4.13	4.90	
Heptachlor epoxide	1	5.69	5.64	5.74	5.00	2
	2	4.86	4.81	4.91	4.90	
Endrin	1	6.57	6.52	6.62	5.10	2
	2	5.78	5.73	5.83	5.00	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

COMP-ROLLOFFMSD

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: Q3399-02MSD

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column: (2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	5.20	1.9
	2	6.74	6.69	6.79	5.30	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	4.90	4.2
	2	3.72	3.67	3.77	4.70	
Heptachlor	1	4.92	4.87	4.97	5.80	18.9
	2	4.08	4.03	4.13	4.80	
Heptachlor epoxide	1	5.68	5.63	5.73	5.40	1.9
	2	4.86	4.81	4.91	5.30	
Endrin	1	6.57	6.52	6.62	5.00	0
	2	5.78	5.73	5.83	5.00	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB170218BS

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab Sample ID: PB170218BS

Date(s) Analyzed: 10/22/2025 10/22/2025

Instrument ID (1): ECD_D

Instrument ID (2): ECD_D

GC Column: (1): ZB-MR1

ID: 0.32 (mm)

GC Column:(2): ZB-MR2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
Methoxychlor	1	7.49	7.44	7.54	0.46	4.4
	2	6.75	6.70	6.80	0.48	
gamma-BHC (Lindane)	1	4.33	4.28	4.38	0.44	1.5
	2	3.72	3.67	3.77	0.45	
Heptachlor	1	4.92	4.87	4.97	0.52	13.2
	2	4.07	4.02	4.12	0.46	
Heptachlor epoxide	1	5.69	5.64	5.74	0.45	0.2
	2	4.86	4.81	4.91	0.45	
Endrin	1	6.57	6.52	6.62	0.45	3.2
	2	5.78	5.73	5.83	0.46	