

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3399

Analytical Method: 8081B

Client: ICE Service Group, Inc.

Datafile : PD090766.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB170218BS (Column 1)	gamma-BHC (Lindane)	0.5	0.44	ug/L	88				59	134		
	Heptachlor	0.5	0.52	ug/L	105				54	130		
	Heptachlor epoxide	0.5	0.45	ug/L	90				61	133		
	Endrin	0.5	0.45	ug/L	89				60	138		
	Methoxychlor	0.5	0.46	ug/L	91				54	145		
PB170218BS (Column 2)	gamma-BHC (Lindane)	0.5	0.45	ug/L	90				59	134		
	Heptachlor	0.5	0.46	ug/L	92				54	130		
	Heptachlor epoxide	0.5	0.45	ug/L	90				61	133		
	Endrin	0.5	0.46	ug/L	92				60	138		
	Methoxychlor	0.5	0.48	ug/L	95				54	145		