

Surrogate Summary

SW-846

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170206BL	PB170206BL	2-Fluorophenol	150	125	84		35	115
		Phenol-d6	150	129	86		34	127
		Nitrobenzene-d5	100	86.7	87		37	122
		2-Fluorobiphenyl	100	84.7	85		44	115
		2,4,6-Tribromophenol	150	152	102		39	132
		Terphenyl-d14	100	105	105		54	127
PB170206BS	PB170206BS	2-Fluorophenol	150	108	72		35	115
		Phenol-d6	150	112	75		34	127
		Nitrobenzene-d5	100	78.4	78		37	122
		2-Fluorobiphenyl	100	77.8	78		44	115
		2,4,6-Tribromophenol	150	121	81		39	132
		Terphenyl-d14	100	95.7	96		54	127
Q3399-01	COMP-ROLLOFF	2-Fluorophenol	150	93.5	62		35	115
		Phenol-d6	150	89.9	60		34	127
		Nitrobenzene-d5	100	67.4	67		37	122
		2-Fluorobiphenyl	100	70.9	71		44	115
		2,4,6-Tribromophenol	150	107	71		39	132
		Terphenyl-d14	100	48.6	49	*	54	127
Q3399-01MS	COMP-ROLLOFFMS	2-Fluorophenol	150	90.7	60		35	115
		Phenol-d6	150	85.1	57		34	127
		Nitrobenzene-d5	100	66.5	66		37	122
		2-Fluorobiphenyl	100	64.8	65		44	115
		2,4,6-Tribromophenol	150	104	69		39	132
		Terphenyl-d14	100	50.5	51	*	54	127
Q3399-01MSD	COMP-ROLLOFFMSD	2-Fluorophenol	150	95.5	64		35	115
		Phenol-d6	150	89.5	60		34	127
		Nitrobenzene-d5	100	68.4	68		37	122
		2-Fluorobiphenyl	100	67.5	68		44	115
		2,4,6-Tribromophenol	150	105	70		39	132
		Terphenyl-d14	100	53.4	53	*	54	127