

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3399

Analytical Method: SW8270E

Client: ICE Service Group, Inc.

DataFile: BF144045.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3399-01MS		Client Sample ID: COMP-ROLLOFFMS									
Benzaldehyde	1800	0	920	ug/Kg	51				10	161	
Phenol	1800	0	1600	ug/Kg	89				34	121	
bis(2-Chloroethyl)ether	1800	0	1600	ug/Kg	89				31	120	
2-Chlorophenol	1800	0	1700	ug/Kg	94				34	121	
2-Methylphenol	1800	0	1600	ug/Kg	89				32	122	
2,2-oxybis(1-Chloropropane)	1800	0	1300	ug/Kg	72				33	131	
Acetophenone	1800	0	2000	ug/Kg	111				33	115	
3+4-Methylphenols	1800	0	1600	ug/Kg	89				34	119	
N-Nitroso-di-n-propylamine	1800	0	1500	ug/Kg	83				36	120	
Hexachloroethane	1800	0	1800	ug/Kg	100				28	117	
Nitrobenzene	1800	0	1800	ug/Kg	100				34	122	
Isophorone	1800	0	1700	ug/Kg	94				30	122	
2-Nitrophenol	1800	0	2000	ug/Kg	111				36	123	
2,4-Dimethylphenol	1800	0	1900	ug/Kg	106				30	127	
bis(2-Chloroethoxy)methane	1800	0	1700	ug/Kg	94				36	121	
2,4-Dichlorophenol	1800	0	1800	ug/Kg	100				40	122	
Naphthalene	1800	0	1800	ug/Kg	100				35	123	
4-Chloroaniline	1800	0	780	ug/Kg	43				17	106	
Hexachlorobutadiene	1800	0	2000	ug/Kg	111				32	123	
Caprolactam	1800	0	1600	ug/Kg	89				46	117	
4-Chloro-3-methylphenol	1800	0	1700	ug/Kg	94				45	122	
2-Methylnaphthalene	1800	0	1600	ug/Kg	89				38	122	
Hexachlorocyclopentadiene	3700	0	2500	ug/Kg	68				43	112	
2,4,6-Trichlorophenol	1800	0	1800	ug/Kg	100				39	126	
2,4,5-Trichlorophenol	1800	0	1800	ug/Kg	100				41	124	
1,1-Biphenyl	1800	0	2000	ug/Kg	111				40	117	
2-Chloronaphthalene	1800	0	1800	ug/Kg	100				41	114	
2-Nitroaniline	1800	0	1800	ug/Kg	100				44	127	
Dimethylphthalate	1800	0	1800	ug/Kg	100				48	124	
Acenaphthylene	1800	0	2000	ug/Kg	111				32	132	
2,6-Dinitrotoluene	1800	0	2100	ug/Kg	117				46	124	
3-Nitroaniline	1800	0	1300	ug/Kg	72				33	119	
Acenaphthene	1800	0	1800	ug/Kg	100				40	123	
2,4-Dinitrophenol	3700	0	2800	ug/Kg	76				15	130	
4-Nitrophenol	3700	0	3100	ug/Kg	84				30	132	
Dibenzofuran	1800	0	1800	ug/Kg	100				44	120	
2,4-Dinitrotoluene	1800	0	2000	ug/Kg	111				48	126	
Diethylphthalate	1800	0	1800	ug/Kg	100				50	124	
4-Chlorophenyl-phenylether	1800	0	1800	ug/Kg	100				45	121	
Fluorene	1800	0	1800	ug/Kg	100				43	125	
4-Nitroaniline	1800	0	1700	ug/Kg	94				35	115	
4,6-Dinitro-2-methylphenol	1800	0	2000	ug/Kg	111				29	132	
N-Nitrosodiphenylamine	1800	0	1900	ug/Kg	106				38	127	
4-Bromophenyl-phenylether	1800	0	1900	ug/Kg	106				46	124	
Hexachlorobenzene	1800	0	2000	ug/Kg	111				45	122	
Atrazine	1800	0	2000	ug/Kg	111				47	127	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3399

Analytical Method: SW8270E

Client: ICE Service Group, Inc.

DataFile: BF144045.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Pentachlorophenol	3700	0	3500	ug/Kg	95				25	133	
Phenanthrene	1800	0	1800	ug/Kg	100				50	121	
Anthracene	1800	0	1900	ug/Kg	106				47	123	
Carbazole	1800	0	1900	ug/Kg	106				50	123	
Di-n-butylphthalate	1800	0	1900	ug/Kg	106				51	128	
Fluoranthene	1800	0	2000	ug/Kg	111				50	127	
Pyrene	1800	0	1400	ug/Kg	78				47	127	
Butylbenzylphthalate	1800	0	1700	ug/Kg	94				48	132	
3,3-Dichlorobenzidine	1800	0	1600	ug/Kg	89				22	121	
Benzo(a)anthracene	1800	0	1800	ug/Kg	100				49	126	
Chrysene	1800	0	1900	ug/Kg	106				50	124	
bis(2-Ethylhexyl)phthalate	1800	0	2000	ug/Kg	111				51	133	
Di-n-octyl phthalate	1800	0	2200	ug/Kg	122				45	140	
Benzo(b)fluoranthene	1800	0	1900	ug/Kg	106				45	132	
Benzo(k)fluoranthene	1800	0	1800	ug/Kg	100				47	132	
Benzo(a)pyrene	1800	0	1900	ug/Kg	106				45	129	
Indeno(1,2,3-cd)pyrene	1800	0	1500	ug/Kg	83				45	133	
Dibenz(a,h)anthracene	1800	0	1600	ug/Kg	89				45	134	
Benzo(g,h,i)perylene	1800	0	1400	ug/Kg	78				43	134	
1,2,4,5-Tetrachlorobenzene	1800	0	2100	ug/Kg	117				37	119	
1,4-Dioxane	1800	0	1600	ug/Kg	89				70	130	
2,3,4,6-Tetrachlorophenol	1800	0	1900	ug/Kg	106				44	125	