

Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
 Method File : 8270-BG102425.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 03 18:16:26 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BG064569.D 5 =BG064570.D 10 =BG064571.D 20 =BG064572.D 40 =BG064573.D 50 =BG064576.D 60 =BG064574.D 80 =BG064575.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.570	0.581	0.524	0.525	0.474	0.559	0.460	0.528	8.85	
3)	Pyridine	1.630	1.608	1.552	1.669	1.471	1.733	1.427	1.584	6.84	
4)	n-Nitrosodimet...		0.912	0.807	0.900	0.780	0.921	0.776	0.850	8.08	
5) S	2-Fluorophenol	1.098	1.207	1.155	1.292	1.139	1.318	1.132	1.192	7.09	
6)	Aniline	2.203	2.320	2.131	2.386	2.090	2.440	2.004	2.225	7.28	
7) S	Phenol-d6	1.469	1.654	1.566	1.771	1.597	1.864	1.524	1.635	8.55	
8)	2-Chlorophenol	1.051	1.207	1.197	1.373	1.225	1.461	1.216	1.247	10.65	
9)	Benzaldehyde		0.942	0.994	0.940	0.836	0.981	0.662	0.893	14.09	
10) C	Phenol	1.665	1.820	1.748	1.944	1.761	2.027	1.693	1.808	7.36	
11)	bis(2-Chloroet...	1.307	1.409	1.292	1.383	1.273	1.443	1.198	1.329	6.47	
12)	1,3-Dichlorobe...	1.435	1.516	1.394	1.520	1.342	1.542	1.299	1.435	6.62	
13) C	1,4-Dichlorobe...	1.431	1.546	1.442	1.534	1.381	1.561	1.323	1.460	6.21	
14)	1,2-Dichlorobe...	1.371	1.518	1.364	1.483	1.316	1.520	1.283	1.408	6.97	
15)	Benzyl Alcohol		1.281	1.263	1.425	1.306	1.533	1.283	1.348	7.97	
16)	2,2'-oxybis(1-...	3.421	3.604	3.396	3.733	3.301	3.802	3.154	3.487	6.74	
17)	2-Methylphenol	1.103	1.225	1.178	1.273	1.159	1.345	1.120	1.201	7.22	
18)	Hexachloroethane	0.468	0.535	0.505	0.541	0.496	0.557	0.483	0.512	6.41	
19) P	n-Nitroso-di-n...	1.044	1.038	1.141	1.129	1.227	1.115	1.289	1.057	1.130	7.93
20)	3+4-Methylphenols		1.652	1.581	1.818	1.612	1.895	1.571	1.688	8.04	
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.535	0.565	0.528	0.584	0.516	0.590	0.509	0.547	6.04	
23) S	Nitrobenzene-d5	0.239	0.286	0.314	0.377	0.356	0.403	0.356	0.333	17.02	
24)	Nitrobenzene	0.274	0.312	0.344	0.410	0.377	0.438	0.380	0.362	15.64	
25)	Isophorone	0.751	0.796	0.771	0.869	0.764	0.891	0.751	0.799	7.25	
26) C	2-Nitrophenol	0.076	0.095	0.110	0.137	0.135	0.152	0.144	0.121	23.20	
27)	2,4-Dimethylph...	0.236	0.294	0.285	0.335	0.297	0.343	0.290	0.297	11.93	
28)	bis(2-Chloroet...	0.452	0.464	0.443	0.490	0.426	0.493	0.417	0.455	6.48	
29) C	2,4-Dichloroph...	0.243	0.299	0.315	0.360	0.327	0.373	0.322	0.320	13.34	
30)	1,2,4-Trichlor...	0.338	0.369	0.348	0.391	0.349	0.399	0.349	0.363	6.44	
31)	Naphthalene	1.083	1.144	1.084	1.177	1.036	1.201	1.037	1.109	5.96	
32)	Benzoic acid		0.142	0.164	0.223	0.215	0.260	0.237	0.207	21.74	
33)	4-Chloroaniline	0.390	0.456	0.411	0.459	0.410	0.486	0.405	0.431	8.29	
34) C	Hexachlorobuta...	0.266	0.291	0.282	0.299	0.271	0.306	0.273	0.284	5.33	
35)	Caprolactam		0.112	0.114	0.134	0.118	0.142	0.116	0.123	9.85	
36) C	4-Chloro-3-met...	0.323	0.381	0.361	0.426	0.376	0.454	0.370	0.384	11.18	
37)	2-Methylnaphth...	0.764	0.858	0.797	0.865	0.772	0.888	0.755	0.814	6.74	
38)	1-Methylnaphth...	0.798	0.849	0.776	0.853	0.752	0.878	0.735	0.806	6.85	

Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
 Method File : 8270-BG102425.M


39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.657	0.684	0.648	0.702	0.633	0.717	0.657	0.671	4.55	
41) P	Hexachlorocycl...	0.268	0.279	0.324	0.297	0.340	0.305	0.302		8.97	
42) S	2,4,6-Tribromo...	0.221	0.266	0.272	0.318	0.292	0.346	0.306	0.289	14.04	
43) C	2,4,6-Trichlor...	0.294	0.333	0.375	0.423	0.399	0.456	0.408	0.384	14.41	
44)	2,4,5-Trichlor...	0.339	0.425	0.439	0.484	0.443	0.517	0.454	0.443	12.48	
45) S	2-Fluorobiphenyl	1.310	1.389	1.306	1.384	1.235	1.384	1.228	1.319	5.24	
46)	1,1'-Biphenyl	1.456	1.450	1.401	1.489	1.326	1.498	1.333	1.422	4.98	
47)	2-Chloronaphth...	1.065	1.082	1.053	1.132	1.013	1.145	1.023	1.073	4.71	
48)	2-Nitroaniline	0.201	0.234	0.277	0.348	0.343	0.404	0.367	0.311	24.06	
49)	Acenaphthylene	1.667	1.726	1.647	1.786	1.593	1.827	1.604	1.693	5.30	
50)	Dimethylphthalate	1.389	1.570	1.488	1.641	1.462	1.699	1.462	1.530	7.21	
51)	2,6-Dinitrotol...	0.127	0.180	0.193	0.247	0.244	0.283	0.254	0.218	24.69	
52) C	Acenaphthene	1.126	1.178	1.115	1.223	1.093	1.261	1.101	1.157	5.63	
53)	3-Nitroaniline	0.180	0.211	0.252	0.294	0.281	0.325	0.297	0.263	19.70	
54) P	2,4-Dinitrophenol	0.086	0.110	0.137	0.139	0.165	0.154	0.132		22.11	
55)	Dibenzofuran	1.881	1.972	1.826	1.963	1.734	1.988	1.725	1.870	5.96	
56) P	4-Nitrophenol	0.191	0.213	0.261	0.250	0.283	0.267	0.244		14.40	
57)	2,4-Dinitrotol...	0.223	0.269	0.300	0.385	0.373	0.434	0.396	0.340	22.67	
58)	Fluorene	1.471	1.559	1.426	1.531	1.356	1.549	1.329	1.460	6.38	
59)	2,3,4,6-Tetrac...	0.312	0.385	0.402	0.466	0.444	0.508	0.449	0.424	15.04	
60)	Diethylphthalate	1.417	1.645	1.514	1.657	1.486	1.718	1.502	1.563	7.06	
61)	4-Chlorophenyl...	0.800	0.864	0.792	0.866	0.763	0.888	0.763	0.820	6.38	
62)	4-Nitroaniline	0.194	0.252	0.286	0.327	0.310	0.355	0.324	0.293	18.55	
63)	Azobenzene	1.362	1.504	1.377	1.538	1.338	1.564	1.353	1.434	6.80	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...	0.055	0.064	0.083	0.083	0.097	0.093	0.079		21.00	
66) c	n-Nitrosodiphe...	0.510	0.551	0.525	0.583	0.518	0.594	0.500	0.540	6.79	
67)	4-Bromophenyl-...	0.218	0.236	0.230	0.258	0.229	0.264	0.224	0.237	7.28	
68)	Hexachlorobenzene	0.254	0.266	0.262	0.292	0.260	0.302	0.256	0.270	6.99	
69)	Atrazine	0.215	0.227	0.249	0.245	0.219	0.261	0.211	0.232	8.27	
70) C	Pentachlorophenol	0.132	0.152	0.180	0.172	0.199	0.177	0.169		13.91	
71)	Phenanthrene	1.084	1.133	1.073	1.162	1.020	1.179	1.005	1.094	6.17	
72)	Anthracene	1.087	1.144	1.101	1.187	1.041	1.204	1.025	1.113	6.17	
73)	Carbazole	0.982	1.037	0.982	1.031	0.910	1.028	0.897	0.981	5.90	
74)	Di-n-butylphth...	1.011	1.156	1.179	1.240	1.084	1.198	1.075	1.134	7.11	
75) C	Fluoranthene	1.385	1.463	1.380	1.440	1.273	1.403	1.247	1.370	5.91	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.392	0.475	0.490	0.416	0.533	0.383	0.448		13.40	
78)	Pyrene	1.287	1.331	1.284	1.411	1.233	1.406	1.196	1.307	6.24	
79) S	Terphenyl-d14	1.074	1.120	1.066	1.142	0.987	1.102	0.928	1.060	7.21	
80)	Butylbenzylphth...	0.284	0.359	0.388	0.450	0.408	0.467	0.410	0.395	15.43	
81)	Benzo(a)anthra...	1.294	1.355	1.292	1.416	1.253	1.430	1.243	1.326	5.71	
82)	3,3'-Dichlorob...	0.418	0.428	0.477	0.424	0.486	0.424	0.443		6.88	
83)	Chrysene	1.249	1.291	1.229	1.344	1.188	1.359	1.173	1.262	5.76	
84)	Bis(2-ethylhex...	0.484	0.555	0.581	0.671	0.585	0.667	0.588	0.590	10.95	
85) c	Di-n-octyl pht...	0.871	0.932	1.087	0.976	1.117	0.986	0.995		9.32	

Method Path : Z:\svoasrv\HPCHEM1\BNA_G\Methods\
Method File : 8270-BG102425.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.383	1.491	1.410	1.568	1.396	1.642	1.410	1.471	6.80
88)	Benzo(b)fluora...	1.138	1.246	1.169	1.322	1.173	1.353	1.183	1.226	6.77
89)	Benzo(k)fluora...	1.175	1.260	1.206	1.285	1.151	1.364	1.149	1.227	6.51
90) C	Benzo(a)pyrene	1.061	1.144	1.093	1.202	1.067	1.246	1.079	1.127	6.42
91)	Dibenzo(a,h)an...	1.120	1.223	1.158	1.268	1.130	1.324	1.143	1.195	6.55
92)	Benzo(g,h,i)pe...	1.080	1.162	1.089	1.213	1.078	1.280	1.092	1.142	6.95

(#) = Out of Range


2-Nitrophenol



Response = 1.495e-001 * Amt - 2.340e-002
 Coef of Det (r^2) = 0.995585 Curve Fit: wlr(1/a)
 Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Calibration Table Last Updated: Fri Oct 24 16:40:26 2025

Benzoic acid

Response Ratio



$$\text{Response} = 2.554\text{e-}001 * \text{Amt} - 8.844\text{e-}002$$


Coef of Det (r^2) = 0.993384 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M

Calibration Table Last Updated: Fri Oct 24 16:40:26 2025

2-Nitroaniline

Response Ratio



$$\text{Response} = 3.863\text{e-}001 * \text{Amt} - 6.228\text{e-}002$$


Coef of Det (r^2) = 0.991893 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M

Calibration Table Last Updated: Fri Oct 24 16:40:26 2025

2,6-Dinitrotoluene

Response Ratio




$$\text{Response} = 2.705\text{e-}001 * \text{Amt} - 4.319\text{e-}002$$

Coef of Det (r^2) = 0.992850 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M

Calibration Table Last Updated: Fri Oct 24 16:40:26 2025


2,4-Dinitrophenol



Response = 1.666e-001 * Amt - 4.640e-002
 Coef of Det (r^2) = 0.992196 Curve Fit: wlr(1/a)
 Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
 Calibration Table Last Updated: Fri Oct 24 16:40:26 2025

2,4-Dinitrotoluene

Response Ratio




$$\text{Response} = 4.169\text{e-}001 * \text{Amt} - 6.360\text{e-}002$$

Coef of Det (r^2) = 0.992833 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M

Calibration Table Last Updated: Fri Oct 24 16:40:26 2025

4,6-Dinitro-2-methylphenol



Response = $9.922 \times 10^{-2} \times \text{Amt} - 2.692 \times 10^{-2}$
Coef of Det (r^2) = 0.993360 Curve Fit: wlr(1/a)
Method Name: Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG102425.M
Calibration Table Last Updated: Fri Oct 24 16:40:26 2025