

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3399</u>
Instrument ID:	<u>BNA_G</u>	Calibration Date/Time:	<u>11/03/2025 23:27</u>
Lab File ID:	<u>BG064685.D</u>	Init. Calib. Date(s):	<u>10/24/2025 10/24/2025</u>
EPA Sample No.:	<u>SSTDCCC040EC</u>	Init. Calib. Time(s):	<u>09:00 15:32</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.192	1.219		2.3	50.0
Benzaldehyde	0.893	0.893		0.0	50.0
Phenol-d6	1.635	1.692		3.5	50.0
Phenol	1.808	1.851		2.4	50.0
bis(2-Chloroethyl) ether	1.329	1.297		-2.4	50.0
2-Chlorophenol	1.247	1.371		9.9	50.0
2-Methylphenol	1.201	1.233		2.7	50.0
2,2-oxybis(1-Chloropropane)	3.487	3.201		-8.2	50.0
Acetophenone	0.547	0.547		0.0	50.0
3+4-Methylphenols	1.688	1.731		2.5	50.0
n-Nitroso-di-n-propylamine	1.130	1.089	0.050	-3.6	50.0
Nitrobenzene-d5	0.333	0.413		24.0	50.0
Hexachloroethane	0.512	0.554		8.2	50.0
Nitrobenzene	0.362	0.425		17.4	50.0
Isophorone	0.799	0.784		-1.9	50.0
2-Nitrophenol	0.121	0.193		59.5	50.0
2,4-Dimethylphenol	0.297	0.313		5.4	50.0
bis(2-Chloroethoxy) methane	0.455	0.448		-1.5	50.0
2,4-Dichlorophenol	0.320	0.380		18.8	50.0
Naphthalene	1.109	1.162		4.8	50.0
4-Chloroaniline	0.431	0.443		2.8	50.0
Hexachlorobutadiene	0.284	0.337		18.7	50.0
Caprolactam	0.123	0.110		-10.6	50.0
4-Chloro-3-methylphenol	0.384	0.430		12.0	50.0
2-Methylnaphthalene	0.814	0.877		7.7	50.0
Hexachlorocyclopentadiene	0.302	0.384	0.050	27.2	50.0
2,4,6-Trichlorophenol	0.384	0.469		22.1	50.0
2-Fluorobiphenyl	1.319	1.442		9.3	50.0
2,4,5-Trichlorophenol	0.443	0.507		14.4	50.0
1,1-Biphenyl	1.422	1.491		4.9	50.0
2-Chloronaphthalene	1.073	1.143		6.5	50.0
2-Nitroaniline	0.311	0.410		31.8	50.0
Dimethylphthalate	1.530	1.618		5.8	50.0
Acenaphthylene	1.693	1.806		6.7	50.0
2,6-Dinitrotoluene	0.218	0.317		45.4	50.0
3-Nitroaniline	0.263	0.343		30.4	50.0
Acenaphthene	1.157	1.250		8.0	50.0
2,4-Dinitrophenol	0.132	0.140	0.050	6.1	50.0
4-Nitrophenol	0.244	0.219	0.050	-10.2	50.0

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3399</u>
Instrument ID:	<u>BNA_G</u>	Calibration Date/Time:	<u>11/03/2025 23:27</u>
Lab File ID:	<u>BG064685.D</u>	Init. Calib. Date(s):	<u>10/24/2025 10/24/2025</u>
EPA Sample No.:	<u>SSTDCCC040EC</u>	Init. Calib. Time(s):	<u>09:00 15:32</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.870	1.992		6.5	50.0
2,4-Dinitrotoluene	0.340	0.486		42.9	50.0
Diethylphthalate	1.563	1.606		2.8	50.0
4-Chlorophenyl-phenylether	0.820	0.941		14.8	50.0
Fluorene	1.460	1.562		7.0	50.0
4-Nitroaniline	0.293	0.360		22.9	50.0
4,6-Dinitro-2-methylphenol	0.079	0.115		45.6	50.0
n-Nitrosodiphenylamine	0.540	0.558		3.3	50.0
2,4,6-Tribromophenol	0.289	0.386		33.6	50.0
4-Bromophenyl-phenylether	0.237	0.269		13.5	50.0
Hexachlorobenzene	0.270	0.310		14.8	50.0
Atrazine	0.232	0.240		3.4	50.0
Pentachlorophenol	0.169	0.085		-49.7	50.0
Phenanthrene	1.094	1.127		3.0	50.0
Anthracene	1.113	1.150		3.3	50.0
Carbazole	0.981	0.988		0.7	50.0
Di-n-butylphthalate	1.135	1.155		1.8	50.0
Fluoranthene	1.370	1.467		7.1	50.0
Pyrene	1.307	1.267		-3.1	50.0
Terphenyl-d14	1.060	1.087		2.5	50.0
Butylbenzylphthalate	0.395	0.435		10.1	50.0
3,3-Dichlorobenzidine	0.443	0.479		8.1	50.0
Benzo (a) anthracene	1.326	1.382		4.2	50.0
Chrysene	1.262	1.316		4.3	50.0
Bis(2-ethylhexyl)phthalate	0.590	0.625		5.9	50.0
Di-n-octyl phthalate	0.995	1.082		8.7	50.0
Benzo (b) fluoranthene	1.226	1.263		3.0	50.0
Benzo (k) fluoranthene	1.227	1.262		2.9	50.0
Benzo (a) pyrene	1.127	1.168		3.6	50.0
Indeno (1,2,3-cd) pyrene	1.471	1.514		2.9	50.0
Dibenzo (a,h) anthracene	1.195	1.241		3.8	50.0
Benzo (g,h,i) perylene	1.142	1.175		2.9	50.0
1,2,4,5-Tetrachlorobenzene	0.671	0.779		16.1	50.0
1,4-Dioxane	0.528	0.425		-19.5	50.0
2,3,4,6-Tetrachlorophenol	0.424	0.476		12.3	50.0

All other compounds must meet a minimum RRF of 0.010.