

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG102425

Review By	rahul	Review On	10/27/2025 10:16:00 AM		
Supervise By	Jagrut	Supervise On	10/27/2025 12:25:10 PM		
SubDirectory	BG102425	HP Acquire Method	BNA_G	HP Processing Method	BG102425
STD. NAME		STD REF.#			
Tune/Reschk		SP6875			
Initial Calibration Stds		SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864			
CCC		SP6860			
Internal Standard/PEM		S13179,10ul/1000ul sample			
ICV/I.BLK		SP6872			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064568.D	24 Oct 2025 8:20		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BG064569.D	24 Oct 2025 9:00		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BG064570.D	24 Oct 2025 9:41		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BG064571.D	24 Oct 2025 11:02		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BG064572.D	24 Oct 2025 12:23	Compound #32 kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BG064573.D	24 Oct 2025 13:03	Compound #26,48,51,54,57,65 are kept on LR	RC/JU	Ok,M
7	SSTDICC060	SSTDICC060	BG064574.D	24 Oct 2025 13:44		RC/JU	Ok,M
8	SSTDICC080	SSTDICC080	BG064575.D	24 Oct 2025 14:24		RC/JU	Ok,M
9	SSTDICC050	SSTDICC050	BG064576.D	24 Oct 2025 15:32		RC/JU	Ok,M
10	SSTDICV040	ICVBG102425	BG064577.D	24 Oct 2025 16:47	compound #77 failed in the ICV.	RC/JU	Ok
11	PB170222BL	PB170222BL	BG064578.D	24 Oct 2025 17:28	Analyzed for contamination check	RC/JU	Not Ok
12	Q3405-02	PAD-10-20-25-B	BG064579.D	24 Oct 2025 18:08		RC/JU	Ok
13	Q3405-03	PAD-10-20-25-C	BG064580.D	24 Oct 2025 18:49		RC/JU	Ok
14	Q3405-04	SW-10-20-25	BG064581.D	24 Oct 2025 19:30		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_G

Daily Analysis Runlog For Sequence/QC Batch ID # BG110325

Review By	rahul	Review On	11/3/2025 3:19:29 PM
Supervise By		Supervise On	
SubDirectory	BG110325	HP Acquire Method	BNA_G
		HP Processing Method	BG102425
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13181,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BG064671.D	3 Nov 2025 13:54		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BG064672.D	3 Nov 2025 14:35		RC/JU	Ok,NR
3	PB170206BL	PB170206BL	BG064673.D	3 Nov 2025 15:16	Good for DOD Projects	RC/JU	Ok
4	PB170206BS	PB170206BS	BG064674.D	3 Nov 2025 15:57		RC/JU	Ok,NR
5	PB170216BL	PB170216BL	BG064675.D	3 Nov 2025 16:38		RC/JU	Ok
6	PB170216BS	PB170216BS	BG064676.D	3 Nov 2025 17:19		RC/JU	Ok,NR
7	PB170172TB	PB170172TB	BG064677.D	3 Nov 2025 18:00		RC/JU	Ok
8	Q3399-02	COMP-ROLLOFF	BG064678.D	3 Nov 2025 18:41		RC/JU	Ok
9	Q3368-06	OUTSIDE-DUMPSTER	BG064679.D	3 Nov 2025 19:22		RC/JU	Ok
10	Q3368-06MS	OUTSIDE-DUMPSTER	BG064680.D	3 Nov 2025 20:02		RC/JU	Ok,NR
11	Q3368-06MSD	OUTSIDE-DUMPSTER	BG064681.D	3 Nov 2025 20:43		RC/JU	Ok,NR
12	Q3399-01	COMP-ROLLOFF	BG064682.D	3 Nov 2025 21:24		RC/JU	Ok
13	Q3399-01MS	COMP-ROLLOFFMS	BG064683.D	3 Nov 2025 22:05		RC/JU	Ok,NR
14	Q3399-01MSD	COMP-ROLLOFFMSD	BG064684.D	3 Nov 2025 22:46		RC/JU	Ok,NR
15	SSTDCCC040	SSTDCCC040EC	BG064685.D	3 Nov 2025 23:27		RC/JU	Ok,NR

M : Manual Integration