

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3399

Client ID : SSTDCCC040

Date Analyzed: 11/03/2025

Lab File ID: BG064672.D

Time Analyzed: 14:35

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	31906	8.207	126610	11.04	99069	14.83
UPPER LIMIT	63812	8.707	253220	11.539	198138	15.334
LOWER LIMIT	15953	7.707	63305	10.539	49534.5	14.334
EPA SAMPLE NO.						
01 PB170206BL	33303	8.21	120360	11.04	97938	14.83
02 PB170206BS	31473	8.21	125369	11.04	100193	14.83
03 COMP-ROLLOFF	32053	8.21	122752	11.04	95326	14.83
04 COMP-ROLLOFFMS	30518	8.21	124719	11.03	97711	14.83
05 COMP-ROLLOFFMSD	29939	8.21	118752	11.03	97631	14.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.