

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3399

Client ID: SSTDCCC040

Date Analyzed: 11/03/2025

Lab File ID: BG064672.D

Time Analyzed: 14:35

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	239125	17.572	279624	21.85	296512	25.187
UPPER LIMIT	478250	18.072	559248	22.35	593024	25.687
LOWER LIMIT	119563	17.072	139812	21.35	148256	24.687
EPA SAMPLE NO.						
01 PB170206BL	255703	17.57	321187	21.84	341510	25.17
02 PB170206BS	249048	17.57	291284	21.85	304737	25.18
03 COMP-ROLLOFF	251613	17.57	301190	21.84	331540	25.17
04 COMP-ROLLOFFMS	238341	17.57	278646	21.84	306904	25.17
05 COMP-ROLLOFFMSD	238127	17.57	275359	21.84	303393	25.18

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.