

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP101525

Review By	rahul	Review On	10/16/2025 4:25:02 PM
Supervise By	Jagrut	Supervise On	10/16/2025 4:29:28 PM
SubDirectory	BP101525	HP Acquire Method	BNA_P
		HP Processing Method	BP101525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13178,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025971.D	15 Oct 2025 11:04		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP025972.D	15 Oct 2025 11:45	Method failed for Comp. #56	RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP025973.D	15 Oct 2025 12:26		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP025974.D	15 Oct 2025 13:07	Comp. #56 not detected in 10ppm	RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP025975.D	15 Oct 2025 13:48		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BP025976.D	15 Oct 2025 14:29	Comp. #41,54 on Linear Regression	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP025977.D	15 Oct 2025 15:51		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BP025978.D	15 Oct 2025 17:13		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BP025979.D	15 Oct 2025 18:35		RC/JU	Ok
10	SSTDICV040	ICVBP101525	BP025980.D	15 Oct 2025 22:00	ICV failed for comp. #09,77	RC/JU	Ok,M
11	PB170067BL	PB170067BL	BP025981.D	15 Oct 2025 22:41	Analyzed for contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP102225

Review By	rahul	Review On	10/28/2025 9:34:38 AM
Supervise By	mohammad	Supervise On	10/28/2025 9:37:45 AM
SubDirectory	BP102225	HP Acquire Method	BNA_P
		HP Processing Method	BP101525

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13179,10ul/1000ul sample
ICV/I.BLK	SP6872
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP026000.D	22 Oct 2025 15:52		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP026001.D	22 Oct 2025 16:33		RC/JU	Ok,M
3	PB170216BL	PB170216BL	BP026002.D	22 Oct 2025 17:14		RC/JU	Ok
4	PB170216BS	PB170216BS	BP026003.D	22 Oct 2025 17:55		RC/JU	Ok,M
5	PB170172TB	PB170172TB	BP026004.D	22 Oct 2025 18:36		RC/JU	Ok
6	Q3368-06	OUTSIDE-DUMPSTER	BP026005.D	22 Oct 2025 19:17		RC/JU	Ok
7	Q3368-06MS	OUTSIDE-DUMPSTER	BP026006.D	22 Oct 2025 19:58		RC/JU	Ok
8	Q3368-06MSD	OUTSIDE-DUMPSTER	BP026007.D	22 Oct 2025 20:39		RC/JU	Ok
9	Q3399-02	COMP-ROLLOFF	BP026008.D	22 Oct 2025 21:20		RC/JU	Ok
10	Q3402-03	41-LA	BP026009.D	22 Oct 2025 22:01		RC/JU	Ok
11	Q3402-06	41-RA	BP026010.D	22 Oct 2025 22:42		RC/JU	Ok
12	Q3402-09	41-LB	BP026011.D	22 Oct 2025 23:22		RC/JU	Ok
13	Q3402-12	41-RB	BP026012.D	23 Oct 2025 00:03		RC/JU	Ok
14	Q3397-04	TP-7-WC	BP026013.D	23 Oct 2025 00:44		RC/JU	Ok
15	Q3409-01	EXCAVATION-AREA-A	BP026014.D	23 Oct 2025 01:24		RC/JU	Ok
16	Q3409-02	EXCAVATION-AREA-B	BP026015.D	23 Oct 2025 02:05		RC/JU	Ok
17	Q3368-07	PRESS-SLUDGE	BP026016.D	23 Oct 2025 02:46		RC/JU	Ok

M : Manual Integration