

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3399

Analytical Method: 8270E

Client: ICE Service Group, Inc.

DataFile: BP026003.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170216BS	Pyridine	50	35.7	ug/L	71				10	153	
	1,4-Dichlorobenzene	50	42.6	ug/L	85				29	112	
	2-Methylphenol	50	50.9	ug/L	102				30	117	
	3+4-Methylphenols	50	49.5	ug/L	99				29	110	
	Hexachloroethane	50	42.5	ug/L	85				21	115	
	Nitrobenzene	50	45.3	ug/L	91				45	121	
	Hexachlorobutadiene	50	41.2	ug/L	82				22	124	
	2,4,6-Trichlorophenol	50	45.6	ug/L	91				50	125	
	2,4,5-Trichlorophenol	50	47.0	ug/L	94				53	123	
	2,4-Dinitrotoluene	50	47.2	ug/L	94				57	128	
	Hexachlorobenzene	50	42.2	ug/L	84				53	125	
	Pentachlorophenol	100	94.1	ug/L	94				35	138	