

Surrogate Summary

SW-846

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170172TB	PB170172TB	2-Fluorophenol	150	111	74		19	119
		Phenol-d6	150	104	69		10	130
		Nitrobenzene-d5	100	89.1	89		44	120
		2-Fluorobiphenyl	100	82.7	83		44	119
		2,4,6-Tribromophenol	150	141	94		43	140
		Terphenyl-d14	100	81.4	81		50	134
PB170216BL	PB170216BL	2-Fluorophenol	150	104	69		19	119
		Phenol-d6	150	95.4	64		10	130
		Nitrobenzene-d5	100	83.4	83		44	120
		2-Fluorobiphenyl	100	74.1	74		44	119
		2,4,6-Tribromophenol	150	126	84		43	140
		Terphenyl-d14	100	75.9	76		50	134
PB170216BS	PB170216BS	2-Fluorophenol	150	116	77		19	119
		Phenol-d6	150	122	81		10	130
		Nitrobenzene-d5	100	92.8	93		44	120
		2-Fluorobiphenyl	100	80.9	81		44	119
		2,4,6-Tribromophenol	150	150	100		43	140
		Terphenyl-d14	100	82.9	83		50	134
Q3368-06MS	OUTSIDE-DUMPSTERMS	2-Fluorophenol	150	134	90		19	119
		Phenol-d6	150	123	82		10	130
		Nitrobenzene-d5	100	112	112		44	120
		2-Fluorobiphenyl	100	96.2	96		44	119
		2,4,6-Tribromophenol	150	184	123		43	140
		Terphenyl-d14	100	71.0	71		50	134
Q3368-06MSD	OUTSIDE-DUMPSTERMSD	2-Fluorophenol	150	135	90		19	119
		Phenol-d6	150	123	82		10	130
		Nitrobenzene-d5	100	113	113		44	120
		2-Fluorobiphenyl	100	96.1	96		44	119
		2,4,6-Tribromophenol	150	178	119		43	140
		Terphenyl-d14	100	70.5	71		50	134
Q3399-02	COMP-ROLLOFF	2-Fluorophenol	150	113	76		19	119
		Phenol-d6	150	93.3	62		10	130
		Nitrobenzene-d5	100	101	101		44	120
		2-Fluorobiphenyl	100	87.1	87		44	119
		2,4,6-Tribromophenol	150	143	95		43	140
		Terphenyl-d14	100	89.2	89		50	134