

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3399

Analytical Method: 8270E

Client: ICE Service Group, Inc.

DataFile: BG064676.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170216BS	Pyridine	50	26.0	ug/L	52				10	153	
	1,4-Dichlorobenzene	50	42.0	ug/L	84				29	112	
	2-Methylphenol	50	43.4	ug/L	87				30	117	
	3+4-Methylphenols	50	41.8	ug/L	84				29	110	
	Hexachloroethane	50	42.1	ug/L	84				21	115	
	Nitrobenzene	50	47.0	ug/L	94				45	121	
	Hexachlorobutadiene	50	43.6	ug/L	87				22	124	
	2,4,6-Trichlorophenol	50	51.2	ug/L	102				50	125	
	2,4,5-Trichlorophenol	50	48.9	ug/L	98				53	123	
	2,4-Dinitrotoluene	50	51.4	ug/L	103				57	128	
	Hexachlorobenzene	50	45.5	ug/L	91				53	125	
	Pentachlorophenol	100	100	ug/L	100				35	138	