

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3399

Client ID : SSTDCCC040

Date Analyzed: 11/03/2025

Lab File ID: BG064672.D

Time Analyzed: 14:35

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	31906	8.207	126610	11.04	99069	14.83
UPPER LIMIT	63812	8.707	253220	11.539	198138	15.334
LOWER LIMIT	15953	7.707	63305	10.539	49534.5	14.334
EPA SAMPLE NO.						
01 PB170216BL	35344	8.21	128956	11.03	100482	14.83
02 PB170216BS	36665	8.21	145545	11.03	116501	14.84
03 PB170172TB	38354	8.21	143910	11.04	111139	14.83
04 OUTSIDE-DUMPSTERMS	30368	8.21	117215	11.03	90360	14.83
05 OUTSIDE-DUMPSTERMSD	29562	8.21	113027	11.03	89136	14.83
06 COMP-ROLLOFF	28350	8.21	106669	11.03	82537	14.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.