

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BG064671.D
Instrument ID: BNA_G

Contract: ICES02
SDG NO.: Q3399
DFTPP Injection Date: 11/03/2025
DFTPP Injection Time: 13:54

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.2 (0.7) 1
69	Mass 69 relative abundance	31.2
70	Less than 2.0% of mass 69	0.2 (0.7) 1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	6.3
441	Present, but less than mass 443	16.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	20.3 (20.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BG064672.D	11/03/2025	14:35
PB170216BL	PB170216BL	BG064675.D	11/03/2025	16:38
PB170216BS	PB170216BS	BG064676.D	11/03/2025	17:19
PB170172TB	PB170172TB	BG064677.D	11/03/2025	18:00
COMP-ROLLOFF	Q3399-02	BG064678.D	11/03/2025	18:41
OUTSIDE-DUMPSTERMS	Q3368-06MS	BG064680.D	11/03/2025	20:02
OUTSIDE-DUMPSTERMSD	Q3368-06MSD	BG064681.D	11/03/2025	20:43
SSTDCCC040EC	SSTDCCC040	BG064685.D	11/03/2025	23:27