

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3399</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/24/2025</u> <u>09:16</u>
Lab File ID:	<u>VX048339.D</u>	Init. Calib. Date(s):	<u>10/20/2025</u> <u>10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59</u> <u>13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.444	0.627		41.22	20
1,1-Dichloroethene	0.525	0.572		8.95	20
2-Butanone	0.304	0.271		-10.85	20
Carbon Tetrachloride	0.599	0.550		-8.18	20
Chloroform	1.123	1.261		12.29	20
Benzene	1.409	1.480		5.04	20
1,2-Dichloroethane	0.520	0.592		13.85	20
Trichloroethene	0.372	0.374		0.54	20
Tetrachloroethene	0.373	0.372		-0.27	20
Chlorobenzene	1.125	1.142	0.3	1.51	20
1,2-Dichloroethane-d4	0.701	0.754		7.56	20
Dibromofluoromethane	0.341	0.277		-18.77	20
Toluene-d8	1.196	1.133		-5.27	20
4-Bromofluorobenzene	0.455	0.434		-4.61	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.