

## Report of Analysis

Client: ICE Service Group, Inc.  
Project: NWIRP Northrup Grumman Site – Bethpage, NY  
Client Sample ID: COMP-ROLLOFF  
Lab Sample ID: Q3399-04  
Analytical Method: 8260D  
Sample Wt/Vol: 5 mL

Level: LOW  
Final Vol: 5000 uL

Date Collected: 10/20/25  
Date Received: 10/20/25  
SDG No.: Q3399  
Matrix: TCLP  
% Solid: 0  
Test: TCLP VOA

| CAS Number                | Parameter              | Conc.             | Qua. | DF | MDL      | LOD  | LOQ / CRQL | Units   | Date Ana.      | BatchID  |
|---------------------------|------------------------|-------------------|------|----|----------|------|------------|---------|----------------|----------|
| <b>TARGETS</b>            |                        |                   |      |    |          |      |            |         |                |          |
| 75-01-4                   | Vinyl Chloride         | 0.75              | UQ   | 1  | 0.26     | 0.75 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 75-35-4                   | 1,1-Dichloroethene     | 0.75              | U    | 1  | 0.23     | 0.75 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 78-93-3                   | 2-Butanone             | 2.50              | U    | 1  | 0.98     | 2.50 | 25.0       | ug/L    | 10/24/25 11:03 | VX102425 |
| 56-23-5                   | Carbon Tetrachloride   | 0.50              | U    | 1  | 0.25     | 0.50 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 67-66-3                   | Chloroform             | 0.50              | U    | 1  | 0.25     | 0.50 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 71-43-2                   | Benzene                | 0.50              | U    | 1  | 0.15     | 0.50 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 107-06-2                  | 1,2-Dichloroethane     | 0.75              | U    | 1  | 0.22     | 0.75 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 79-01-6                   | Trichloroethene        | 0.75              | U    | 1  | 0.090    | 0.75 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 127-18-4                  | Tetrachloroethene      | 0.50              | U    | 1  | 0.23     | 0.50 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| 108-90-7                  | Chlorobenzene          | 0.50              | U    | 1  | 0.12     | 0.50 | 5.00       | ug/L    | 10/24/25 11:03 | VX102425 |
| <b>SURROGATES</b>         |                        |                   |      |    |          |      |            |         |                |          |
| 17060-07-0                | 1,2-Dichloroethane-d4  | 54.5              |      |    | 81 - 118 |      | 109%       | SPK: 50 |                |          |
| 1868-53-7                 | Dibromofluoromethane   | 39.9              |      |    | 80 - 119 |      | 80%        | SPK: 50 |                |          |
| 2037-26-5                 | Toluene-d8             | 47.1              |      |    | 89 - 112 |      | 94%        | SPK: 50 |                |          |
| 460-00-4                  | 4-Bromofluorobenzene   | 50.4              |      |    | 85 - 114 |      | 101%       | SPK: 50 |                |          |
| <b>INTERNAL STANDARDS</b> |                        |                   |      |    |          |      |            |         |                |          |
|                           |                        | <b>Area Count</b> |      |    |          |      |            |         |                |          |
| 363-72-4                  | Pentafluorobenzene     | 137000            |      |    |          |      |            |         |                |          |
| 540-36-3                  | 1,4-Difluorobenzene    | 271000            |      |    |          |      |            |         |                |          |
| 3114-55-4                 | Chlorobenzene-d5       | 270000            |      |    |          |      |            |         |                |          |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4 | 124000            |      |    |          |      |            |         |                |          |

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products