

LAB CHRONICLE

OrderID:	Q3399	OrderDate:	10/20/2025 1:23:08 PM
Client:	ICE Service Group, Inc.	Project:	NWIRP Northrup Grumman Site – Bethpage, NY
Contact:	Dennis D. Morgan, II	Location:	D41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3399-03	COMP-ROLLOFF	SOIL	VOC-TCLVOA-10	8260D	10/20/25		10/21/25	10/20/25
Q3399-04	COMP-ROLLOFF	TCLP	TCLP VOA	8260D	10/20/25		10/24/25	10/20/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3399

Client: ICE Service Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:



SAMPLE DATA

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: COMP-ROLLOFF
Lab Sample ID: Q3399-04
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected: 10/20/25
Date Received: 10/20/25
SDG No.: Q3399
Matrix: TCLP
% Solid: 0
Test: TCLP VOA

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-01-4	Vinyl Chloride	0.75	UQ	1	0.26	0.75	5.00	ug/L	10/24/25 11:03	VX102425
75-35-4	1,1-Dichloroethene	0.75	U	1	0.23	0.75	5.00	ug/L	10/24/25 11:03	VX102425
78-93-3	2-Butanone	2.50	U	1	0.98	2.50	25.0	ug/L	10/24/25 11:03	VX102425
56-23-5	Carbon Tetrachloride	0.50	U	1	0.25	0.50	5.00	ug/L	10/24/25 11:03	VX102425
67-66-3	Chloroform	0.50	U	1	0.25	0.50	5.00	ug/L	10/24/25 11:03	VX102425
71-43-2	Benzene	0.50	U	1	0.15	0.50	5.00	ug/L	10/24/25 11:03	VX102425
107-06-2	1,2-Dichloroethane	0.75	U	1	0.22	0.75	5.00	ug/L	10/24/25 11:03	VX102425
79-01-6	Trichloroethene	0.75	U	1	0.090	0.75	5.00	ug/L	10/24/25 11:03	VX102425
127-18-4	Tetrachloroethene	0.50	U	1	0.23	0.50	5.00	ug/L	10/24/25 11:03	VX102425
108-90-7	Chlorobenzene	0.50	U	1	0.12	0.50	5.00	ug/L	10/24/25 11:03	VX102425
SURROGATES										
17060-07-0	1,2-Dichloroethane-d4	54.5			81 - 118		109%	SPK: 50		
1868-53-7	Dibromofluoromethane	39.9			80 - 119		80%	SPK: 50		
2037-26-5	Toluene-d8	47.1			89 - 112		94%	SPK: 50		
460-00-4	4-Bromofluorobenzene	50.4			85 - 114		101%	SPK: 50		
INTERNAL STANDARDS										
		Area Count								
363-72-4	Pentafluorobenzene	137000								
540-36-3	1,4-Difluorobenzene	271000								
3114-55-4	Chlorobenzene-d5	270000								
3855-82-1	1,4-Dichlorobenzene-d4	124000								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3399-04	COMP-ROLLOFF	1,2-Dichloroethane-d4	50	54.5	109		81	118
		Dibromofluoromethane	50	39.9	80		80	119
		Toluene-d8	50	47.1	94		89	112
		4-Bromofluorobenzene	50	50.4	101		85	114

Surrogate Summary

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
VX1024WBL01	VX1024WBL01	1,2-Dichloroethane-d4	50	52.2	104		81	118
		Dibromofluoromethane	50	39.0	78	*	80	119
		Toluene-d8	50	43.1	86	*	89	112
		4-Bromofluorobenzene	50	42.7	85		85	114
VX1024WBS01	VX1024WBS01	1,2-Dichloroethane-d4	50	51.3	103		81	118
		Dibromofluoromethane	50	41.6	83		80	119
		Toluene-d8	50	49.6	99		89	112
		4-Bromofluorobenzene	50	48.3	97		85	114
VX1024WBSD0	VX1024WBSD01	1,2-Dichloroethane-d4	50	53.1	106		81	118
		Dibromofluoromethane	50	40.4	81		80	119
		Toluene-d8	50	49.6	99		89	112
		4-Bromofluorobenzene	50	48.4	97		85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048341.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1024WBS01	Vinyl chloride	20	27.8	ug/L	139		*	58	137	
	1,1-Dichloroethene	20	20.7	ug/L	104			71	131	
	2-Butanone	100	84.6	ug/L	85			56	143	
	Carbon Tetrachloride	20	18.5	ug/L	93			72	136	
	Chloroform	20	20.8	ug/L	104			79	124	
	Benzene	20	20.2	ug/L	101			79	120	
	1,2-Dichloroethane	20	21.0	ug/L	105			73	128	
	Trichloroethene	20	19.2	ug/L	96			79	123	
	Tetrachloroethene	20	20.9	ug/L	104			74	129	
	Chlorobenzene	20	19.8	ug/L	99			82	118	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048342.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1024WBSD01	Vinyl chloride	20	26.6	ug/L	133	4		58	137	20
	1,1-Dichloroethene	20	19.5	ug/L	98	6		71	131	20
	2-Butanone	100	95.2	ug/L	95	11		56	143	20
	Carbon Tetrachloride	20	19.0	ug/L	95	2		72	136	20
	Chloroform	20	20.8	ug/L	104	0		79	124	20
	Benzene	20	20.1	ug/L	101	0		79	120	20
	1,2-Dichloroethane	20	21.6	ug/L	108	3		73	128	20
	Trichloroethene	20	19.6	ug/L	98	2		79	123	20
	Tetrachloroethene	20	21.2	ug/L	106	2		74	129	20
	Chlorobenzene	20	19.7	ug/L	99	0		82	118	20

VOLATILE METHOD BLANK SUMMARY

Client ID

VX1024WBL01

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab File ID: VX048340.D

Lab Sample ID: VX1024WBL01

Date Analyzed: 10/24/2025

Time Analyzed: 09:48

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1024WBS01	VX1024WBS01	VX048341.D	10/24/2025
VX1024WBSD01	VX1024WBSD01	VX048342.D	10/24/2025
COMP-ROLLOFF	Q3399-04	VX048343.D	10/24/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab File ID: VX048226.D

BFB Injection Date: 10/20/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:15

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.7
75	30.0 - 60.0% of mass 95	50.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.3 (0.4) 1
174	50.0 - 100.0% of mass 95	76.5
175	5.0 - 9.0% of mass 174	5.8 (7.5) 1
176	95.0 - 101.0% of mass 174	73.8 (96.4) 1
177	5.0 - 9.0% of mass 176	4.8 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX048229.D	10/20/2025	10:59
VSTDIC020	VSTDIC020	VX048231.D	10/20/2025	11:40
VSTDIC050	VSTDIC050	VX048232.D	10/20/2025	12:01
VSTDIC100	VSTDIC100	VX048233.D	10/20/2025	12:22
VSTDIC150	VSTDIC150	VX048234.D	10/20/2025	12:43
VSTDIC005	VSTDIC005	VX048236.D	10/20/2025	13:49

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab File ID: VX048338.D

BFB Injection Date: 10/24/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:19

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.2
75	30.0 - 60.0% of mass 95	54.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	72.8
175	5.0 - 9.0% of mass 174	5.6 (7.8) 1
176	95.0 - 101.0% of mass 174	71.6 (98.4) 1
177	5.0 - 9.0% of mass 176	4.8 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX048339.D	10/24/2025	09:16
VX1024WBL01	VX1024WBL01	VX048340.D	10/24/2025	09:48
VX1024WBS01	VX1024WBS01	VX048341.D	10/24/2025	10:15
VX1024WBSD01	VX1024WBSD01	VX048342.D	10/24/2025	10:43
COMP-ROLLOFF	Q3399-04	VX048343.D	10/24/2025	11:03
VSTDCCC050EC	VSTDCCC050	VX048353.D	10/24/2025	15:49

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ICES02
 Lab Code: ACE SDG NO.: Q3399
 Lab File ID: VX048339.D Date Analyzed: 10/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	125053	5.53	210685	6.74	187763	10.04
UPPER LIMIT	250106	6.031	421370	7.239	375526	10.537
LOWER LIMIT	62526.5	5.031	105343	6.239	93881.5	9.537
EPA SAMPLE NO.						
COMP-ROLLOFF	137283	5.54	270507	6.75	270324	10.04
VX1024WBL01	90540	5.53	176490	6.74	165587	10.04
VX1024WBS01	145979	5.53	238661	6.74	208902	10.04
VX1024WBSD01	137701	5.53	225696	6.74	197290	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>
Lab File ID:	<u>VX048339.D</u>	Date Analyzed:	<u>10/24/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:16</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	93479	12.00€				
UPPER LIMIT	186958	12.50€				
LOWER LIMIT	46739.5	11.50€				
EPA SAMPLE NO.						
COMP-ROLLOFF	123520	12.01				
VX1024WBL01	72164	12.01				
VX1024WBS01	103563	12.01				
VX1024WBSD01	98612	12.00				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1024WBL01
Lab Sample ID: VX1024WBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3399
Matrix: TCLP
% Solid: 0
Test: TCLP VOA

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-01-4	Vinyl Chloride	0.75	U	1	0.26	0.75	1.00	ug/L	10/24/25 09:48	VX102425
75-35-4	1,1-Dichloroethene	0.75	U	1	0.23	0.75	1.00	ug/L	10/24/25 09:48	VX102425
78-93-3	2-Butanone	2.50	U	1	0.98	2.50	5.00	ug/L	10/24/25 09:48	VX102425
56-23-5	Carbon Tetrachloride	0.50	U	1	0.25	0.50	1.00	ug/L	10/24/25 09:48	VX102425
67-66-3	Chloroform	0.50	U	1	0.25	0.50	1.00	ug/L	10/24/25 09:48	VX102425
71-43-2	Benzene	0.50	U	1	0.15	0.50	1.00	ug/L	10/24/25 09:48	VX102425
107-06-2	1,2-Dichloroethane	0.50	U	1	0.22	0.50	1.00	ug/L	10/24/25 09:48	VX102425
79-01-6	Trichloroethene	0.75	U	1	0.090	0.75	1.00	ug/L	10/24/25 09:48	VX102425
127-18-4	Tetrachloroethene	0.50	U	1	0.23	0.50	1.00	ug/L	10/24/25 09:48	VX102425
108-90-7	Chlorobenzene	0.50	U	1	0.12	0.50	1.00	ug/L	10/24/25 09:48	VX102425
SURROGATES										
17060-07-0	1,2-Dichloroethane-d4	52.2			81 - 118		104%	SPK: 50		
1868-53-7	Dibromofluoromethane	39.0	*		80 - 119		78%	SPK: 50		
2037-26-5	Toluene-d8	43.1	*		89 - 112		86%	SPK: 50		
460-00-4	4-Bromofluorobenzene	42.7			85 - 114		85%	SPK: 50		
INTERNAL STANDARDS										
		Area Count								
363-72-4	Pentafluorobenzene	90500								
540-36-3	1,4-Difluorobenzene	176000								
3114-55-4	Chlorobenzene-d5	166000								
3855-82-1	1,4-Dichlorobenzene-d4	72200								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1024WBS01
Lab Sample ID: VX1024WBS01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3399
Matrix: TCLP
% Solid: 0
Test: TCLP VOA

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-01-4	Vinyl Chloride	27.8		1	0.26	0.75	1.00	ug/L	10/24/25 10:15	VX102425
75-35-4	1,1-Dichloroethene	20.7		1	0.23	0.75	1.00	ug/L	10/24/25 10:15	VX102425
78-93-3	2-Butanone	84.6		1	0.98	2.50	5.00	ug/L	10/24/25 10:15	VX102425
56-23-5	Carbon Tetrachloride	18.5		1	0.25	0.50	1.00	ug/L	10/24/25 10:15	VX102425
67-66-3	Chloroform	20.8		1	0.25	0.50	1.00	ug/L	10/24/25 10:15	VX102425
71-43-2	Benzene	20.2		1	0.15	0.50	1.00	ug/L	10/24/25 10:15	VX102425
107-06-2	1,2-Dichloroethane	21.0		1	0.22	0.50	1.00	ug/L	10/24/25 10:15	VX102425
79-01-6	Trichloroethene	19.2		1	0.090	0.75	1.00	ug/L	10/24/25 10:15	VX102425
127-18-4	Tetrachloroethene	20.9		1	0.23	0.50	1.00	ug/L	10/24/25 10:15	VX102425
108-90-7	Chlorobenzene	19.8		1	0.12	0.50	1.00	ug/L	10/24/25 10:15	VX102425
SURROGATES										
17060-07-0	1,2-Dichloroethane-d4	51.3			81 - 118		103%	SPK: 50		
1868-53-7	Dibromofluoromethane	41.6			80 - 119		83%	SPK: 50		
2037-26-5	Toluene-d8	49.6			89 - 112		99%	SPK: 50		
460-00-4	4-Bromofluorobenzene	48.3			85 - 114		97%	SPK: 50		
INTERNAL STANDARDS										
		Area Count								
363-72-4	Pentafluorobenzene	146000								
540-36-3	1,4-Difluorobenzene	239000								
3114-55-4	Chlorobenzene-d5	209000								
3855-82-1	1,4-Dichlorobenzene-d4	104000								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: VX1024WBSD01
Lab Sample ID: VX1024WBSD01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level: LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3399
Matrix: TCLP
% Solid: 0
Test: TCLP VOA

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
75-01-4	Vinyl Chloride	26.6		1	0.26	0.75	1.00	ug/L	10/24/25 10:43	VX102425
75-35-4	1,1-Dichloroethene	19.5		1	0.23	0.75	1.00	ug/L	10/24/25 10:43	VX102425
78-93-3	2-Butanone	95.2		1	0.98	2.50	5.00	ug/L	10/24/25 10:43	VX102425
56-23-5	Carbon Tetrachloride	19.0		1	0.25	0.50	1.00	ug/L	10/24/25 10:43	VX102425
67-66-3	Chloroform	20.8		1	0.25	0.50	1.00	ug/L	10/24/25 10:43	VX102425
71-43-2	Benzene	20.1		1	0.15	0.50	1.00	ug/L	10/24/25 10:43	VX102425
107-06-2	1,2-Dichloroethane	21.6		1	0.22	0.50	1.00	ug/L	10/24/25 10:43	VX102425
79-01-6	Trichloroethene	19.6		1	0.090	0.75	1.00	ug/L	10/24/25 10:43	VX102425
127-18-4	Tetrachloroethene	21.2		1	0.23	0.50	1.00	ug/L	10/24/25 10:43	VX102425
108-90-7	Chlorobenzene	19.7		1	0.12	0.50	1.00	ug/L	10/24/25 10:43	VX102425
SURROGATES										
17060-07-0	1,2-Dichloroethane-d4	53.1			81 - 118		106%	SPK: 50		
1868-53-7	Dibromofluoromethane	40.4			80 - 119		81%	SPK: 50		
2037-26-5	Toluene-d8	49.6			89 - 112		99%	SPK: 50		
460-00-4	4-Bromofluorobenzene	48.4			85 - 114		97%	SPK: 50		
INTERNAL STANDARDS										
		Area Count								
363-72-4	Pentafluorobenzene	138000								
540-36-3	1,4-Difluorobenzene	226000								
3114-55-4	Chlorobenzene-d5	197000								
3855-82-1	1,4-Dichlorobenzene-d4	98600								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>ICES02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3399</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>10/20/2025</u> <u>10/20/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>10:59</u> <u>13:49</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF001 = VX048229.D		RRF020 = VX048231.D		RRF050 = VX048232.D		
		RRF100 = VX048233.D		RRF150 = VX048234.D		RRF005 = VX048236.D		
COMPOUND	RRF001	RRF020	RRF050	RRF100	RRF150	RRF005	RRF	% RSD
Vinyl Chloride	0.447	0.437	0.416	0.462	0.445	0.458	0.444	3.7
1,1-Dichloroethene	0.515	0.548	0.485	0.558	0.531	0.514	0.525	5
2-Butanone	0.311	0.329	0.281	0.301	0.290	0.315	0.304	5.7
Carbon Tetrachloride	0.650	0.594	0.553	0.604	0.567	0.622	0.599	5.9
Chloroform	1.147	1.208	1.046	1.127	1.067	1.143	1.123	5.2
Benzene	1.493	1.480	1.301	1.426	1.347	1.405	1.409	5.3
1,2-Dichloroethane	0.529	0.566	0.500	0.528	0.496	0.499	0.520	5.2
Trichloroethene	0.400	0.373	0.348	0.380	0.363	0.366	0.372	4.7
Tetrachloroethene	0.418	0.378	0.328	0.370	0.361	0.385	0.373	7.9
Chlorobenzene	1.149	1.159	1.045	1.148	1.102	1.145	1.125	3.9
1,2-Dichloroethane-d4		0.709	0.682	0.690	0.698	0.725	0.701	2.4
Dibromofluoromethane		0.345	0.330	0.346	0.348	0.338	0.341	2.2
Toluene-d8		1.185	1.185	1.196	1.195	1.221	1.196	1.2
4-Bromofluorobenzene		0.445	0.444	0.446	0.441	0.500	0.455	5.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3399</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/24/2025</u> <u>09:16</u>
Lab File ID:	<u>VX048339.D</u>	Init. Calib. Date(s):	<u>10/20/2025</u> <u>10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59</u> <u>13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.444	0.627		41.22	20
1,1-Dichloroethene	0.525	0.572		8.95	20
2-Butanone	0.304	0.271		-10.85	20
Carbon Tetrachloride	0.599	0.550		-8.18	20
Chloroform	1.123	1.261		12.29	20
Benzene	1.409	1.480		5.04	20
1,2-Dichloroethane	0.520	0.592		13.85	20
Trichloroethene	0.372	0.374		0.54	20
Tetrachloroethene	0.373	0.372		-0.27	20
Chlorobenzene	1.125	1.142	0.3	1.51	20
1,2-Dichloroethane-d4	0.701	0.754		7.56	20
Dibromofluoromethane	0.341	0.277		-18.77	20
Toluene-d8	1.196	1.133		-5.27	20
4-Bromofluorobenzene	0.455	0.434		-4.61	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3399</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/24/2025 15:49</u>
Lab File ID:	<u>VX048353.D</u>	Init. Calib. Date(s):	<u>10/20/2025 10/20/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:59 13:49</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.444	0.606		36.49	50
1,1-Dichloroethene	0.525	0.541		3.05	50
2-Butanone	0.304	0.265		-12.83	50
Carbon Tetrachloride	0.599	0.525		-12.35	50
Chloroform	1.123	1.134		0.98	50
Benzene	1.409	1.352		-4.05	50
1,2-Dichloroethane	0.520	0.518		-0.38	50
Trichloroethene	0.372	0.346		-6.99	50
Tetrachloroethene	0.373	0.355		-4.83	50
Chlorobenzene	1.125	1.054	0.3	-6.31	50
1,2-Dichloroethane-d4	0.701	0.681		-2.85	50
Dibromofluoromethane	0.341	0.269		-21.11	50
Toluene-d8	1.196	1.117		-6.61	50
4-Bromofluorobenzene	0.455	0.412		-9.45	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.