

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048342.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1024WBSD01	Vinyl chloride	20	26.6	ug/L	133	4		58	137	20
	1,1-Dichloroethene	20	19.5	ug/L	98	6		71	131	20
	2-Butanone	100	95.2	ug/L	95	11		56	143	20
	Carbon Tetrachloride	20	19.0	ug/L	95	2		72	136	20
	Chloroform	20	20.8	ug/L	104	0		79	124	20
	Benzene	20	20.1	ug/L	101	0		79	120	20
	1,2-Dichloroethane	20	21.6	ug/L	108	3		73	128	20
	Trichloroethene	20	19.6	ug/L	98	2		79	123	20
	Tetrachloroethene	20	21.2	ug/L	106	2		74	129	20
	Chlorobenzene	20	19.7	ug/L	99	0		82	118	20