

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ICES02
 Lab Code: ACE SDG NO.: Q3399
 Lab File ID: VX048339.D Date Analyzed: 10/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	125053	5.53	210685	6.74	187763	10.04
UPPER LIMIT	250106	6.031	421370	7.239	375526	10.537
LOWER LIMIT	62526.5	5.031	105343	6.239	93881.5	9.537
EPA SAMPLE NO.						
COMP-ROLLOFF	137283	5.54	270507	6.75	270324	10.04
VX1024WBL01	90540	5.53	176490	6.74	165587	10.04
VX1024WBS01	145979	5.53	238661	6.74	208902	10.04
VX1024WBSD01	137701	5.53	225696	6.74	197290	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.