

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ICES02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3399</u>
Lab File ID:	<u>VX048339.D</u>	Date Analyzed:	<u>10/24/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:16</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	93479	12.00€				
UPPER LIMIT	186958	12.50€				
LOWER LIMIT	46739.5	11.50€				
EPA SAMPLE NO.						
COMP-ROLLOFF	123520	12.01				
VX1024WBL01	72164	12.01				
VX1024WBS01	103563	12.01				
VX1024WBSD01	98612	12.00				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.