

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3399</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>VX048284.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1022WBSD01	Vinyl chloride	20	17.8	ug/L	89	8		58	137	20
	1,1-Dichloroethene	20	17.0	ug/L	85	8		71	131	20
	2-Butanone	100	88.8	ug/L	89	0		56	143	20
	Carbon Tetrachloride	20	18.3	ug/L	92	3		72	136	20
	Chloroform	20	19.2	ug/L	96	1		79	124	20
	Benzene	20	18.6	ug/L	93	2		79	120	20
	1,2-Dichloroethane	20	19.6	ug/L	98	4		73	128	20
	Trichloroethene	20	18.8	ug/L	94	3		79	123	20
	Tetrachloroethene	20	18.3	ug/L	92	2		74	129	20
	Chlorobenzene	20	19.1	ug/L	96	1		82	118	20