

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ICES02
 Lab Code: ACE SDG NO.: Q3399
 Lab File ID: VX048280.D Date Analyzed: 10/22/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:37
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	76595	12				
UPPER LIMIT	153190	12.5				
LOWER LIMIT	38297.5	11.5				
EPA SAMPLE NO.						
COMP-ROLLOFF	141254	12.01				
VX1022WBL01	99844	12.01				
VX1022WBS01	91086	12.01				
VX1022WBSD01	92297	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.