

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ICES02
 Lab Code: ACE SDG NO.: Q3399
 Lab File ID: VX048339.D Date Analyzed: 10/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	93479	12.00€				
UPPER LIMIT	186958	12.50€				
LOWER LIMIT	46739.5	11.50€				
EPA SAMPLE NO.						
COMP-ROLLOFFRE	123520	12.01				
VX1024WBL01	72164	12.01				
VX1024WBS01	103563	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.