

Surrogate Summary

SDG No.: Q3399

Client: ICE Service Group, Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3399-03	COMP-ROLLOFF	1,2-Dichloroethane-d4	50	50.1	100		71	136
		Dibromofluoromethane	50	51.1	102		78	119
		Toluene-d8	50	49.4	99		85	116
		4-Bromofluorobenzene	50	42.8	86		79	119
VY1021SBL01	VY1021SBL01	1,2-Dichloroethane-d4	50	46.0	92		71	136
		Dibromofluoromethane	50	49.7	99		78	119
		Toluene-d8	50	48.1	96		85	116
		4-Bromofluorobenzene	50	36.4	73	*	79	119
VY1021SBS01	VY1021SBS01	1,2-Dichloroethane-d4	50	46.9	94		71	136
		Dibromofluoromethane	50	50.9	102		78	119
		Toluene-d8	50	51.3	103		85	116
		4-Bromofluorobenzene	50	49.9	100		79	119
VY1021SBSD01	VY1021SBSD01	1,2-Dichloroethane-d4	50	45.6	91		71	136
		Dibromofluoromethane	50	49.6	99		78	119
		Toluene-d8	50	49.7	99		85	116
		4-Bromofluorobenzene	50	47.6	95		79	119