

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3399

Lab File ID: VY023520.D

BFB Injection Date: 10/21/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:29

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.5
75	30.0 - 60.0% of mass 95	48.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1.3 (1.4) 1
174	50.0 - 100.0% of mass 95	94.7
175	5.0 - 9.0% of mass 174	6.8 (7.1) 1
176	95.0 - 101.0% of mass 174	92.5 (97.7) 1
177	5.0 - 9.0% of mass 176	6.2 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023521.D	10/21/2025	09:05
VY1021SBL01	VY1021SBL01	VY023522.D	10/21/2025	09:34
VY1021SBS01	VY1021SBS01	VY023523.D	10/21/2025	10:04
VY1021SBSD01	VY1021SBSD01	VY023524.D	10/21/2025	10:27
COMP-ROLLOFF	Q3399-03	VY023534.D	10/21/2025	14:59