



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ICE Service Group, Inc.

SDG No.: Q3400

Project: NWIRP Northrup Grumman Site – Bethpage, NY

RunNo.: LB137606

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90–110	10/22/2025
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90–110	10/22/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90–110	10/22/2025