

Duplicate Sample Summary

Client:	ICE Service Group, Inc.	SDG No.:	Q3400
Project:	NWIRP Northrup Grumman Site – Bethpage, NY	Sample ID:	Q3385-04
Client ID:	EFF-WWDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	6.50		6.30		1	3.13		10/22/2025

Duplicate Sample Summary

Client: ICE Service Group, Inc.	SDG No.: Q3400
Project: NWIRP Northrup Grumman Site – Bethpage, NY	Sample ID: Q3398-01
Client ID: A748DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	9.43		9.45		1	0.21		10/22/2025

Duplicate Sample Summary

Client:	ICE Service Group, Inc.	SDG No.:	Q3400
Project:	NWIRP Northrup Grumman Site – Bethpage, NY	Sample ID:	Q3400-01
Client ID:	FRAC-TANK-0760450DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.62		7.64		1	0.26		10/22/2025

Duplicate Sample Summary

Client: ICE Service Group, Inc.	SDG No.: Q3400
Project: NWIRP Northrup Grumman Site – Bethpage, NY	Sample ID: Q3417-01
Client ID: 101425DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	5.60		5.62		1	0.36		10/22/2025