

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137605

Review By	mohan	Review On	10/22/2025 12:03:39 PM
Supervise By	jaswal	Supervise On	10/22/2025 12:07:45 PM

STD. NAME	STD REF.#
ICAL Standard	MP87670,MP87671,MP87672,MP87673,MP87674,MP87675
ICV Standard	MP87676
CCV Standard	MP87678
ICSA Standard	
CRI Standard	MP87680
LCS Standard	
Chk Standard	MP87677,MP87679,MP87681,MP87683

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/21/25 16:46		mohan	OK
2	S0.2	S0.2	CAL2	10/21/25 16:49		mohan	OK
3	S2.5	S2.5	CAL3	10/21/25 16:52		mohan	OK
4	S5	S5	CAL4	10/21/25 16:54		mohan	OK
5	S7.5	S7.5	CAL5	10/21/25 16:59		mohan	OK
6	S10	S10	CAL6	10/21/25 17:07		mohan	OK
7	ICV82	ICV82	ICV	10/21/25 17:10		mohan	OK
8	ICB82	ICB82	ICB	10/21/25 17:12		mohan	OK
9	CCV68	CCV68	CCV	10/21/25 17:15		mohan	OK
10	CCB68	CCB68	CCB	10/21/25 17:17		mohan	OK
11	CRA	CRA	CRDL	10/21/25 17:19		mohan	OK
12	HighStd	HighStd	HIGH STD	10/21/25 17:21		mohan	OK
13	ChkStd	ChkStd	SAM	10/21/25 17:24		mohan	OK
14	PB170194BL	PB170194BL	MB	10/21/25 17:26		mohan	OK
15	PB170194BS	PB170194BS	LCS	10/21/25 17:28		mohan	OK
16	Q3386-03	3443	SAM	10/21/25 17:31		mohan	OK
17	Q3386-04	3444	SAM	10/21/25 17:33		mohan	OK
18	Q3398-01	A748	SAM	10/21/25 17:36		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137605

Review By	mohan	Review On	10/22/2025 12:03:39 PM
Supervise By	jaswal	Supervise On	10/22/2025 12:07:45 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87670,MP87671,MP87672,MP87673,MP87674,MP87675		
ICV Standard	MP87676		
CCV Standard	MP87678		
ICSA Standard			
CRI Standard	MP87680		
LCS Standard			
Chk Standard	MP87677,MP87679,MP87681,MP87683		

19	Q3400-01	FRAC-TANK-0760450	SAM	10/21/25 17:38		mohan	OK
20	Q3400-01DUP	FRAC-TANK-0760450	DUP	10/21/25 17:41		mohan	OK
21	CCV69	CCV69	CCV	10/21/25 17:43		mohan	OK
22	CCB69	CCB69	CCB	10/21/25 17:45		mohan	OK
23	Q3400-01MS	FRAC-TANK-0760450	MS	10/21/25 17:47		mohan	OK
24	Q3400-01MSD	FRAC-TANK-0760450	MSD	10/21/25 17:50		mohan	OK
25	Q3400-01L	FRAC-TANK-0760450	SD	10/21/25 17:52		mohan	OK
26	Q3400-01A	FRAC-TANK-0760450	PS	10/21/25 17:55		mohan	OK
27	CCV70	CCV70	CCV	10/21/25 18:00		mohan	OK
28	CCB70	CCB70	CCB	10/21/25 18:03		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/21/25 12:03		Jaswal	OK
2	S1	S1	CAL2	10/21/25 12:08		Jaswal	OK
3	S2	S2	CAL3	10/21/25 12:12		Jaswal	OK
4	S3	S3	CAL4	10/21/25 12:16		Jaswal	OK
5	S4	S4	CAL5	10/21/25 12:21		Jaswal	OK
6	S5	S5	CAL6	10/21/25 12:25		Jaswal	OK
7	ICV01	ICV01	ICV	10/21/25 12:40		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/21/25 12:44		Jaswal	OK
9	ICB01	ICB01	ICB	10/21/25 12:50		Jaswal	OK
10	CRI01	CRI01	CRDL	10/21/25 12:54		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/21/25 12:59		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/21/25 13:03		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/21/25 13:07		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/21/25 13:11		Jaswal	OK
15	CCV01	CCV01	CCV	10/21/25 13:45		Jaswal	OK
16	CCB01	CCB01	CCB	10/21/25 13:54		Jaswal	OK
17	LR-1	LR-1	HIGH STD	10/21/25 13:57		Jaswal	OK
18	LR-2	LR-2	HIGH STD	10/21/25 14:02		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

19	PB170187BL	PB170187BL	MB	10/21/25 14:11		Jaswal	OK
20	PB170187BS	PB170187BS	LCS	10/21/25 14:16		Jaswal	OK
21	PB170183BL	PB170183BL	MB	10/21/25 14:25		Jaswal	OK
22	PB170183BS	PB170183BS	LCS	10/21/25 14:29		Jaswal	OK
23	LR-3	LR-3	HIGH STD	10/21/25 14:33		Jaswal	OK
24	Q3381-02	VNJ-222	SAM	10/21/25 14:39		Jaswal	OK
25	Q3381-04	VNJ-245	SAM	10/21/25 14:43		Jaswal	OK
26	Q3381-06	RBR200059	SAM	10/21/25 14:47		Jaswal	OK
27	CCV02	CCV02	CCV	10/21/25 15:03		Jaswal	OK
28	CCB02	CCB02	CCB	10/21/25 15:07		Jaswal	OK
29	Q3381-08	VNJ-255	SAM	10/21/25 15:11		Jaswal	OK
30	Q3369-04	MH-22	SAM	10/21/25 15:16		Jaswal	OK
31	Q3369-08	MH-21	SAM	10/21/25 15:20		Jaswal	OK
32	Q3374-04	TP-4-WC	SAM	10/21/25 15:25		Jaswal	OK
33	Q3374-08	TP-5-WC	SAM	10/21/25 15:29		Jaswal	OK
34	Q3388-02	461	SAM	10/21/25 15:34		Jaswal	OK
35	Q3390-02	4174	SAM	10/21/25 15:38		Jaswal	OK
36	Q3392-04	MH-20	SAM	10/21/25 15:42		Jaswal	OK
37	Q3392-08	MH-19	SAM	10/21/25 15:47		Jaswal	OK
38	Q3392-08DUP	MH-19DUP	DUP	10/21/25 15:51		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

39	CCV03	CCV03	CCV	10/21/25 15:56		Jaswal	OK
40	CCB03	CCB03	CCB	10/21/25 16:00		Jaswal	OK
41	Q3392-08L	MH-19L	SD	10/21/25 16:04		Jaswal	OK
42	Q3392-08MS	MH-19MS	MS	10/21/25 16:09		Jaswal	OK
43	Q3392-08MSD	MH-19MSD	MSD	10/21/25 16:13		Jaswal	OK
44	Q3392-08A	MH-19A	PS	10/21/25 16:17	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
45	PB170147TB	PB170147TB	MB	10/21/25 16:21		Jaswal	OK
46	Q3411-01	Q4-WASTEWATER	SAM	10/21/25 16:26		Jaswal	OK
47	Q3401-01	41-CONCRETE	SAM	10/21/25 16:30		Jaswal	OK
48	Q3402-01	41-LA	SAM	10/21/25 16:34		Jaswal	OK
49	Q3402-04	41-RA	SAM	10/21/25 16:40		Jaswal	OK
50	Q3402-07	41-LB	SAM	10/21/25 16:44		Jaswal	OK
51	CCV04	CCV04	CCV	10/21/25 16:48		Jaswal	OK
52	CCB04	CCB04	CCB	10/21/25 16:52		Jaswal	OK
53	Q3402-10	41-RB	SAM	10/21/25 16:56		Jaswal	OK
54	Q3403-01	MH-18	SAM	10/21/25 17:00		Jaswal	OK
55	Q3397-01	TP-7-WC	SAM	10/21/25 17:04		Jaswal	OK
56	Q3397-01DUP	TP-7-WCDUP	DUP	10/21/25 17:08		Jaswal	OK
57	Q3397-01L	TP-7-WCL	SD	10/21/25 17:12		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

58	Q3397-01MS	TP-7-WCMS	MS	10/21/25 17:16		Jaswal	OK
59	Q3397-01MSD	TP-7-WCMSD	MSD	10/21/25 17:20		Jaswal	OK
60	Q3397-01A	TP-7-WCA	PS	10/21/25 17:24	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
61	PB170192BL	PB170192BL	MB	10/21/25 17:28	LCS fail for Cd	Jaswal	Not Ok
62	PB170192BS	PB170192BS	LCS	10/21/25 17:33	LCS fail for Cd	Jaswal	Not Ok
63	CCV05	CCV05	CCV	10/21/25 17:37		Jaswal	OK
64	CCB05	CCB05	CCB	10/21/25 17:41		Jaswal	OK
65	Q3365-01	MW25	SAM	10/21/25 17:45	Na is high, Need dilution	Jaswal	Dilution
66	Q3365-02	MW2D	SAM	10/21/25 17:49	Na is high, Need dilution	Jaswal	Dilution
67	Q3365-03	SUPPLY	SAM	10/21/25 17:53		Jaswal	OK
68	Q3365-04	17 HARVEY	SAM	10/21/25 17:58		Jaswal	OK
69	Q3365-05	19 HARVEY	SAM	10/21/25 18:02		Jaswal	OK
70	Q3370-01	WATER-TREATMENT	SAM	10/21/25 18:06		Jaswal	OK
71	Q3370-01DUP	WATER-TREATMENT	DUP	10/21/25 18:11		Jaswal	OK
72	Q3370-01L	WATER-TREATMENT	SD	10/21/25 18:15		Jaswal	OK
73	Q3370-01MS	WATER-TREATMENT	MS	10/21/25 18:19		Jaswal	OK
74	Q3370-01MSD	WATER-TREATMENT	MSD	10/21/25 18:23		Jaswal	OK
75	CCV06	CCV06	CCV	10/21/25 18:27		Jaswal	OK
76	CCB06	CCB06	CCB	10/21/25 18:31		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638		
ICV Standard	MP87644,MP87643		
CCV Standard	MP87647		
ICSA Standard	MP87645,MP87646		
CRI Standard	MP87643		
LCS Standard			
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650		

77	Q3370-01A	WATER-TREATMENT	PS	10/21/25 18:36	0.1ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
78	Q3376-07	PR132-FB01-202510	SAM	10/21/25 18:40		Jaswal	OK
79	Q3400-01	FRAC-TANK-0760450	SAM	10/21/25 18:44		Jaswal	OK
80	Q3398-01	A748	SAM	10/21/25 18:48	BAD Injection	Jaswal	Not Ok
81	Q3399-01	COMP-ROLLOFF	SAM	10/21/25 19:01		Jaswal	OK
82	Q3405-01	PAD-10-20-25-A	SAM	10/21/25 19:05		Jaswal	OK
83	Q3405-02	PAD-10-20-25-B	SAM	10/21/25 19:09		Jaswal	OK
84	Q3405-03	PAD-10-20-25-C	SAM	10/21/25 19:14		Jaswal	OK
85	CCV07	CCV07	CCV	10/21/25 19:18		Jaswal	OK
86	CCB07	CCB07	CCB	10/21/25 19:22		Jaswal	OK
87	Q3405-04	SW-10-20-25	SAM	10/21/25 19:26		Jaswal	OK
88	Q3409-01	EXCAVATION-AREA-	SAM	10/21/25 19:31		Jaswal	OK
89	Q3409-02	EXCAVATION-AREA-	SAM	10/21/25 19:35		Jaswal	OK
90	Q3412-01	TP-6	SAM	10/21/25 19:39	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
91	Q3412-01DUP	TP-6DUP	DUP	10/21/25 19:43	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
92	Q3412-01L	TP-6L	SD	10/21/25 19:47	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
93	Q3412-01MS	TP-6MS	MS	10/21/25 19:51	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
94	Q3412-01MSD	TP-6MSD	MSD	10/21/25 19:55	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137613

Review By	Janvi	Review On	10/23/2025 4:36:41 PM
Supervise By	jaswal	Supervise On	10/24/2025 6:34:14 PM

STD. NAME	STD REF.#
ICAL Standard	MP87637,MP87643,MP87641,MP87640,MP87639,MP87638
ICV Standard	MP87644,MP87643
CCV Standard	MP87647
ICSA Standard	MP87645,MP87646
CRI Standard	MP87643
LCS Standard	
Chk Standard	MP87668,MP87648,MP87651,MP87649,MP87650

95	Q3412-01A	TP-6A	PS	10/21/25 19:59	MS/MSD Fail for more then 50% Parameters	Jaswal	Not Ok
96	CCV08	CCV08	CCV	10/21/25 20:03		Jaswal	OK
97	CCB08	CCB08	CCB	10/21/25 20:07		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/03/25 11:43		Jaswal	OK
2	S1	S1	CAL2	11/03/25 11:47		Jaswal	OK
3	S2	S2	CAL3	11/03/25 11:51		Jaswal	OK
4	S3	S3	CAL4	11/03/25 11:55		Jaswal	OK
5	S4	S4	CAL5	11/03/25 11:59		Jaswal	OK
6	S5	S5	CAL6	11/03/25 12:04		Jaswal	OK
7	ICV01	ICV01	ICV	11/03/25 12:08		Jaswal	OK
8	LLICV01	LLICV01	LLICV	11/03/25 12:27		Jaswal	OK
9	ICB01	ICB01	ICB	11/03/25 12:31		Jaswal	OK
10	CRI01	CRI01	CRDL	11/03/25 12:35		Jaswal	OK
11	ICSA01	ICSA01	ICSA	11/03/25 12:40		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	11/03/25 12:44		Jaswal	OK
13	ICSADL	ICSADL	ICSA	11/03/25 12:48		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	11/03/25 12:52		Jaswal	OK
15	CCV01	CCV01	CCV	11/03/25 12:56		Jaswal	OK
16	CCB01	CCB01	CCB	11/03/25 13:01		Jaswal	OK
17	LR-1	LR-1	HIGH STD	11/03/25 13:06		Jaswal	OK
18	LR-2	LR-2	HIGH STD	11/03/25 13:10		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	LR-3	LR-3	HIGH STD	11/03/25 13:17		Jaswal	OK
20	PB170192BL	PB170192BL	MB	11/03/25 13:22		Jaswal	OK
21	PB170295TB	PB170295TB	MB	11/03/25 13:36		Jaswal	OK
22	PB170192BS	PB170192BS	LCS	11/03/25 13:45		Jaswal	OK
23	PB170333TB	PB170333TB	MB	11/03/25 13:50		Jaswal	OK
24	Q3482-08	WC-1	SAM	11/03/25 13:54		Jaswal	OK
25	Q3482-09	WC-2	SAM	11/03/25 13:59		Jaswal	OK
26	Q3482-10	WC-3	SAM	11/03/25 14:03		Jaswal	OK
27	CCV02	CCV02	CCV	11/03/25 14:07		Jaswal	OK
28	CCB02	CCB02	CCB	11/03/25 14:11		Jaswal	OK
29	Q3482-11	WC-4	SAM	11/03/25 14:16		Jaswal	OK
30	Q3482-12	WC-5	SAM	11/03/25 14:20		Jaswal	OK
31	Q3482-13	WC-6	SAM	11/03/25 14:25		Jaswal	OK
32	Q3482-14	WC-7	SAM	11/03/25 14:29		Jaswal	OK
33	Q3488-03	MOO-25-0299	SAM	11/03/25 14:34		Jaswal	OK
34	Q3488-05	MOO-25-0302-0310	SAM	11/03/25 14:38		Jaswal	OK
35	Q3488-07	MOO-25-0311-0313	SAM	11/03/25 14:42		Jaswal	OK
36	Q3490-02	LAW-25-0151	SAM	11/03/25 14:47		Jaswal	OK
37	Q3499-02	RBR251235	SAM	11/03/25 14:51		Jaswal	OK
38	Q3501-02	461	SAM	11/03/25 14:55		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	CCV03	CCV03	CCV	11/03/25 15:00		Jaswal	OK
40	CCB03	CCB03	CCB	11/03/25 15:04		Jaswal	OK
41	PB170355BL	PB170355BL	MB	11/03/25 15:08		Jaswal	OK
42	PB170355BS	PB170355BS	LCS	11/03/25 15:12		Jaswal	OK
43	Q3502-01DL	CATCH-CAN-USED-C	SAM	11/03/25 15:16		Jaswal	Confirms
44	Q3503-02	322	SAM	11/03/25 15:21		Jaswal	OK
45	Q3503-04	323	SAM	11/03/25 15:25	Cd is high, Need dilution	Jaswal	Dilution
46	Q3503-06	324	SAM	11/03/25 15:30		Jaswal	OK
47	Q3504-02	72-11986	SAM	11/03/25 15:34		Jaswal	OK
48	Q3504-02DUP	72-11986DUP	DUP	11/03/25 15:38		Jaswal	OK
49	Q3504-02L	72-11986L	SD	11/03/25 15:43		Jaswal	OK
50	Q3504-02MS	72-11986MS	MS	11/03/25 15:47		Jaswal	OK
51	CCV04	CCV04	CCV	11/03/25 15:51		Jaswal	OK
52	CCB04	CCB04	CCB	11/03/25 15:55		Jaswal	OK
53	Q3504-02MSD	72-11986MSD	MSD	11/03/25 16:00		Jaswal	OK
54	Q3504-02A	72-11986A	PS	11/03/25 16:04		Jaswal	OK
55	PB170357BL	PB170357BL	MB	11/03/25 16:08		Jaswal	OK
56	PB170357BS	PB170357BS	LCS	11/03/25 16:12		Jaswal	OK
57	Q3500-02	462	SAM	11/03/25 16:16		Jaswal	OK
58	Q3500-03	463	SAM	11/03/25 16:20		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

59	Q3500-04DL	465DL	SAM	11/03/25 16:25		Jaswal	Confirms
60	Q3500-05	461-LIQUID	SAM	11/03/25 16:29		Jaswal	OK
61	Q3497-01	OK-01-103025	SAM	11/03/25 16:33		Jaswal	OK
62	Q3501-01	461	SAM	11/03/25 16:37		Jaswal	OK
63	CCV05	CCV05	CCV	11/03/25 16:41		Jaswal	OK
64	CCB05	CCB05	CCB	11/03/25 16:45		Jaswal	OK
65	Q3502-04	PARTS-CLEANER	SAM	11/03/25 16:50		Jaswal	OK
66	Q3495-03	TW-1025-1	SAM	11/03/25 16:54		Jaswal	OK
67	Q3495-03DUP	TW-1025-1DUP	DUP	11/03/25 16:58		Jaswal	OK
68	Q3495-03L	TW-1025-1L	SD	11/03/25 17:03		Jaswal	OK
69	Q3495-03MS	TW-1025-1MS	MS	11/03/25 17:07		Jaswal	OK
70	Q3495-03MSD	TW-1025-1MSD	MSD	11/03/25 17:11		Jaswal	OK
71	Q3495-03A	TW-1025-1A	PS	11/03/25 17:15		Jaswal	OK
72	PB170353BL	PB170353BL	MB	11/03/25 17:19		Jaswal	OK
73	PB170353BS	PB170353BS	LCS	11/03/25 17:23		Jaswal	OK
74	Q3496-01	DRILL-CUTTINGS	SAM	11/03/25 17:27		Jaswal	OK
75	CCV06	CCV06	CCV	11/03/25 17:32		Jaswal	OK
76	CCB06	CCB06	CCB	11/03/25 17:36		Jaswal	OK
77	Q3495-01	SB-1025-1	SAM	11/03/25 17:40	Q3495-01DUP bd injection	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

78	Q3495-01DUP	SB-1025-1DUP	DUP	11/03/25 17:44	Q3495-01DUP bad injection	Jaswal	Not Ok
79	Q3495-01L	SB-1025-1L	SD	11/03/25 17:48	Q3495-01DUP bad injection	Jaswal	Not Ok
80	Q3495-01MS	SB-1025-1MS	MS	11/03/25 17:53	Q3495-01DUP bad injection	Jaswal	Not Ok
81	Q3495-01MSD	SB-1025-1MSD	MSD	11/03/25 17:56	Q3495-01DUP bad injection	Jaswal	Not Ok
82	Q3495-01A	SB-1025-1A	PS	11/03/25 18:00	Q3495-01DUP bad injection	Jaswal	Not Ok
83	Q3495-02	SB-1025-2	SAM	11/03/25 18:04		Jaswal	OK
84	Q3475-01DL	GW-DRUMDL	SAM	11/03/25 18:08	5X For Ag, still oversaturated.	Jaswal	Dilution
85	Q3502-01	CATCH-CAN-USED-C	SAM	11/03/25 18:13		Jaswal	OK
86	Q3519-01	PL-01-103125	SAM	11/03/25 18:17	Not use	Jaswal	Not Ok
87	CCV07	CCV07	CCV	11/03/25 18:21		Jaswal	OK
88	CCB07	CCB07	CCB	11/03/25 18:25		Jaswal	OK
89	PB170374BL	PB170374BL	MB	11/03/25 18:29	Not use	Jaswal	Not Ok
90	PB170374BS	PB170374BS	LCS	11/03/25 18:34	Not use	Jaswal	Not Ok
91	Q3512-01	10200	SAM	11/03/25 18:38	Not use	Jaswal	Not Ok
92	Q3515-01	SU-04-103125	SAM	11/03/25 18:42	Not use	Jaswal	Not Ok
93	Q3516-01	HD-01-103125	SAM	11/03/25 18:46	Not use	Jaswal	Not Ok
94	Q3520-01	EO-02-103125	SAM	11/03/25 18:51	Not use	Jaswal	Not Ok
95	Q3520-01DUP	EO-02-103125DUP	DUP	11/03/25 18:55	Not use	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137750

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

96	Q3520-01L	EO-02-103125L	SD	11/03/25 18:59	Not use	Jaswal	Not Ok
97	Q3520-01MS	EO-02-103125MS	MS	11/03/25 19:03	Not use	Jaswal	Not Ok
98	Q3520-01MSD	EO-02-103125MSD	MSD	11/03/25 19:07	Not use	Jaswal	Not Ok
99	CCV08	CCV08	CCV	11/03/25 19:11	Not use	Jaswal	Not Ok
100	CCB08	CCB08	CCB	11/03/25 19:15	Not use	Jaswal	Not Ok
101	Q3520-01A	EO-02-103125A	PS	11/03/25 19:20	Not use	Jaswal	Not Ok
102	Q3503-04DL	323DL	SAM	11/03/25 19:24	Not use	Jaswal	Not Ok
103	PB170382BL	PB170382BL	MB	11/03/25 19:28	Not use	Jaswal	Not Ok
104	PB170382BS	PB170382BS	LCS	11/03/25 19:33	Not use	Jaswal	Not Ok
105	Q3527-01	RW8-SP100-2025103	SAM	11/03/25 19:37	Not use	Jaswal	Not Ok
106	Q3527-02	RW8-SP303-2025103	SAM	11/03/25 19:41	Not use	Jaswal	Not Ok
107	Q3527-02DUP	RW8-SP303-2025103	DUP	11/03/25 19:45	Not use	Jaswal	Not Ok
108	Q3527-02L	RW8-SP303-2025103	SD	11/03/25 19:49	Not use	Jaswal	Not Ok
109	Q3527-02MS	RW8-SP303-2025103	MS	11/03/25 19:54	Not use	Jaswal	Not Ok
110	Q3527-02MSD	RW8-SP303-2025103	MSD	11/03/25 19:58	Not use	Jaswal	Not Ok
111	CCV09	CCV09	CCV	11/03/25 20:02		Jaswal	OK
112	CCB09	CCB09	CCB	11/03/25 20:06		Jaswal	OK
113	Q3527-02A	RW8-SP303-2025103	PS	11/03/25 20:10	Not use	Jaswal	Not Ok
114	CCV10	CCV10	CCV	11/03/25 20:15		Jaswal	OK
115	CCB10	CCB10	CCB	11/03/25 20:19		Jaswal	OK