

Surrogate Summary

SDG No.: **Q3400**

Client: **ICE Service Group, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
I.BLK-PQ071090.D	PIBLK-PQ071090.D	Tetrachloro-m-xyl	1	20	20.9	104		60	140
		Decachlorobiphen	1	20	23.7	118		60	140
		Tetrachloro-m-xyl	2	20	19.3	97		60	140
		Decachlorobiphen	2	20	23.1	115		60	140
PB170211BL	PB170211BL	Tetrachloro-m-xyl	1	20	24.8	124		30	173
		Decachlorobiphen	1	20	28.2	141		10	173
		Tetrachloro-m-xyl	2	20	23.0	115		30	173
		Decachlorobiphen	2	20	27.3	137		10	173
Q3400-01	FRAC-TANK-0760450	Tetrachloro-m-xyl	1	20	23.4	117		30	173
		Decachlorobiphen	1	20	9.38	47		10	173
		Tetrachloro-m-xyl	2	20	21.7	109		30	173
		Decachlorobiphen	2	20	9.32	47		10	173
I.BLK-PQ071102.D	PIBLK-PQ071102.D	Tetrachloro-m-xyl	1	20	20.6	103		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
		Tetrachloro-m-xyl	2	20	19.1	95		60	140
		Decachlorobiphen	2	20	22.9	114		60	140
I.BLK-PQ071120.D	PIBLK-PQ071120.D	Tetrachloro-m-xyl	1	20	20.3	101		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
		Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	21.9	110		60	140
PB170211BS	PB170211BS	Tetrachloro-m-xyl	1	20	21.6	108		30	173
		Decachlorobiphen	1	20	25.2	126		10	173
		Tetrachloro-m-xyl	2	20	16.9	85		30	173
		Decachlorobiphen	2	20	23.4	117		10	173
PB170211BSD	PB170211BSD	Tetrachloro-m-xyl	1	20	22.4	112		30	173
		Decachlorobiphen	1	20	25.7	128		10	173
		Tetrachloro-m-xyl	2	20	18.3	91		30	173
		Decachlorobiphen	2	20	24.2	121		10	173
I.BLK-PQ071128.D	PIBLK-PQ071128.D	Tetrachloro-m-xyl	1	20	20.8	104		60	140
		Decachlorobiphen	1	20	24.2	121		60	140
		Tetrachloro-m-xyl	2	20	18.4	92		60	140
		Decachlorobiphen	2	20	22.9	115		60	140