

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114834.D **Time Analyzed:** 08:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.488	5.392	5.592	558.010	500.000	11.6
Aroclor-1016-2	5.509	5.414	5.614	532.430	500.000	6.5
Aroclor-1016-3	5.571	5.476	5.676	532.150	500.000	6.4
Aroclor-1016-4	5.667	5.571	5.771	564.480	500.000	12.9
Aroclor-1016-5	5.958	5.863	6.063	564.580	500.000	12.9
Aroclor-1260-1	7.071	6.975	7.175	552.040	500.000	10.4
Aroclor-1260-2	7.325	7.229	7.429	550.680	500.000	10.1
Aroclor-1260-3	7.682	7.585	7.785	532.380	500.000	6.5
Aroclor-1260-4	7.904	7.807	8.007	534.270	500.000	6.9
Aroclor-1260-5	8.210	8.113	8.313	510.800	500.000	2.2
Decachlorobiphenyl	9.931	9.834	10.034	51.810	50.000	3.6
Tetrachloro-m-xylene	4.349	4.255	4.455	54.380	50.000	8.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114834.D **Time Analyzed:** 08:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.643	4.843	467.830	500.000	-6.4
Aroclor-1016-2	4.748	4.662	4.862	471.560	500.000	-5.7
Aroclor-1016-3	4.922	4.836	5.036	501.070	500.000	0.2
Aroclor-1016-4	4.963	4.879	5.079	512.000	500.000	2.4
Aroclor-1016-5	5.176	5.092	5.292	502.560	500.000	0.5
Aroclor-1260-1	6.206	6.125	6.325	519.290	500.000	3.9
Aroclor-1260-2	6.394	6.313	6.513	530.830	500.000	6.2
Aroclor-1260-3	6.547	6.466	6.666	501.170	500.000	0.2
Aroclor-1260-4	7.017	6.939	7.139	580.950	500.000	16.2
Aroclor-1260-5	7.258	7.179	7.379	558.650	500.000	11.7
Decachlorobiphenyl	8.642	8.571	8.771	51.040	50.000	2.1
Tetrachloro-m-xylene	3.651	3.560	3.760	46.540	50.000	-6.9