

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114849.D **Time Analyzed:** 14:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.487	5.392	5.592	482.500	500.000	-3.5
Aroclor-1016-2	5.509	5.414	5.614	451.400	500.000	-9.7
Aroclor-1016-3	5.570	5.476	5.676	442.490	500.000	-11.5
Aroclor-1016-4	5.666	5.571	5.771	489.320	500.000	-2.1
Aroclor-1016-5	5.957	5.863	6.063	524.730	500.000	4.9
Aroclor-1260-1	7.071	6.975	7.175	474.850	500.000	-5.0
Aroclor-1260-2	7.324	7.229	7.429	457.380	500.000	-8.5
Aroclor-1260-3	7.681	7.585	7.785	431.740	500.000	-13.7
Aroclor-1260-4	7.903	7.807	8.007	443.690	500.000	-11.3
Aroclor-1260-5	8.210	8.113	8.313	430.500	500.000	-13.9
Decachlorobiphenyl	9.932	9.834	10.034	45.480	50.000	-9.0
Tetrachloro-m-xylene	4.349	4.255	4.455	53.590	50.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114849.D **Time Analyzed:** 14:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.729	4.643	4.843	450.780	500.000	-9.8
Aroclor-1016-2	4.748	4.662	4.862	453.430	500.000	-9.3
Aroclor-1016-3	4.923	4.836	5.036	423.360	500.000	-15.3
Aroclor-1016-4	4.965	4.879	5.079	442.140	500.000	-11.6
Aroclor-1016-5	5.177	5.092	5.292	472.480	500.000	-5.5
Aroclor-1260-1	6.207	6.125	6.325	436.270	500.000	-12.7
Aroclor-1260-2	6.395	6.313	6.513	431.390	500.000	-13.7
Aroclor-1260-3	6.547	6.466	6.666	406.860	500.000	-18.6
Aroclor-1260-4	7.019	6.939	7.139	440.280	500.000	-11.9
Aroclor-1260-5	7.259	7.179	7.379	451.450	500.000	-9.7
Decachlorobiphenyl	8.643	8.571	8.771	40.090	50.000	-19.8
Tetrachloro-m-xylene	3.652	3.560	3.760	47.250	50.000	-5.5