

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071086.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.481	4.380	4.580	536.950	500.000	7.4
Aroclor-1016-2	4.499	4.399	4.599	535.300	500.000	7.1
Aroclor-1016-3	4.557	4.456	4.656	544.430	500.000	8.9
Aroclor-1016-4	4.647	4.546	4.746	540.610	500.000	8.1
Aroclor-1016-5	4.931	4.831	5.031	533.730	500.000	6.7
Aroclor-1260-1	6.021	5.919	6.119	523.570	500.000	4.7
Aroclor-1260-2	6.277	6.175	6.375	524.480	500.000	4.9
Aroclor-1260-3	6.628	6.526	6.726	542.650	500.000	8.5
Aroclor-1260-4	6.842	6.740	6.940	551.920	500.000	10.4
Aroclor-1260-5	7.147	7.045	7.245	548.300	500.000	9.7
Decachlorobiphenyl	8.499	8.396	8.596	54.810	50.000	9.6
Tetrachloro-m-xylene	3.400	3.300	3.500	53.710	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071086.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	448.240	500.000	-10.4
Aroclor-1016-2	3.792	3.693	3.893	444.160	500.000	-11.2
Aroclor-1016-3	3.952	3.853	4.053	458.590	500.000	-8.3
Aroclor-1016-4	4.000	3.902	4.102	462.320	500.000	-7.5
Aroclor-1016-5	4.195	4.097	4.297	456.210	500.000	-8.8
Aroclor-1260-1	5.187	5.089	5.289	455.010	500.000	-9.0
Aroclor-1260-2	5.376	5.278	5.478	454.650	500.000	-9.1
Aroclor-1260-3	5.518	5.419	5.619	457.030	500.000	-8.6
Aroclor-1260-4	5.977	5.879	6.079	459.370	500.000	-8.1
Aroclor-1260-5	6.222	6.124	6.324	445.620	500.000	-10.9
Decachlorobiphenyl	7.532	7.433	7.633	52.490	50.000	5.0
Tetrachloro-m-xylene	2.777	2.678	2.878	47.420	50.000	-5.2