

LAB CHRONICLE

OrderID:	Q3400	OrderDate:	10/20/2025 1:25:31 PM
Client:	ICE Service Group, Inc.	Project:	NWIRP Northrup Grumman Site – Bethpage, NY
Contact:	Dennis D. Morgan, II	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3400-01	FRAC-TANK-0760450	WATER	PCB	8082A	10/20/25	10/22/25	10/23/25	10/20/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.:

Q3400

Order ID:

Q3400

Client:

ICE Service Group, Inc.

Project ID:

NWIRP Northrup Grumman Site – Be

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				Total Concentration:	0.000				



SAMPLE DATA

Report of Analysis

Client:	ICE Service Group, Inc.	Date Collected:	10/20/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY	Date Received:	10/20/25
Client Sample ID:	FRAC-TANK-0760450	SDG No.:	Q3400
Lab Sample ID:	Q3400-01	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	990 mL	Final Vol:	10000 uL
Prep Method:		Test:	PCB
	Prep Date:	10/22/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.098	0.25	0.51	ug/L	10/23/25 11:55	PB170211
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.51	ug/L	10/23/25 11:55	PB170211
11141-16-5	Aroclor-1232	0.25	U	1	0.097	0.25	0.51	ug/L	10/23/25 11:55	PB170211
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.51	ug/L	10/23/25 11:55	PB170211
12672-29-6	Aroclor-1248	0.25	U	1	0.072	0.25	0.51	ug/L	10/23/25 11:55	PB170211
11097-69-1	Aroclor-1254	0.25	U	1	0.095	0.25	0.51	ug/L	10/23/25 11:55	PB170211
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.51	ug/L	10/23/25 11:55	PB170211
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.51	ug/L	10/23/25 11:55	PB170211
11096-82-5	Aroclor-1260	0.25	U	1	0.082	0.25	0.51	ug/L	10/23/25 11:55	PB170211
SURROGATES										
877-09-8	Tetrachloro-m-xylene	23.4			35 - 137		117%	SPK: 20		
2051-24-3	Decachlorobiphenyl	9.38			40 - 135		47%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q3400

Client: ICE Service Group, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114312.D	PIBLK-PO114312.D	Tetrachloro-m-xyl	1	20	18.6	93		60	140
		Decachlorobiphen	1	20	19.2	96		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	19.0	95		60	140
I.BLK-PO114838.D	PIBLK-PO114838.D	Tetrachloro-m-xyl	1	20	20.9	104		60	140
		Decachlorobiphen	1	20	20.8	104		60	140
		Tetrachloro-m-xyl	2	20	18.5	93		60	140
		Decachlorobiphen	2	20	21.6	108		60	140
PB170211BL	PB170211BL	Tetrachloro-m-xyl	1	20	20.6	103		35	137
		Decachlorobiphen	1	20	20.4	102		40	135
		Tetrachloro-m-xyl	2	20	18.3	92		35	137
		Decachlorobiphen	2	20	21.0	105		40	135
I.BLK-PO114853.D	PIBLK-PO114853.D	Tetrachloro-m-xyl	1	20	21.5	107		60	140
		Decachlorobiphen	1	20	19.6	98		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	18.5	92		60	140
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
I.BLK-PQ071090.D	PIBLK-PQ071090.D	Tetrachloro-m-xyl	1	20	20.9	104		60	140
		Decachlorobiphen	1	20	23.7	118		60	140
		Tetrachloro-m-xyl	2	20	19.3	97		60	140
		Decachlorobiphen	2	20	23.1	115		60	140
Q3400-01	FRAC-TANK-0760450	Tetrachloro-m-xyl	1	20	23.4	117		35	137
		Decachlorobiphen	1	20	9.38	47		40	135
		Tetrachloro-m-xyl	2	20	21.7	109		35	137
		Decachlorobiphen	2	20	9.32	47		40	135
I.BLK-PQ071102.D	PIBLK-PQ071102.D	Tetrachloro-m-xyl	1	20	20.6	103		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
		Tetrachloro-m-xyl	2	20	19.1	95		60	140
		Decachlorobiphen	2	20	22.9	114		60	140
I.BLK-PQ071120.D	PIBLK-PQ071120.D	Tetrachloro-m-xyl	1	20	20.3	101		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
		Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	21.9	110		60	140
PB170211BS	PB170211BS	Tetrachloro-m-xyl	1	20	21.6	108		35	137
		Decachlorobiphen	1	20	25.2	126		40	135
		Tetrachloro-m-xyl	2	20	16.9	85		35	137
		Decachlorobiphen	2	20	23.4	117		40	135
PB170211BSD	PB170211BSD	Tetrachloro-m-xyl	1	20	22.4	112		35	137

Surrogate Summary

SDG No.: Q3400

Client: ICE Service Group, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB170211BSD	PB170211BSD	Decachlorobiphen	1	20	25.7	128		40	135
		Tetrachloro-m-xyl	2	20	18.3	91		35	137
		Decachlorobiphen	2	20	24.2	121		40	135
I.BLK-PQ071128.D	PIBLK-PQ071128.D	Tetrachloro-m-xyl	1	20	20.8	104		60	140
		Decachlorobiphen	1	20	24.2	121		60	140
		Tetrachloro-m-xyl	2	20	18.4	92		60	140
		Decachlorobiphen	2	20	22.9	115		60	140



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3400	Analytical Method:	8082A
Client:	ICE Service Group, Inc.	Datafile :	PQ071121.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170211BS (Column 1)	AR1016	5	4.90	ug/L	98				46	129	
	AR1260	5	5.30	ug/L	106				45	134	
PB170211BS (Column 2)	AR1016	5	4.10	ug/L	82				46	129	
	AR1260	5	4.20	ug/L	84				45	134	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3400</u>	Analytical Method:	<u>8082A</u>
Client:	<u>ICE Service Group, Inc.</u>	Datafile :	<u>PQ071122.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170211BSD (Column 1)	AR1016	5	5.30	ug/L	106	8			46	129	20
	AR1260	5	5.40	ug/L	108	2			45	134	20
PB170211BSD (Column 2)	AR1016	5	4.30	ug/L	86	5			46	129	20
	AR1260	5	4.40	ug/L	88	5			45	134	20

4C

PESTICIDE METHOD BLANK SUMMARY

Client ID

PB170211BL

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Lab Sample ID: PB170211BL

Lab File ID: PO114839.D

Matrix: (soil/water) WATER

Extraction: (Type) SEPF

Sulfur Cleanup: (Y/N) N

Date Extracted: 10/22/2025

Date Analyzed (1): 11/03/2025

Date Analyzed (2): 11/03/2025

Time Analyzed (1): 10:20

Time Analyzed (2): 10:20

Instrument ID (1): ECD_O

Instrument ID (2): ECD_O

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
FRAC-TANK-0760450	Q3400-01	PQ071095.D	10/23/2025	10/23/2025
PB170211BS	PB170211BS	PQ071121.D	10/24/2025	10/24/2025
PB170211BSD	PB170211BSD	PQ071122.D	10/24/2025	10/24/2025

COMMENTS: _____



CALIBRATION SUMMARY

A
B
C
D
E
F
G

Calibration Times: 14:14 22:29

LAB FILE ID: RT 1000 = PO114313.D RT 750 = PO114314.D
RT 500 = PO114315.D RT 250 = PO114316.D RT 050 = PO114317.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	9.93	9.93	9.93	9.93	9.93	9.93	9.83	10.03
Tetrachloro-m-xylene	4.36	4.36	4.36	4.36	4.36	4.36	4.26	4.46

A
B
C
D
E
F
G

A
B
C
D
E
F
G

Calibration Times: 14:14 22:29

LAB FILE ID: RT 1000 = PO114313.D RT 750 = PO114314.D
RT 500 = PO114315.D RT 250 = PO114316.D RT 050 = PO114317.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.67	8.67	8.67	8.67	8.67	8.67	8.57	8.77
Tetrachloro-m-xylene	3.66	3.66	3.66	3.66	3.66	3.66	3.56	3.76

A
B
C
D
E
F
G

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ICES02
SDG NO.: Q3400

Calibration Date(s): 10/09/2025 10/09/2025
Calibration Times: 14:14 22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:								
CF 1000 =		PO114313.D		CF 750 =		PO114314.D		
CF 500 =		PO114315.D		CF 250 =		PO114316.D		
CF 050 =		PO114317.D						
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	174192886	179895124	183856868	183322864	180520760	180357700	2
Aroclor-1016-2	(2)	293939915	305561101	305220730	315140900	301624180	304297365	3
Aroclor-1016-3	(3)	165633600	172668037	174260432	182505220	189193940	176852246	5
Aroclor-1016-4	(4)	134425999	138097755	140122574	142368188	132758060	137554515	3
Aroclor-1016-5	(5)	123446822	127769624	128908938	130194584	121008000	126265594	3
Aroclor-1260-1	(1)	213791672	221280347	222839010	218586200	213158120	217931070	2
Aroclor-1260-2	(2)	302622462	312056085	309594538	307167896	311081240	308504444	1
Aroclor-1260-3	(3)	259467311	268042008	265717042	260457100	228668800	256470452	6
Aroclor-1260-4	(4)	230199629	235131995	233767838	232769056	232729060	232919516	1
Aroclor-1260-5	(5)	586428065	605840996	593634282	581137220	533761820	580160477	5
Decachlorobiphenyl		3664039320	3792684880	3728036680	3664938040	3355866400	3641113064	5
Tetrachloro-m-xylene		5506498360	5677498373	5642459220	5552412920	4766000200	5428973815	7
Aroclor-1242-1	(1)	126706434	132179268	135020258	135246300	128666340	131563720	3
Aroclor-1242-2	(2)	212392946	219812959	222351800	226444372	216757700	219551955	2
Aroclor-1242-3	(3)	119144042	124121997	125717214	131545340	116119900	123329699	5
Aroclor-1242-4	(4)	99737299	101870191	102531118	105496736	103337180	102594505	2
Aroclor-1242-5	(5)	103047071	106230547	111781010	123521904	134589100	115833926	11
Decachlorobiphenyl		3803036960	3909619067	3861003240	3833959680	3524921000	3786507989	4
Tetrachloro-m-xylene		5804343740	5928892613	5893981820	5915994040	5454317000	5799505843	3
Aroclor-1248-1	(1)	98567829	102136305	106179528	101725532	83182260	98358291	9
Aroclor-1248-2	(2)	137892609	142528752	146526406	145313792	145562060	143564724	2
Aroclor-1248-3	(3)	151681692	159333787	163420432	159028692	160593320	158811585	3
Aroclor-1248-4	(4)	182190012	190636973	195926736	195051484	215146120	195790265	6
Aroclor-1248-5	(5)	174543681	183607891	190759426	202813192	249246460	200194130	15
Decachlorobiphenyl		3816291730	3937073800	3957474960	3763925080	3735520600	3842057234	3
Tetrachloro-m-xylene		5779379270	6044791093	6117390360	5832290520	5648377200	5884445689	3
Aroclor-1254-1	(1)	171902953	180224452	182326136	179059532	158754700	174453555	6
Aroclor-1254-2	(2)	248603985	257164111	266428592	263066468	238270000	254706631	4
Aroclor-1254-3	(3)	278966197	288101712	293337044	293741552	290556100	288940521	2
Aroclor-1254-4	(4)	232417367	241793461	248310406	242500032	218210100	236646273	5
Aroclor-1254-5	(5)	221022379	230048296	231810568	224607672	189890120	219475807	8
Decachlorobiphenyl		3792373920	3896541893	3936938560	3860958960	3451332800	3787629227	5
Tetrachloro-m-xylene		5822824890	5988302693	6053484060	5960886280	5298842000	5824867985	5
Aroclor-1268-1	(1)	742630944	763047828	755049378	744463164	658875500	732813363	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	632443459	650139448	640918178	630546628	553465960	621502735	6
Aroclor-1268-3	(3)	576556870	593571136	586459800	580217260	523967180	572154449	5
Aroclor-1268-4	(4)	239235256	248654241	247759484	273329844	226227580	247041281	7
Aroclor-1268-5	(5)	1713485078	1761561424	1737860666	1722182616	1634161480	1713850253	3
Decachlorobiphenyl		7124253780	7355062013	7282856320	7142144040	6358111400	7052485511	6
Tetrachloro-m-xylene		6131677880	6338649440	6321767300	6220727560	5845855000	6171735436	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ICES02
SDG NO.: Q3400

Calibration Date(s): 10/09/2025 10/09/2025
Calibration Times: 14:14 22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO114313.D	CF 750 =	PO114314.D			
CF 500 =		PO114315.D	CF 250 =	PO114316.D	CF 050 =	PO114317.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	158257841	168187716	172362186	178585052	164529320	168384423	5
Aroclor-1016-2	(2)	229568804	240612359	245734900	252904664	241358100	242035765	4
Aroclor-1016-3	(3)	119655603	124240307	128221016	131086968	117564320	124153643	5
Aroclor-1016-4	(4)	92316708	96506141	100242646	102698056	90863720	96525454	5
Aroclor-1016-5	(5)	119349433	124629577	128669076	134240076	132389300	127855492	5
Aroclor-1260-1	(1)	215558470	231498535	235626690	245673392	249059400	235483297	6
Aroclor-1260-2	(2)	315603721	330098711	337167224	350489224	338584660	334388708	4
Aroclor-1260-3	(3)	256430713	268520155	278444642	296438324	374872640	294941295	16
Aroclor-1260-4	(4)	211317075	223266731	226148960	232910768	220902880	222909283	4
Aroclor-1260-5	(5)	594775480	624015233	623481606	632255524	590282220	612962013	3
Decachlorobiphenyl		3170001190	3348740240	3399801360	3490868680	3256491400	3333180574	4
Tetrachloro-m-xylene		3928229090	4070663293	4096822280	4116538520	3799017800	4002254197	3
Aroclor-1242-1	(1)	122498620	129064381	131010080	137236660	133254420	130612832	4
Aroclor-1242-2	(2)	175860987	183257420	186752992	191657052	196407220	186787134	4
Aroclor-1242-3	(3)	92492282	96111452	98072652	99600996	92244160	95704308	3
Aroclor-1242-4	(4)	86300671	89905741	92299888	94852936	98359060	92343659	5
Aroclor-1242-5	(5)	113144641	119163080	122745918	127985400	129764100	122560628	6
Decachlorobiphenyl		3517109590	3681723480	3757541820	3835681800	3777005200	3713812378	3
Tetrachloro-m-xylene		4432077070	4412224160	4379772960	4471865640	4248743200	4388936606	2
Aroclor-1248-1	(1)	94570009	99404269	103594360	101758388	93150140	98495433	5
Aroclor-1248-2	(2)	125502158	133300196	137663402	137604864	136198920	134053908	4
Aroclor-1248-3	(3)	130740348	139010419	143306254	142181184	142470360	139541713	4
Aroclor-1248-4	(4)	154697640	163551519	169016486	167510084	167044800	164364106	4
Aroclor-1248-5	(5)	159814127	168355577	173516216	171305900	174900140	169578392	4
Decachlorobiphenyl		3661201490	3833278560	3951800380	3911090520	3937109200	3858896030	3
Tetrachloro-m-xylene		4430508460	4634013587	4677547100	4537709360	4261034200	4508162541	4
Aroclor-1254-1	(1)	247222467	259048601	270361008	279309448	272529160	265694137	5
Aroclor-1254-2	(2)	219192417	228941475	239248508	249551080	254226320	238231960	6
Aroclor-1254-3	(3)	340127494	353950264	366242670	368614488	343101920	354407367	4
Aroclor-1254-4	(4)	246402938	256874435	266498340	267498992	253508600	258156661	3
Aroclor-1254-5	(5)	309302184	322642697	332806976	330987116	305278780	320203551	4
Decachlorobiphenyl		3673134130	3816212200	3925129200	3949127480	3714385800	3815597762	3
Tetrachloro-m-xylene		4426752370	4520704027	4624441060	4560126680	4117729800	4449950787	4
Aroclor-1268-1	(1)	790806319	820672996	823414120	829776224	826084500	818150832	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	668389064	693704507	696355354	697040636	678270940	686752100	2
Aroclor-1268-3	(3)	577681654	598471028	600065844	598539168	577278220	590407183	2
Aroclor-1268-4	(4)	233329938	243075305	244332122	245097676	230283080	239223624	3
Aroclor-1268-5	(5)	1599907631	1653346000	1653968404	1645996908	1596574960	1629958781	2
Decachlorobiphenyl		6770675710	7072875573	7158178740	7244610400	7053861200	7060040325	3
Tetrachloro-m-xylene		4643395640	4793601640	4792629120	4749683120	4526037200	4701069344	2

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ICES02

Lab Code: ACE SDG NO.: Q3400

Instrument ID: ECD_O Date(s) Analyzed: 10/09/2025 10/09/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.56	4.46	4.66	65029800
		2	4.64	4.54	4.74	45851400
		3	4.71	4.61	4.81	150477000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.71	4.61	4.81	122061000
		2	5.23	5.13	5.33	54066000
		3	5.52	5.42	5.62	121355000
		4	5.67	5.57	5.77	54031600
		5	5.76	5.66	5.86	37348800
Aroclor-1262	500	1	7.69	7.59	7.79	365212000
		2	8.21	8.11	8.31	604682000
		3	8.52	8.42	8.62	365448000
		4	8.60	8.50	8.70	263614000
		5	9.22	9.12	9.32	214854000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ICES02

Lab Code: ACE SDG NO.: Q3400

Instrument ID: ECD_O Date(s) Analyzed: 10/09/2025 10/09/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.87	3.77	3.97	54832000
		2	3.96	3.86	4.06	40465400
		3	4.03	3.93	4.13	128449000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.03	3.93	4.13	104652000
		2	4.76	4.66	4.86	102880000
		3	4.94	4.84	5.04	51237200
		4	5.02	4.92	5.12	43634600
		5	5.19	5.09	5.29	48052000
Aroclor-1262	500	1	6.78	6.68	6.88	425654000
		2	7.28	7.18	7.38	673478000
		3	7.56	7.46	7.66	276888000
		4	7.63	7.53	7.73	456530000
		5	8.12	8.02	8.22	210780000

A
B
C
D
E
F
G

Calibration Times: 08:57 16:43

LAB FILE ID: RT 1000 = PQ070973.D RT 750 = PQ070974.D
RT 500 = PQ070975.D RT 250 = PQ070976.D RT 050 = PQ070977.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	8.50	8.50	8.50	8.50	8.50	8.50	8.40	8.60
Tetrachloro-m-xylene	3.40	3.40	3.40	3.40	3.40	3.40	3.30	3.50

A

B

C

D

E

F

G

A
B
C
D
E
F
G

Calibration Times: 08:57 16:43

LAB FILE ID: RT 1000 = PQ070973.D RT 750 = PQ070974.D
RT 500 = PQ070975.D RT 250 = PQ070976.D RT 050 = PQ070977.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	7.53	7.53	7.53	7.53	7.53	7.53	7.43	7.63
Tetrachloro-m-xylene	2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88

A
B
C
D
E
F
G

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q

Contract: ICES02
SDG NO.: Q3400

Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PQ070973.D</u>		CF 750 = <u>PQ070974.D</u>			
CF 500 = <u>PQ070975.D</u>		CF 250 = <u>PQ070976.D</u>		CF 050 = <u>PQ070977.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	152632258	161381460	169566884	175492100	161647220	164143984
Aroclor-1016-2	(2)	235274296	256838307	264605432	271518004	253235720	256294352
Aroclor-1016-3	(3)	145675938	156782704	164416996	169596776	159119400	159118363
Aroclor-1016-4	(4)	120162951	129378400	134848252	138303760	126953020	129929277
Aroclor-1016-5	(5)	117231883	128545740	133336348	136306768	128593380	128802824
Aroclor-1260-1	(1)	208928914	234195365	241877518	247859876	239493660	234471067
Aroclor-1260-2	(2)	243695794	275266251	287710272	296720404	279652200	276608984
Aroclor-1260-3	(3)	178017013	196424459	204653606	209590628	207855920	199308325
Aroclor-1260-4	(4)	191556315	214263107	223690164	228814336	216092480	214883280
Aroclor-1260-5	(5)	400986941	436898331	447507618	451939420	431553500	433777162
Decachlorobiphenyl		2603997920	2888430733	3017037660	3097201720	2913248400	2903983287
Tetrachloro-m-xylene		4048991050	4420325093	4447782420	4497118040	4337178200	4350278961
Aroclor-1242-1	(1)	117929900	120501580	121917368	124177888	110233240	118951995
Aroclor-1242-2	(2)	179238730	188762288	194471730	197719720	193371880	190712870
Aroclor-1242-3	(3)	111298009	117265485	120728326	123754708	121434700	118896246
Aroclor-1242-4	(4)	91702396	97376287	98944784	101196952	108558560	99555796
Aroclor-1242-5	(5)	63322787	67097277	82718820	102665812	77804800	78721899
Decachlorobiphenyl		3082568640	3243406933	3352805760	3359769360	3286790800	3265068299
Tetrachloro-m-xylene		4441893010	4577290613	4623065280	4605855840	4702516200	4590124189
Aroclor-1248-1	(1)	89951061	92802913	99278242	97399904	67891380	89464700
Aroclor-1248-2	(2)	126849870	131764987	138900528	142257364	142462060	136446962
Aroclor-1248-3	(3)	152980867	157544657	166146850	167481096	162764480	161383590
Aroclor-1248-4	(4)	167077290	173126060	182827440	186713196	157721500	173493097
Aroclor-1248-5	(5)	163667698	168136284	132211082	121389384	159185140	148917918
Decachlorobiphenyl		3177348440	3248518907	3402766540	3434790520	3353916400	3323468161
Tetrachloro-m-xylene		4546508980	4599154480	4739413180	4685828360	4518221600	4617825320
Aroclor-1254-1	(1)	186317660	198212097	200308250	222421948	174131700	196278331
Aroclor-1254-2	(2)	272465681	301191412	300558232	318683640	259516480	290483089
Aroclor-1254-3	(3)	294493285	320137152	318530722	335107472	267769020	307207530
Aroclor-1254-4	(4)	212636724	232647637	229798130	237628972	184991000	219540493
Aroclor-1254-5	(5)	229095618	241104957	242576814	249415576	197285560	231895705
Decachlorobiphenyl		3135993520	3237057933	3300114540	3432442200	2724290800	3165979799
Tetrachloro-m-xylene		4449551480	4584042107	4604497820	4625575480	3701892800	4393111937
Aroclor-1268-1	(1)	555593291	572438855	583365466	598015136	603747960	582632142

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	510325576	524057337	534075228	546100468	539884140	530888550	3
Aroclor-1268-3	(3)	416538407	427532203	436646098	443325192	442459640	433300308	3
Aroclor-1268-4	(4)	179373252	185531188	189624920	193669728	190155160	187670850	3
Aroclor-1268-5	(5)	1283167757	1308010297	1321380362	1326514116	1318957680	1311606042	1
Decachlorobiphenyl		5757514290	5913646120	6037365900	6192444040	6223042400	6024802550	3
Tetrachloro-m-xylene		4734874420	4871269707	4916225920	4964521920	4932484000	4883875193	2

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_Q

Contract: ICES02
SDG NO.: Q3400

Calibration Date(s): 10/10/2025 10/10/2025
Calibration Times: 08:57 16:43

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:								
CF 1000 = <u>PQ070973.D</u>		CF 750 = <u>PQ070974.D</u>		CF 500 = <u>PQ070975.D</u>		CF 250 = <u>PQ070976.D</u>		
CF 500 = <u>PQ070975.D</u>		CF 250 = <u>PQ070976.D</u>		CF 050 = <u>PQ070977.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	198094739	217218485	228189888	237783008	238723760	224001976	8
Aroclor-1016-2	(2)	306974778	336928535	345611618	359075648	352024660	340123048	6
Aroclor-1016-3	(3)	164197912	180535471	187035142	195263012	182262860	181858879	6
Aroclor-1016-4	(4)	139678855	154088441	161162796	170890484	161242760	157412667	7
Aroclor-1016-5	(5)	171310856	187991799	196963118	205633060	197310680	191841903	7
Aroclor-1260-1	(1)	322104344	352779140	365742242	382158164	379245720	360405922	7
Aroclor-1260-2	(2)	392921251	432078265	450339936	466790948	457838380	439993756	7
Aroclor-1260-3	(3)	374441100	409997701	423595070	436050024	430913160	414999411	6
Aroclor-1260-4	(4)	324169048	356082777	368301826	380117112	370143240	359762801	6
Aroclor-1260-5	(5)	773923407	822481391	873089514	880740576	851175200	840282018	5
Decachlorobiphenyl		4511855410	4959881947	5165429800	5359918520	5250729800	5049563095	7
Tetrachloro-m-xylene		4790824740	5183285133	5313707940	5302586120	4809696600	5080020107	5
Aroclor-1242-1	(1)	149120136	160366853	160219200	165386980	179219740	162862582	7
Aroclor-1242-2	(2)	228406508	237737935	244055020	248772624	247903320	241375081	4
Aroclor-1242-3	(3)	123267302	131780304	132089372	135429308	136535900	131820437	4
Aroclor-1242-4	(4)	107929318	113409847	120761922	128344628	102401880	114569519	9
Aroclor-1242-5	(5)	175313049	196019013	175387368	167275268	156704980	174139936	8
Decachlorobiphenyl		5194557400	5483323560	5616478880	5698756520	6104877800	5619598832	6
Tetrachloro-m-xylene		5511103540	5712989200	5725543240	5446452960	5436162600	5566450308	3
Aroclor-1248-1	(1)	118147108	120641916	127443132	132200888	134031720	126492953	6
Aroclor-1248-2	(2)	172958749	177705395	189325858	196101732	206725040	188563355	7
Aroclor-1248-3	(3)	167593672	172934769	178216680	182062792	197467520	179655087	6
Aroclor-1248-4	(4)	207432918	214191484	230649658	232116576	237264740	224331075	6
Aroclor-1248-5	(5)	211078475	216720032	243430792	250826296	242157620	232842643	8
Decachlorobiphenyl		5357284330	5460780160	5730809660	5844375240	6130897000	5704829278	5
Tetrachloro-m-xylene		5499268270	5571294053	5722936200	5664492400	5408993600	5573396905	2
Aroclor-1254-1	(1)	326099704	341297420	345505280	377987212	295169840	337211891	9
Aroclor-1254-2	(2)	283781194	302336180	304849560	326229636	268575420	297154398	7
Aroclor-1254-3	(3)	455841488	477545104	484627186	501619228	415693040	467065209	7
Aroclor-1254-4	(4)	291449906	303039177	308578712	317073776	256115840	295251482	8
Aroclor-1254-5	(5)	417470468	427079459	439495106	452352324	365350180	420349507	8
Decachlorobiphenyl		5360763820	5577066440	5716894620	5972893640	5256420800	5576807864	5
Tetrachloro-m-xylene		5311856950	5451428933	5484489360	5522984680	4238156400	5201783265	10
Aroclor-1268-1	(1)	931455367	961712979	973652708	974423320	1030986200	974446115	4

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	858872879	891158693	900594916	906701084	936258500	898717214	3
Aroclor-1268-3	(3)	746016914	769745548	779615614	791820108	821422560	781724149	4
Aroclor-1268-4	(4)	330162118	339473573	345636662	356393996	350955500	344524370	3
Aroclor-1268-5	(5)	2374681378	2422263161	2432472106	2427996872	2444443520	2420371407	1
Decachlorobiphenyl		10401085910	10736618840	10941521580	11269357960	11896046400	11048926138	5
Tetrachloro-m-xylene		5581659480	5729850200	5733063640	5828463760	5669436400	5708494696	2

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ICES02

Lab Code: ACE SDG NO.: Q3400

Instrument ID: ECD_Q Date(s) Analyzed: 10/10/2025 10/10/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.59	3.49	3.69	59190800
		2	3.67	3.57	3.77	43146800
		3	3.74	3.64	3.84	133548000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.74	3.64	3.84	108948000
		2	4.23	4.13	4.33	57502800
		3	4.50	4.40	4.60	107099000
		4	4.65	4.55	4.75	54014200
		5	4.79	4.69	4.89	43713200
Aroclor-1262	500	1	6.84	6.74	6.94	274018000
		2	7.15	7.05	7.25	498476000
		3	7.42	7.32	7.52	328660000
		4	7.49	7.39	7.59	252340000
		5	7.99	7.89	8.09	164350000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ICES02

Lab Code: ACE SDG NO.: Q3400

Instrument ID: ECD_Q Date(s) Analyzed: 10/10/2025 10/10/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	2.98	2.88	3.08	73819200
		2	3.05	2.95	3.15	60003200
		3	3.12	3.02	3.22	171184000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	3.12	3.02	3.22	139342000
		2	3.79	3.69	3.89	138900000
		3	3.95	3.85	4.05	73450200
		4	4.04	3.94	4.14	63034400
		5	4.20	4.10	4.30	71607200
Aroclor-1262	500	1	5.73	5.63	5.83	452486000
		2	5.98	5.88	6.08	427702000
		3	6.50	6.40	6.60	332708000
		4	6.56	6.46	6.66	616820000
		5	7.05	6.95	7.15	302986000

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 11/03/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 08:49

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.58	5.48	5.68	0.01
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.08	6.98	7.18	0.01
Aroclor-1260-2	(2)	7.33	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.69	7.59	7.79	0.01
Aroclor-1260-4	(4)	7.90	7.91	7.81	8.01	0.01
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.36	4.26	4.46	0.01
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 11/03/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 08:49

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1016-2	(2)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-3	(3)	4.92	4.94	4.84	5.04	0.02
Aroclor-1016-4	(4)	4.96	4.98	4.88	5.08	0.02
Aroclor-1016-5	(5)	5.18	5.19	5.09	5.29	0.01
Aroclor-1260-1	(1)	6.21	6.23	6.13	6.33	0.02
Aroclor-1260-2	(2)	6.39	6.41	6.31	6.51	0.02
Aroclor-1260-3	(3)	6.55	6.57	6.47	6.67	0.02
Aroclor-1260-4	(4)	7.02	7.04	6.94	7.14	0.02
Aroclor-1260-5	(5)	7.26	7.28	7.18	7.38	0.02
Tetrachloro-m-xylene		3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl		8.64	8.67	8.57	8.77	0.03

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114834.D **Time Analyzed:** 08:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.488	5.392	5.592	558.010	500.000	11.6
Aroclor-1016-2	5.509	5.414	5.614	532.430	500.000	6.5
Aroclor-1016-3	5.571	5.476	5.676	532.150	500.000	6.4
Aroclor-1016-4	5.667	5.571	5.771	564.480	500.000	12.9
Aroclor-1016-5	5.958	5.863	6.063	564.580	500.000	12.9
Aroclor-1260-1	7.071	6.975	7.175	552.040	500.000	10.4
Aroclor-1260-2	7.325	7.229	7.429	550.680	500.000	10.1
Aroclor-1260-3	7.682	7.585	7.785	532.380	500.000	6.5
Aroclor-1260-4	7.904	7.807	8.007	534.270	500.000	6.9
Aroclor-1260-5	8.210	8.113	8.313	510.800	500.000	2.2
Decachlorobiphenyl	9.931	9.834	10.034	51.810	50.000	3.6
Tetrachloro-m-xylene	4.349	4.255	4.455	54.380	50.000	8.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114834.D **Time Analyzed:** 08:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.643	4.843	467.830	500.000	-6.4
Aroclor-1016-2	4.748	4.662	4.862	471.560	500.000	-5.7
Aroclor-1016-3	4.922	4.836	5.036	501.070	500.000	0.2
Aroclor-1016-4	4.963	4.879	5.079	512.000	500.000	2.4
Aroclor-1016-5	5.176	5.092	5.292	502.560	500.000	0.5
Aroclor-1260-1	6.206	6.125	6.325	519.290	500.000	3.9
Aroclor-1260-2	6.394	6.313	6.513	530.830	500.000	6.2
Aroclor-1260-3	6.547	6.466	6.666	501.170	500.000	0.2
Aroclor-1260-4	7.017	6.939	7.139	580.950	500.000	16.2
Aroclor-1260-5	7.258	7.179	7.379	558.650	500.000	11.7
Decachlorobiphenyl	8.642	8.571	8.771	51.040	50.000	2.1
Tetrachloro-m-xylene	3.651	3.560	3.760	46.540	50.000	-6.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 11/03/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 14:33

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.49	5.49	5.39	5.59	0.00
Aroclor-1016-2	(2)	5.51	5.51	5.41	5.61	0.00
Aroclor-1016-3	(3)	5.57	5.58	5.48	5.68	0.01
Aroclor-1016-4	(4)	5.67	5.67	5.57	5.77	0.00
Aroclor-1016-5	(5)	5.96	5.96	5.86	6.06	0.00
Aroclor-1260-1	(1)	7.07	7.08	6.98	7.18	0.01
Aroclor-1260-2	(2)	7.32	7.33	7.23	7.43	0.01
Aroclor-1260-3	(3)	7.68	7.69	7.59	7.79	0.01
Aroclor-1260-4	(4)	7.90	7.91	7.81	8.01	0.01
Aroclor-1260-5	(5)	8.21	8.21	8.11	8.31	0.00
Tetrachloro-m-xylene		4.35	4.36	4.26	4.46	0.01
Decachlorobiphenyl		9.93	9.93	9.83	10.03	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 11/03/2025

Initial Calibration Date(s): 10/09/2025

10/09/2025

Continuing Calib Time: 14:33

Initial Calibration Time(s): 14:14

22:29

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1016-2	(2)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-3	(3)	4.92	4.94	4.84	5.04	0.02
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.18	5.19	5.09	5.29	0.01
Aroclor-1260-1	(1)	6.21	6.23	6.13	6.33	0.02
Aroclor-1260-2	(2)	6.40	6.41	6.31	6.51	0.02
Aroclor-1260-3	(3)	6.55	6.57	6.47	6.67	0.02
Aroclor-1260-4	(4)	7.02	7.04	6.94	7.14	0.02
Aroclor-1260-5	(5)	7.26	7.28	7.18	7.38	0.02
Tetrachloro-m-xylene		3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl		8.64	8.67	8.57	8.77	0.03

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114849.D **Time Analyzed:** 14:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.487	5.392	5.592	482.500	500.000	-3.5
Aroclor-1016-2	5.509	5.414	5.614	451.400	500.000	-9.7
Aroclor-1016-3	5.570	5.476	5.676	442.490	500.000	-11.5
Aroclor-1016-4	5.666	5.571	5.771	489.320	500.000	-2.1
Aroclor-1016-5	5.957	5.863	6.063	524.730	500.000	4.9
Aroclor-1260-1	7.071	6.975	7.175	474.850	500.000	-5.0
Aroclor-1260-2	7.324	7.229	7.429	457.380	500.000	-8.5
Aroclor-1260-3	7.681	7.585	7.785	431.740	500.000	-13.7
Aroclor-1260-4	7.903	7.807	8.007	443.690	500.000	-11.3
Aroclor-1260-5	8.210	8.113	8.313	430.500	500.000	-13.9
Decachlorobiphenyl	9.932	9.834	10.034	45.480	50.000	-9.0
Tetrachloro-m-xylene	4.349	4.255	4.455	53.590	50.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/09/2025 10/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 11/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114849.D **Time Analyzed:** 14:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.729	4.643	4.843	450.780	500.000	-9.8
Aroclor-1016-2	4.748	4.662	4.862	453.430	500.000	-9.3
Aroclor-1016-3	4.923	4.836	5.036	423.360	500.000	-15.3
Aroclor-1016-4	4.965	4.879	5.079	442.140	500.000	-11.6
Aroclor-1016-5	5.177	5.092	5.292	472.480	500.000	-5.5
Aroclor-1260-1	6.207	6.125	6.325	436.270	500.000	-12.7
Aroclor-1260-2	6.395	6.313	6.513	431.390	500.000	-13.7
Aroclor-1260-3	6.547	6.466	6.666	406.860	500.000	-18.6
Aroclor-1260-4	7.019	6.939	7.139	440.280	500.000	-11.9
Aroclor-1260-5	7.259	7.179	7.379	451.450	500.000	-9.7
Decachlorobiphenyl	8.643	8.571	8.771	40.090	50.000	-19.8
Tetrachloro-m-xylene	3.652	3.560	3.760	47.250	50.000	-5.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 08:59

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 08:59

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071086.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.481	4.380	4.580	536.950	500.000	7.4
Aroclor-1016-2	4.499	4.399	4.599	535.300	500.000	7.1
Aroclor-1016-3	4.557	4.456	4.656	544.430	500.000	8.9
Aroclor-1016-4	4.647	4.546	4.746	540.610	500.000	8.1
Aroclor-1016-5	4.931	4.831	5.031	533.730	500.000	6.7
Aroclor-1260-1	6.021	5.919	6.119	523.570	500.000	4.7
Aroclor-1260-2	6.277	6.175	6.375	524.480	500.000	4.9
Aroclor-1260-3	6.628	6.526	6.726	542.650	500.000	8.5
Aroclor-1260-4	6.842	6.740	6.940	551.920	500.000	10.4
Aroclor-1260-5	7.147	7.045	7.245	548.300	500.000	9.7
Decachlorobiphenyl	8.499	8.396	8.596	54.810	50.000	9.6
Tetrachloro-m-xylene	3.400	3.300	3.500	53.710	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071086.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	448.240	500.000	-10.4
Aroclor-1016-2	3.792	3.693	3.893	444.160	500.000	-11.2
Aroclor-1016-3	3.952	3.853	4.053	458.590	500.000	-8.3
Aroclor-1016-4	4.000	3.902	4.102	462.320	500.000	-7.5
Aroclor-1016-5	4.195	4.097	4.297	456.210	500.000	-8.8
Aroclor-1260-1	5.187	5.089	5.289	455.010	500.000	-9.0
Aroclor-1260-2	5.376	5.278	5.478	454.650	500.000	-9.1
Aroclor-1260-3	5.518	5.419	5.619	457.030	500.000	-8.6
Aroclor-1260-4	5.977	5.879	6.079	459.370	500.000	-8.1
Aroclor-1260-5	6.222	6.124	6.324	445.620	500.000	-10.9
Decachlorobiphenyl	7.532	7.433	7.633	52.490	50.000	5.0
Tetrachloro-m-xylene	2.777	2.678	2.878	47.420	50.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 13:37

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/23/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 13:37

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071098.D **Time Analyzed:** 13:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	547.800	500.000	9.6
Aroclor-1016-2	4.498	4.399	4.599	544.520	500.000	8.9
Aroclor-1016-3	4.556	4.456	4.656	550.240	500.000	10.0
Aroclor-1016-4	4.646	4.546	4.746	556.420	500.000	11.3
Aroclor-1016-5	4.931	4.831	5.031	553.620	500.000	10.7
Aroclor-1260-1	6.020	5.919	6.119	549.350	500.000	9.9
Aroclor-1260-2	6.276	6.175	6.375	554.480	500.000	10.9
Aroclor-1260-3	6.627	6.526	6.726	559.460	500.000	11.9
Aroclor-1260-4	6.841	6.740	6.940	571.200	500.000	14.2
Aroclor-1260-5	7.146	7.045	7.245	565.790	500.000	13.2
Decachlorobiphenyl	8.498	8.396	8.596	56.680	50.000	13.4
Tetrachloro-m-xylene	3.399	3.300	3.500	54.440	50.000	8.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071098.D **Time Analyzed:** 13:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	453.960	500.000	-9.2
Aroclor-1016-2	3.792	3.693	3.893	452.590	500.000	-9.5
Aroclor-1016-3	3.951	3.853	4.053	464.930	500.000	-7.0
Aroclor-1016-4	4.000	3.902	4.102	469.520	500.000	-6.1
Aroclor-1016-5	4.195	4.097	4.297	470.610	500.000	-5.9
Aroclor-1260-1	5.188	5.089	5.289	476.540	500.000	-4.7
Aroclor-1260-2	5.377	5.278	5.478	468.050	500.000	-6.4
Aroclor-1260-3	5.518	5.419	5.619	457.860	500.000	-8.4
Aroclor-1260-4	5.978	5.879	6.079	473.740	500.000	-5.3
Aroclor-1260-5	6.222	6.124	6.324	469.640	500.000	-6.1
Decachlorobiphenyl	7.532	7.433	7.633	55.130	50.000	10.3
Tetrachloro-m-xylene	2.776	2.678	2.878	49.120	50.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 08:52

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 08:52

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071119.D **Time Analyzed:** 08:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	529.290	500.000	5.9
Aroclor-1016-2	4.499	4.399	4.599	527.760	500.000	5.6
Aroclor-1016-3	4.557	4.456	4.656	540.370	500.000	8.1
Aroclor-1016-4	4.647	4.546	4.746	539.790	500.000	8.0
Aroclor-1016-5	4.931	4.831	5.031	531.950	500.000	6.4
Aroclor-1260-1	6.021	5.919	6.119	517.090	500.000	3.4
Aroclor-1260-2	6.277	6.175	6.375	516.680	500.000	3.3
Aroclor-1260-3	6.628	6.526	6.726	542.990	500.000	8.6
Aroclor-1260-4	6.842	6.740	6.940	552.950	500.000	10.6
Aroclor-1260-5	7.147	7.045	7.245	543.020	500.000	8.6
Decachlorobiphenyl	8.499	8.396	8.596	54.590	50.000	9.2
Tetrachloro-m-xylene	3.400	3.300	3.500	52.510	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071119.D **Time Analyzed:** 08:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	423.330	500.000	-15.3
Aroclor-1016-2	3.791	3.693	3.893	423.970	500.000	-15.2
Aroclor-1016-3	3.952	3.853	4.053	438.320	500.000	-12.3
Aroclor-1016-4	4.000	3.902	4.102	444.030	500.000	-11.2
Aroclor-1016-5	4.195	4.097	4.297	436.450	500.000	-12.7
Aroclor-1260-1	5.187	5.089	5.289	433.170	500.000	-13.4
Aroclor-1260-2	5.376	5.278	5.478	431.050	500.000	-13.8
Aroclor-1260-3	5.517	5.419	5.619	433.100	500.000	-13.4
Aroclor-1260-4	5.977	5.879	6.079	438.310	500.000	-12.3
Aroclor-1260-5	6.222	6.124	6.324	424.780	500.000	-15.0
Decachlorobiphenyl	7.532	7.433	7.633	51.170	50.000	2.3
Tetrachloro-m-xylene	2.776	2.678	2.878	45.960	50.000	-8.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 10:36

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.48	4.48	4.38	4.58	0.00
Aroclor-1016-2	(2)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-3	(3)	4.56	4.56	4.46	4.66	0.00
Aroclor-1016-4	(4)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-5	(5)	4.93	4.93	4.83	5.03	0.00
Aroclor-1260-1	(1)	6.02	6.02	5.92	6.12	0.00
Aroclor-1260-2	(2)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-3	(3)	6.63	6.63	6.53	6.73	0.00
Aroclor-1260-4	(4)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-5	(5)	7.15	7.15	7.05	7.25	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.50	8.50	8.40	8.60	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ICES02

Lab Code: ACE

SDG NO.: Q3400

Continuing Calib Date: 10/24/2025

Initial Calibration Date(s): 10/10/2025 10/10/2025

Continuing Calib Time: 10:36

Initial Calibration Time(s): 08:57 16:43

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	3.78	3.78	3.68	3.88	0.00
Aroclor-1016-2	(2)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-3	(3)	3.95	3.95	3.85	4.05	0.00
Aroclor-1016-4	(4)	4.00	4.00	3.90	4.10	0.00
Aroclor-1016-5	(5)	4.20	4.20	4.10	4.30	0.00
Aroclor-1260-1	(1)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-2	(2)	5.38	5.38	5.28	5.48	0.00
Aroclor-1260-3	(3)	5.52	5.52	5.42	5.62	0.00
Aroclor-1260-4	(4)	5.98	5.98	5.88	6.08	0.00
Aroclor-1260-5	(5)	6.22	6.22	6.12	6.32	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00
Decachlorobiphenyl		7.53	7.53	7.43	7.63	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071124.D **Time Analyzed:** 10:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.480	4.380	4.580	535.180	500.000	7.0
Aroclor-1016-2	4.499	4.399	4.599	521.020	500.000	4.2
Aroclor-1016-3	4.556	4.456	4.656	530.020	500.000	6.0
Aroclor-1016-4	4.646	4.546	4.746	529.540	500.000	5.9
Aroclor-1016-5	4.930	4.831	5.031	526.950	500.000	5.4
Aroclor-1260-1	6.020	5.919	6.119	515.670	500.000	3.1
Aroclor-1260-2	6.276	6.175	6.375	516.240	500.000	3.2
Aroclor-1260-3	6.627	6.526	6.726	541.410	500.000	8.3
Aroclor-1260-4	6.840	6.740	6.940	553.630	500.000	10.7
Aroclor-1260-5	7.147	7.045	7.245	544.110	500.000	8.8
Decachlorobiphenyl	8.497	8.396	8.596	54.590	50.000	9.2
Tetrachloro-m-xylene	3.400	3.300	3.500	52.250	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ICES02
Lab Code: ACE **SDG NO.:** Q3400
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/10/2025 10/10/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ071124.D **Time Analyzed:** 10:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.776	3.678	3.878	427.210	500.000	-14.6
Aroclor-1016-2	3.792	3.693	3.893	424.110	500.000	-15.2
Aroclor-1016-3	3.951	3.853	4.053	434.870	500.000	-13.0
Aroclor-1016-4	4.000	3.902	4.102	441.890	500.000	-11.6
Aroclor-1016-5	4.195	4.097	4.297	439.520	500.000	-12.1
Aroclor-1260-1	5.187	5.089	5.289	435.740	500.000	-12.9
Aroclor-1260-2	5.376	5.278	5.478	433.990	500.000	-13.2
Aroclor-1260-3	5.517	5.419	5.619	438.480	500.000	-12.3
Aroclor-1260-4	5.978	5.879	6.079	442.200	500.000	-11.6
Aroclor-1260-5	6.222	6.124	6.324	441.170	500.000	-11.8
Decachlorobiphenyl	7.531	7.433	7.633	51.410	50.000	2.8
Tetrachloro-m-xylene	2.776	2.678	2.878	46.210	50.000	-7.6

Analytical Sequence

Client: ICE Service Group, Inc.	SDG No.: Q3400
Project: NWIRP Northrup Grumman Site – Bethpage	Instrument ID: ECD_O
GC Column: ZB-MR1	ID: 0.32 (mm) Inst. Calib. Date(s): 10/09/2025 10/09/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/09/2025	13:55	PO114312.D	9.93	4.36
AR1660ICC1000	AR1660ICC1000	10/09/2025	14:14	PO114313.D	9.93	4.36
AR1660ICC750	AR1660ICC750	10/09/2025	14:32	PO114314.D	9.93	4.36
AR1660ICC500	AR1660ICC500	10/09/2025	14:50	PO114315.D	9.93	4.36
AR1660ICC250	AR1660ICC250	10/09/2025	15:09	PO114316.D	9.93	4.36
AR1660ICC050	AR1660ICC050	10/09/2025	15:27	PO114317.D	9.93	4.36
AR1221ICC500	AR1221ICC500	10/09/2025	15:45	PO114318.D	9.93	4.36
AR1232ICC500	AR1232ICC500	10/09/2025	16:04	PO114319.D	9.93	4.36
AR1242ICC1000	AR1242ICC1000	10/09/2025	16:22	PO114320.D	9.93	4.36
AR1242ICC750	AR1242ICC750	10/09/2025	16:40	PO114321.D	9.93	4.36
AR1242ICC500	AR1242ICC500	10/09/2025	16:59	PO114322.D	9.93	4.36
AR1242ICC250	AR1242ICC250	10/09/2025	17:17	PO114323.D	9.93	4.36
AR1242ICC050	AR1242ICC050	10/09/2025	17:35	PO114324.D	9.93	4.36
AR1248ICC1000	AR1248ICC1000	10/09/2025	17:54	PO114325.D	9.93	4.36
AR1248ICC750	AR1248ICC750	10/09/2025	18:12	PO114326.D	9.93	4.36
AR1248ICC500	AR1248ICC500	10/09/2025	18:30	PO114327.D	9.93	4.36
AR1248ICC250	AR1248ICC250	10/09/2025	18:49	PO114328.D	9.93	4.36
AR1248ICC050	AR1248ICC050	10/09/2025	19:07	PO114329.D	9.93	4.36
AR1254ICC1000	AR1254ICC1000	10/09/2025	19:26	PO114330.D	9.93	4.36
AR1254ICC750	AR1254ICC750	10/09/2025	19:44	PO114331.D	9.93	4.36
AR1254ICC500	AR1254ICC500	10/09/2025	20:02	PO114332.D	9.93	4.36
AR1254ICC250	AR1254ICC250	10/09/2025	20:21	PO114333.D	9.93	4.36
AR1254ICC050	AR1254ICC050	10/09/2025	20:39	PO114334.D	9.93	4.36
AR1262ICC500	AR1262ICC500	10/09/2025	20:57	PO114335.D	9.93	4.36
AR1268ICC1000	AR1268ICC1000	10/09/2025	21:16	PO114336.D	9.93	4.36
AR1268ICC750	AR1268ICC750	10/09/2025	21:34	PO114337.D	9.93	4.36
AR1268ICC500	AR1268ICC500	10/09/2025	21:52	PO114338.D	9.93	4.36
AR1268ICC250	AR1268ICC250	10/09/2025	22:11	PO114339.D	9.93	4.36
AR1268ICC050	AR1268ICC050	10/09/2025	22:29	PO114340.D	9.93	4.36
AR1660CCC500	AR1660CCC500	11/03/2025	08:49	PO114834.D	9.93	4.35
IBLK	IBLK	11/03/2025	10:01	PO114838.D	9.93	4.35
PB170211BL	PB170211BL	11/03/2025	10:20	PO114839.D	9.93	4.35
AR1660CCC500	AR1660CCC500	11/03/2025	14:33	PO114849.D	9.93	4.35
IBLK	IBLK	11/03/2025	15:44	PO114853.D	9.93	4.35
IBLK	IBLK	10/10/2025	08:42	PQ070972.D	8.50	3.40
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	8.50	3.40
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	8.50	3.40
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	8.50	3.40
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	8.50	3.40
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	8.50	3.40
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	8.50	3.40
AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	8.50	3.40

Analytical Sequence

AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	8.50	3.40
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	8.50	3.40
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	8.50	3.40
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	8.50	3.40
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	8.50	3.40
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	8.50	3.40
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	8.50	3.40
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	8.50	3.40
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	8.50	3.40
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	8.50	3.40
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	8.50	3.40
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	8.50	3.40
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	8.50	3.40
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	8.50	3.40
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	8.50	3.40
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	8.50	3.40
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	8.50	3.40
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	8.50	3.40
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	8.50	3.40
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	8.50	3.40
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/23/2025	08:59	PQ071086.D	8.50	3.40
IBLK	IBLK	10/23/2025	10:42	PQ071090.D	8.50	3.40
FRAC-TANK-0760450	Q3400-01	10/23/2025	11:55	PQ071095.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/23/2025	13:37	PQ071098.D	8.50	3.40
IBLK	IBLK	10/23/2025	14:35	PQ071102.D	8.50	3.40
AR1660CCC500	AR1660CCC500	10/24/2025	08:52	PQ071119.D	8.50	3.40
IBLK	IBLK	10/24/2025	09:06	PQ071120.D	8.50	3.40
PB170211BS	PB170211BS	10/24/2025	09:34	PQ071121.D	8.51	3.41
PB170211BSD	PB170211BSD	10/24/2025	09:51	PQ071122.D	8.51	3.41
AR1660CCC500	AR1660CCC500	10/24/2025	10:36	PQ071124.D	8.50	3.40
IBLK	IBLK	10/24/2025	11:46	PQ071128.D	8.51	3.41

Analytical Sequence

Client: ICE Service Group, Inc.	SDG No.: Q3400
Project: NWIRP Northrup Grumman Site – Bethpage	Instrument ID: ECD_O
GC Column: ZB-MR2	ID: 0.32 (mm) Inst. Calib. Date(s): 10/09/2025 10/09/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

CLIENT ID	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	10/09/2025	13:55	PO114312.D	8.67	3.66
AR1660ICC1000	AR1660ICC1000	10/09/2025	14:14	PO114313.D	8.67	3.66
AR1660ICC750	AR1660ICC750	10/09/2025	14:32	PO114314.D	8.67	3.66
AR1660ICC500	AR1660ICC500	10/09/2025	14:50	PO114315.D	8.67	3.66
AR1660ICC250	AR1660ICC250	10/09/2025	15:09	PO114316.D	8.67	3.66
AR1660ICC050	AR1660ICC050	10/09/2025	15:27	PO114317.D	8.67	3.66
AR1221ICC500	AR1221ICC500	10/09/2025	15:45	PO114318.D	8.67	3.66
AR1232ICC500	AR1232ICC500	10/09/2025	16:04	PO114319.D	8.67	3.66
AR1242ICC1000	AR1242ICC1000	10/09/2025	16:22	PO114320.D	8.67	3.66
AR1242ICC750	AR1242ICC750	10/09/2025	16:40	PO114321.D	8.67	3.66
AR1242ICC500	AR1242ICC500	10/09/2025	16:59	PO114322.D	8.67	3.66
AR1242ICC250	AR1242ICC250	10/09/2025	17:17	PO114323.D	8.67	3.66
AR1242ICC050	AR1242ICC050	10/09/2025	17:35	PO114324.D	8.67	3.66
AR1248ICC1000	AR1248ICC1000	10/09/2025	17:54	PO114325.D	8.67	3.66
AR1248ICC750	AR1248ICC750	10/09/2025	18:12	PO114326.D	8.67	3.66
AR1248ICC500	AR1248ICC500	10/09/2025	18:30	PO114327.D	8.67	3.66
AR1248ICC250	AR1248ICC250	10/09/2025	18:49	PO114328.D	8.67	3.66
AR1248ICC050	AR1248ICC050	10/09/2025	19:07	PO114329.D	8.67	3.66
AR1254ICC1000	AR1254ICC1000	10/09/2025	19:26	PO114330.D	8.67	3.66
AR1254ICC750	AR1254ICC750	10/09/2025	19:44	PO114331.D	8.67	3.66
AR1254ICC500	AR1254ICC500	10/09/2025	20:02	PO114332.D	8.67	3.66
AR1254ICC250	AR1254ICC250	10/09/2025	20:21	PO114333.D	8.67	3.66
AR1254ICC050	AR1254ICC050	10/09/2025	20:39	PO114334.D	8.67	3.66
AR1262ICC500	AR1262ICC500	10/09/2025	20:57	PO114335.D	8.67	3.66
AR1268ICC1000	AR1268ICC1000	10/09/2025	21:16	PO114336.D	8.67	3.66
AR1268ICC750	AR1268ICC750	10/09/2025	21:34	PO114337.D	8.67	3.66
AR1268ICC500	AR1268ICC500	10/09/2025	21:52	PO114338.D	8.67	3.66
AR1268ICC250	AR1268ICC250	10/09/2025	22:11	PO114339.D	8.67	3.66
AR1268ICC050	AR1268ICC050	10/09/2025	22:29	PO114340.D	8.67	3.66
AR1660CCC500	AR1660CCC500	11/03/2025	08:49	PO114834.D	8.64	3.65
IBLK	IBLK	11/03/2025	10:01	PO114838.D	8.64	3.65
PB170211BL	PB170211BL	11/03/2025	10:20	PO114839.D	8.64	3.65
AR1660CCC500	AR1660CCC500	11/03/2025	14:33	PO114849.D	8.64	3.65
IBLK	IBLK	11/03/2025	15:44	PO114853.D	8.64	3.65
IBLK	IBLK	10/10/2025	08:42	PQ070972.D	7.53	2.78
AR1660ICC1000	AR1660ICC1000	10/10/2025	08:57	PQ070973.D	7.53	2.78
AR1660ICC750	AR1660ICC750	10/10/2025	09:11	PQ070974.D	7.53	2.78
AR1660ICC500	AR1660ICC500	10/10/2025	09:26	PQ070975.D	7.53	2.78
AR1660ICC250	AR1660ICC250	10/10/2025	09:40	PQ070976.D	7.53	2.78
AR1660ICC050	AR1660ICC050	10/10/2025	09:55	PQ070977.D	7.53	2.78
AR1221ICC500	AR1221ICC500	10/10/2025	10:09	PQ070978.D	7.53	2.78
AR1232ICC500	AR1232ICC500	10/10/2025	10:24	PQ070979.D	7.53	2.78

Analytical Sequence

AR1242ICC1000	AR1242ICC1000	10/10/2025	10:38	PQ070980.D	7.53	2.78
AR1242ICC750	AR1242ICC750	10/10/2025	10:53	PQ070981.D	7.53	2.78
AR1242ICC500	AR1242ICC500	10/10/2025	11:08	PQ070982.D	7.53	2.78
AR1242ICC250	AR1242ICC250	10/10/2025	11:22	PQ070983.D	7.53	2.78
AR1242ICC050	AR1242ICC050	10/10/2025	11:37	PQ070984.D	7.53	2.78
AR1248ICC1000	AR1248ICC1000	10/10/2025	11:51	PQ070985.D	7.53	2.78
AR1248ICC750	AR1248ICC750	10/10/2025	12:06	PQ070986.D	7.53	2.78
AR1248ICC500	AR1248ICC500	10/10/2025	12:20	PQ070987.D	7.53	2.78
AR1248ICC250	AR1248ICC250	10/10/2025	12:35	PQ070988.D	7.53	2.78
AR1248ICC050	AR1248ICC050	10/10/2025	12:50	PQ070989.D	7.53	2.78
AR1254ICC1000	AR1254ICC1000	10/10/2025	13:04	PQ070990.D	7.53	2.78
AR1254ICC750	AR1254ICC750	10/10/2025	13:19	PQ070991.D	7.53	2.78
AR1254ICC500	AR1254ICC500	10/10/2025	13:33	PQ070992.D	7.53	2.78
AR1254ICC250	AR1254ICC250	10/10/2025	13:48	PQ070993.D	7.53	2.78
AR1254ICC050	AR1254ICC050	10/10/2025	15:15	PQ070994.D	7.53	2.78
AR1262ICC500	AR1262ICC500	10/10/2025	15:30	PQ070995.D	7.53	2.78
AR1268ICC1000	AR1268ICC1000	10/10/2025	15:44	PQ070996.D	7.53	2.78
AR1268ICC750	AR1268ICC750	10/10/2025	15:59	PQ070997.D	7.53	2.78
AR1268ICC500	AR1268ICC500	10/10/2025	16:13	PQ070998.D	7.53	2.78
AR1268ICC250	AR1268ICC250	10/10/2025	16:28	PQ070999.D	7.53	2.78
AR1268ICC050	AR1268ICC050	10/10/2025	16:43	PQ071000.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/23/2025	08:59	PQ071086.D	7.53	2.78
IBLK	IBLK	10/23/2025	10:42	PQ071090.D	7.53	2.78
FRAC-TANK-0760450	Q3400-01	10/23/2025	11:55	PQ071095.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/23/2025	13:37	PQ071098.D	7.53	2.78
IBLK	IBLK	10/23/2025	14:35	PQ071102.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/24/2025	08:52	PQ071119.D	7.53	2.78
IBLK	IBLK	10/24/2025	09:06	PQ071120.D	7.53	2.78
PB170211BS	PB170211BS	10/24/2025	09:34	PQ071121.D	7.53	2.77
PB170211BSD	PB170211BSD	10/24/2025	09:51	PQ071122.D	7.53	2.78
AR1660CCC500	AR1660CCC500	10/24/2025	10:36	PQ071124.D	7.53	2.78
IBLK	IBLK	10/24/2025	11:46	PQ071128.D	7.53	2.78



QC SAMPLE DATA

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: PB170211BL
Lab Sample ID: PB170211BL
Analytical Method: 8082A
Sample Wt/Vol: 1000 mL Final Vol: 10000 uL
Prep Method: Prep Date: 10/22/25

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: WATER
% Solid: 0
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	11/03/25 10:20	PB170211
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	11/03/25 10:20	PB170211
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	11/03/25 10:20	PB170211
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	11/03/25 10:20	PB170211
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	11/03/25 10:20	PB170211
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	11/03/25 10:20	PB170211
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	11/03/25 10:20	PB170211
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	11/03/25 10:20	PB170211
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	11/03/25 10:20	PB170211
SURROGATES										
877-09-8	Tetrachloro-m-xylene	20.6			35 - 137		103%	SPK: 20		
2051-24-3	Decachlorobiphenyl	21.0			40 - 135		105%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/09/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/09/25
Client Sample ID:	PIBLK-PO114312.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PO114312.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	10/09/25	po100925
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/09/25	po100925
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/09/25	po100925
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/09/25	po100925
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/09/25	po100925
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/09/25	po100925
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	10/09/25	po100925
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/09/25	po100925
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/09/25	po100925
SURROGATES										
877-09-8	Tetrachloro-m-xylene	18.6			60 - 140		93%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.0			60 - 140		95%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	11/03/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	11/03/25
Client Sample ID:	PIBLK-PO114838.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PO114838.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	11/03/25	PO110325
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	11/03/25	PO110325
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	11/03/25	PO110325
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	11/03/25	PO110325
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	11/03/25	PO110325
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	11/03/25	PO110325
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	11/03/25	PO110325
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	11/03/25	PO110325
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	11/03/25	PO110325
SURROGATES										
877-09-8	Tetrachloro-m-xylene	18.5			60 - 140		93%	SPK: 20		
2051-24-3	Decachlorobiphenyl	20.8			60 - 140		104%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	11/03/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	11/03/25
Client Sample ID:	PIBLK-PO114853.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PO114853.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	11/03/25	po110325
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	11/03/25	po110325
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	11/03/25	po110325
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	11/03/25	po110325
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	11/03/25	po110325
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	11/03/25	po110325
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	11/03/25	po110325
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	11/03/25	po110325
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	11/03/25	po110325
SURROGATES										
877-09-8	Tetrachloro-m-xylene	18.9			60 - 140		94%	SPK: 20		
2051-24-3	Decachlorobiphenyl	18.5			60 - 140		92%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/10/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/10/25
Client Sample ID:	PIBLK-PQ070972.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PQ070972.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	10/10/25	pq101025
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/10/25	pq101025
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/10/25	pq101025
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/10/25	pq101025
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/10/25	pq101025
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/10/25	pq101025
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	10/10/25	pq101025
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/10/25	pq101025
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/10/25	pq101025
SURROGATES										
877-09-8	Tetrachloro-m-xylene	18.3			60 - 140		91%	SPK: 20		
2051-24-3	Decachlorobiphenyl	19.0			60 - 140		95%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/23/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/23/25
Client Sample ID:	PIBLK-PQ071090.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PQ071090.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	10/23/25	pq102325
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/23/25	pq102325
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/23/25	pq102325
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/23/25	pq102325
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/23/25	pq102325
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/23/25	pq102325
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	10/23/25	pq102325
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/23/25	pq102325
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/23/25	pq102325
SURROGATES										
877-09-8	Tetrachloro-m-xylene	19.3			60 - 140		97%	SPK: 20		
2051-24-3	Decachlorobiphenyl	23.1			60 - 140		115%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/23/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/23/25
Client Sample ID:	PIBLK-PQ071102.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PQ071102.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	10/23/25	pq102325
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/23/25	pq102325
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/23/25	pq102325
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/23/25	pq102325
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/23/25	pq102325
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/23/25	pq102325
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	10/23/25	pq102325
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/23/25	pq102325
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/23/25	pq102325
SURROGATES										
877-09-8	Tetrachloro-m-xylene	19.1			60 - 140		95%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.9			60 - 140		114%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/24/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/24/25
Client Sample ID:	PIBLK-PQ071120.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PQ071120.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	10/24/25	PQ102425
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/24/25	PQ102425
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/24/25	PQ102425
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/24/25	PQ102425
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/24/25	PQ102425
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/24/25	PQ102425
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	10/24/25	PQ102425
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/24/25	PQ102425
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/24/25	PQ102425
SURROGATES										
877-09-8	Tetrachloro-m-xylene	18.0			60 - 140		90%	SPK: 20		
2051-24-3	Decachlorobiphenyl	21.9			60 - 140		110%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ICE Service Group, Inc.		Date Collected:	10/24/25
Project:	NWIRP Northrup Grumman Site – Bethpage, NY		Date Received:	10/24/25
Client Sample ID:	PIBLK-PQ071128.D		SDG No.:	Q3400
Lab Sample ID:	I.BLK-PQ071128.D		Matrix:	WATER
Analytical Method:	8082A		% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 10000 uL	Test:	PCB
Prep Method:	5030	Prep Date:		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	0.25	U	1	0.097	0.25	0.50	ug/L	10/24/25	PQ102425
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/24/25	PQ102425
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/24/25	PQ102425
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/24/25	PQ102425
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/24/25	PQ102425
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/24/25	PQ102425
11096-82-5	Aroclor-1260	0.25	U	1	0.081	0.25	0.50	ug/L	10/24/25	PQ102425
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/24/25	PQ102425
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/24/25	PQ102425
SURROGATES										
877-09-8	Tetrachloro-m-xylene	18.4			60 - 140		92%	SPK: 20		
2051-24-3	Decachlorobiphenyl	22.9			60 - 140		115%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: PB170211BS
Lab Sample ID: PB170211BS
Analytical Method: 8082A
Sample Wt/Vol: 1000 mL Final Vol: 10000 uL
Prep Method: Prep Date: 10/22/25

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: WATER
% Solid: 0
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	4.90		1	0.097	0.25	0.50	ug/L	10/24/25 09:34	PB170211
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/24/25 09:34	PB170211
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/24/25 09:34	PB170211
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/24/25 09:34	PB170211
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/24/25 09:34	PB170211
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/24/25 09:34	PB170211
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/24/25 09:34	PB170211
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/24/25 09:34	PB170211
11096-82-5	Aroclor-1260	5.30		1	0.081	0.25	0.50	ug/L	10/24/25 09:34	PB170211
SURROGATES										
877-09-8	Tetrachloro-m-xylene	21.6			35 - 137		108%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.2			40 - 135		126%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client: ICE Service Group, Inc.
Project: NWIRP Northrup Grumman Site – Bethpage, NY
Client Sample ID: PB170211BSD
Lab Sample ID: PB170211BSD
Analytical Method: 8082A
Sample Wt/Vol: 1000 mL Final Vol: 10000 uL
Prep Method: Prep Date: 10/22/25

Date Collected:
Date Received:
SDG No.: Q3400
Matrix: WATER
% Solid: 0
Test: PCB

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS										
12674-11-2	Aroclor-1016	5.30		1	0.097	0.25	0.50	ug/L	10/24/25 09:51	PB170211
11104-28-2	Aroclor-1221	0.40	U	1	0.13	0.40	0.50	ug/L	10/24/25 09:51	PB170211
11141-16-5	Aroclor-1232	0.25	U	1	0.096	0.25	0.50	ug/L	10/24/25 09:51	PB170211
53469-21-9	Aroclor-1242	0.25	U	1	0.12	0.25	0.50	ug/L	10/24/25 09:51	PB170211
12672-29-6	Aroclor-1248	0.25	U	1	0.071	0.25	0.50	ug/L	10/24/25 09:51	PB170211
11097-69-1	Aroclor-1254	0.25	U	1	0.094	0.25	0.50	ug/L	10/24/25 09:51	PB170211
37324-23-5	Aroclor-1262	0.40	U	1	0.14	0.40	0.50	ug/L	10/24/25 09:51	PB170211
11100-14-4	Aroclor-1268	0.25	U	1	0.11	0.25	0.50	ug/L	10/24/25 09:51	PB170211
11096-82-5	Aroclor-1260	5.40		1	0.081	0.25	0.50	ug/L	10/24/25 09:51	PB170211
SURROGATES										
877-09-8	Tetrachloro-m-xylene	22.4			35 - 137		112%	SPK: 20		
2051-24-3	Decachlorobiphenyl	25.7			40 - 135		128%	SPK: 20		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit