

Surrogate Summary

SDG No.: Q3400

Client: ICE Service Group, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO114312.D	PIBLK-PO114312.D	Tetrachloro-m-xyl	1	20	18.6	93		60	140
		Decachlorobiphen	1	20	19.2	96		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	19.0	95		60	140
I.BLK-PO114838.D	PIBLK-PO114838.D	Tetrachloro-m-xyl	1	20	20.9	104		60	140
		Decachlorobiphen	1	20	20.8	104		60	140
		Tetrachloro-m-xyl	2	20	18.5	93		60	140
		Decachlorobiphen	2	20	21.6	108		60	140
PB170211BL	PB170211BL	Tetrachloro-m-xyl	1	20	20.6	103		35	137
		Decachlorobiphen	1	20	20.4	102		40	135
		Tetrachloro-m-xyl	2	20	18.3	92		35	137
		Decachlorobiphen	2	20	21.0	105		40	135
I.BLK-PO114853.D	PIBLK-PO114853.D	Tetrachloro-m-xyl	1	20	21.5	107		60	140
		Decachlorobiphen	1	20	19.6	98		60	140
		Tetrachloro-m-xyl	2	20	18.9	94		60	140
		Decachlorobiphen	2	20	18.5	92		60	140
I.BLK-PQ070972.D	PIBLK-PQ070972.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	19.1	96		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
I.BLK-PQ071090.D	PIBLK-PQ071090.D	Tetrachloro-m-xyl	1	20	20.9	104		60	140
		Decachlorobiphen	1	20	23.7	118		60	140
		Tetrachloro-m-xyl	2	20	19.3	97		60	140
		Decachlorobiphen	2	20	23.1	115		60	140
Q3400-01	FRAC-TANK-0760450	Tetrachloro-m-xyl	1	20	23.4	117		35	137
		Decachlorobiphen	1	20	9.38	47		40	135
		Tetrachloro-m-xyl	2	20	21.7	109		35	137
		Decachlorobiphen	2	20	9.32	47		40	135
I.BLK-PQ071102.D	PIBLK-PQ071102.D	Tetrachloro-m-xyl	1	20	20.6	103		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
		Tetrachloro-m-xyl	2	20	19.1	95		60	140
		Decachlorobiphen	2	20	22.9	114		60	140
I.BLK-PQ071120.D	PIBLK-PQ071120.D	Tetrachloro-m-xyl	1	20	20.3	101		60	140
		Decachlorobiphen	1	20	23.2	116		60	140
		Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	21.9	110		60	140
PB170211BS	PB170211BS	Tetrachloro-m-xyl	1	20	21.6	108		35	137
		Decachlorobiphen	1	20	25.2	126		40	135
		Tetrachloro-m-xyl	2	20	16.9	85		35	137
		Decachlorobiphen	2	20	23.4	117		40	135
PB170211BSD	PB170211BSD	Tetrachloro-m-xyl	1	20	22.4	112		35	137

Surrogate Summary

SDG No.: Q3400

Client: ICE Service Group, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB170211BSD	PB170211BSD	Decachlorobiphen	1	20	25.7	128		40	135
		Tetrachloro-m-xyl	2	20	18.3	91		35	137
		Decachlorobiphen	2	20	24.2	121		40	135
I.BLK-PQ071128.D	PIBLK-PQ071128.D	Tetrachloro-m-xyl	1	20	20.8	104		60	140
		Decachlorobiphen	1	20	24.2	121		60	140
		Tetrachloro-m-xyl	2	20	18.4	92		60	140
		Decachlorobiphen	2	20	22.9	115		60	140