

Surrogate Summary

SW-846

SDG No.: Q3400

Client: ICE Service Group, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170222BL	PB170222BL	2-Fluorophenol	150	162	108		19	119
		Phenol-d6	150	157	105		10	130
		Nitrobenzene-d5	100	96.1	96		44	120
		2-Fluorobiphenyl	100	93.5	94		44	119
		2,4,6-Tribromophenol	150	143	95		43	140
		Terphenyl-d14	100	94.1	94		50	134
PB170222BS	PB170222BS	2-Fluorophenol	150	124	82		19	119
		Phenol-d6	150	123	82		10	130
		Nitrobenzene-d5	100	83.3	83		44	120
		2-Fluorobiphenyl	100	80.1	80		44	119
		2,4,6-Tribromophenol	150	127	84		43	140
		Terphenyl-d14	100	82.9	83		50	134
PB170222BSD	PB170222BSD	2-Fluorophenol	150	145	97		19	119
		Phenol-d6	150	125	83		10	130
		Nitrobenzene-d5	100	88.5	88		44	120
		2-Fluorobiphenyl	100	79.8	80		44	119
		2,4,6-Tribromophenol	150	124	83		43	140
		Terphenyl-d14	100	83.4	83		50	134
Q3400-01	FRAC-TANK-0760450	2-Fluorophenol	150	88.7	59		19	119
		Phenol-d6	150	57.7	38		10	130
		Nitrobenzene-d5	100	100	100		44	120
		2-Fluorobiphenyl	100	84.6	85		44	119
		2,4,6-Tribromophenol	150	137	91		43	140
		Terphenyl-d14	100	68.9	69		50	134