

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3400

Client ID : SSTDCCC040

Date Analyzed: 10/27/2025

Lab File ID: BG064583.D

Time Analyzed: 11:57

Instrument ID: BNA_G

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	39737	8.22	173874	11.05	136456	14.84
UPPER LIMIT	79474	8.72	347748	11.546	272912	15.342
LOWER LIMIT	19868.5	7.72	86937	10.546	68228	14.342
EPA SAMPLE NO.						
01 PB170222BL	45325	8.22	169145	11.04	133935	14.84
02 PB170222BS	45379	8.22	155959	11.05	125067	14.84
03 PB170222BSD	40304	8.22	162009	11.05	130466	14.84
04 FRAC-TANK-0760450	33849.5	8.22	134030	11.05	109277	14.84

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.